

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3607

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3980	4000	100	95 - 105	P	11/12/2025	16:56	LB137883

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3607

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	9.89	12.0	82	80 - 120	P	11/12/2025	17:25	LB137883

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3607

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4730	5000	95	90 - 110	P	11/12/2025	18:01	LB137883
CCV02	Lead	4790	5000	96	90 - 110	P	11/12/2025	19:04	LB137883
CCV03	Lead	4770	5000	95	90 - 110	P	11/12/2025	20:29	LB137883
CCV04	Lead	5200	5000	104	90 - 110	P	11/12/2025	21:46	LB137883
CCV05	Lead	4740	5000	95	90 - 110	P	11/12/2025	22:16	LB137883
CCV06	Lead	4780	5000	96	90 - 110	P	11/12/2025	22:53	LB137883