

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Total Quality Environmental Inc.

SDG No.: Q3607

Contract: TOTA02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	-8.49			-12	12	11/12/2025	17:44	LB137883
ICSAB01	Lead	38.8	49.0	79	37	61	11/12/2025	17:48	LB137883