

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Total Quality Environmental Inc. **level:** low **sdg no.:** Q3607
contract: TOTA02 **lab code:** ACE
matrix: Water **sample id:** Q3607-09 **client id:** main hall left sink 1st drawMS
Percent Solids for Sample: NA **Spiked ID:** Q3607-09MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	444		1.49	J	500	89		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Total Quality Environmental Inc.</u>	level:	<u>low</u>	sdg no.:	<u>Q3607</u>
contract:	<u>TOTA02</u>			lab code:	<u>ACE</u>
matrix:	<u>Water</u>	sample id:	<u>Q3607-09</u>	client id:	<u>main hall left sink 1st drawMSD</u>
Percent Solids for Sample:	NA	Spiked ID:	Q3607-09MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	418		1.49	J	500	83		P