

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3608

Project: Monthly 2025

Run No.: LB137878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137878BS							
TPH	mg/L	20.0	19.0		95	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3608

Project: Monthly 2025

Run No.: LB137878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137878BSD							
TPH	mg/L	20.0	17.8		89	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3608

Project: Monthly 2025

Run No.: LB137907

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137907BS							
BOD5	mg/L	198	190		96	1	84.6-115.4	11/13/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3608

Project: Monthly 2025

Run No.: LB137913

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137913BS							
TSS	mg/L	550	538		98	1	90-110	11/17/2025