

LAB CHRONICLE

OrderID:	Q3614	OrderDate:	11/12/2025 11:51:00 AM
Client:	Kleinfelder	Project:	Logan School
Contact:	Mark Warchol	Location:	J21,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3614-01	COMP-1	SOIL			11/11/25 10:05			11/12/25
			Ammonia	SM4500-NH3		11/17/25	11/17/25 14:34	
			Anions Group1	9056A			11/12/25 17:50	
			Hexavalent Chromium	7196A		11/13/25	11/13/25 15:40	
			Trivalent Chromium	6010D			11/13/25 15:31	
Q3614-02	COMP-2	SOIL			11/11/25 10:45			11/12/25
			Ammonia	SM4500-NH3		11/17/25	11/17/25 13:30	
			Anions Group1	9056A			11/12/25 19:38	
			Hexavalent Chromium	7196A		11/13/25	11/13/25 15:41	
			Trivalent Chromium	6010D			11/13/25 15:35	
Q3614-03	COMP-3	SOIL			11/11/25 11:30			11/12/25
			Ammonia	SM4500-NH3		11/17/25	11/17/25 13:31	
			Anions Group1	9056A			11/12/25 19:59	
			Hexavalent Chromium	7196A		11/13/25	11/13/25 15:42	
			Trivalent Chromium	6010D			11/13/25 15:39	



SAMPLE DATA

A

B

C

D

Report of Analysis

Client: Kleinfelder
Project: Logan School
Client Sample ID: COMP-1
Lab Sample ID: Q3614-01

Date Collected: 11/11/25 10:05
Date Received: 11/12/25
SDG No.: Q3614
Matrix: SOIL
% Solid: 82.3

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	3.20	J	1	2.60	6.00	mg/Kg	11/17/25 10:10	11/17/25 14:34	SM 4500-NH3 B plus G-21
Chloride	4.30	J	1	4.10	14.4	mg/Kg		11/12/25 17:50	9056A
Fluoride	7.50	J	1	2.10	9.60	mg/Kg		11/12/25 17:50	9056A
Sulfate	20.2	J	1	10.5	71.8	mg/Kg		11/12/25 17:50	9056A
Hexavalent Chromium	0.084	U	1	0.084	0.48	mg/Kg	11/13/25 12:00	11/13/25 15:40	7196A
Trivalent Chromium	24.0		1	0.61	0.61	mg/Kg		11/13/25 15:31	6010D

Comments: _____

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client: Kleinfelder
Project: Logan School
Client Sample ID: COMP-2
Lab Sample ID: Q3614-02

Date Collected: 11/11/25 10:45
Date Received: 11/12/25
SDG No.: Q3614
Matrix: SOIL
% Solid: 81.7

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	2.70	U	1	2.70	6.10	mg/Kg	11/17/25 10:10	11/17/25 13:30	SM 4500-NH3 B plus G-21
Chloride	8.10	J	1	4.20	14.6	mg/Kg		11/12/25 19:38	9056A
Fluoride	6.50	J	1	2.10	9.70	mg/Kg		11/12/25 19:38	9056A
Sulfate	21.0	J	1	10.7	73.0	mg/Kg		11/12/25 19:38	9056A
Hexavalent Chromium	0.14	J	1	0.084	0.48	mg/Kg	11/13/25 12:00	11/13/25 15:41	7196A
Trivalent Chromium	25.2		1	0.61	0.61	mg/Kg		11/13/25 15:35	6010D

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Kleinfelder
Project: Logan School
Client Sample ID: COMP-3
Lab Sample ID: Q3614-03

Date Collected: 11/11/25 11:30
Date Received: 11/12/25
SDG No.: Q3614
Matrix: SOIL
% Solid: 83.5

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	4.00	J	1	2.60	5.80	mg/Kg	11/17/25 10:10	11/17/25 13:31	SM 4500-NH3 B plus G-21
Chloride	12.6	J	1	4.10	14.3	mg/Kg		11/12/25 19:59	9056A
Fluoride	5.90	J	1	2.10	9.50	mg/Kg		11/12/25 19:59	9056A
Sulfate	17.2	J	1	10.4	71.3	mg/Kg		11/12/25 19:59	9056A
Hexavalent Chromium	0.082	U	1	0.082	0.47	mg/Kg	11/13/25 12:00	11/13/25 15:42	7196A
Trivalent Chromium	25.0		1	0.60	0.60	mg/Kg		11/13/25 15:39	6010D

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: Kleinfelder
Project: Logan School

SDG No.: Q3614
RunNo.: LB137882

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.7	10	97	90-110	11/03/2025
Chloride	mg/L	2.9	3	97	90-110	11/03/2025
Fluoride	mg/L	2	2	100	90-110	11/03/2025
Nitrite	mg/L	2.9	3	97	90-110	11/03/2025
Nitrate	mg/L	2.4	2.5	96	90-110	11/03/2025
Sulfate	mg/L	14.5	15	97	90-110	11/03/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	11/03/2025
Sample ID: CCV1						
Bromide	mg/L	10.5	10	105	90-110	11/12/2025
Chloride	mg/L	3.2	3	107	90-110	11/12/2025
Fluoride	mg/L	2.1	2	105	90-110	11/12/2025
Nitrite	mg/L	3.1	3	103	90-110	11/12/2025
Nitrate	mg/L	2.6	2.5	104	90-110	11/12/2025
Sulfate	mg/L	15.4	15	103	90-110	11/12/2025
Orthophosphate as P	mg/L	4.7	5	94	90-110	11/12/2025
Sample ID: CCV2						
Bromide	mg/L	10.5	10	105	90-110	11/12/2025
Chloride	mg/L	3.2	3	107	90-110	11/12/2025
Fluoride	mg/L	2.1	2	105	90-110	11/12/2025
Nitrite	mg/L	3.1	3	103	90-110	11/12/2025
Nitrate	mg/L	2.6	2.5	104	90-110	11/12/2025
Sulfate	mg/L	15.5	15	103	90-110	11/12/2025
Orthophosphate as P	mg/L	5	5	100	90-110	11/12/2025
Sample ID: CCV3						
Bromide	mg/L	10.5	10	105	90-110	11/12/2025
Chloride	mg/L	3.2	3	107	90-110	11/12/2025
Fluoride	mg/L	2.1	2	105	90-110	11/12/2025
Nitrite	mg/L	3.1	3	103	90-110	11/12/2025
Nitrate	mg/L	2.6	2.5	104	90-110	11/12/2025
Sulfate	mg/L	15.8	15	105	90-110	11/12/2025
Orthophosphate as P	mg/L	5.5	5	110	90-110	11/12/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

RunNo.: LB137886

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.488	0.5	98	90-110	11/13/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/13/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.491	0.5	98	90-110	11/13/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.485	0.5	97	90-110	11/13/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

RunNo.: LB137922

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	11/17/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.95	1	95	90-110	11/17/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.97	1	97	90-110	11/17/2025

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder
Project: Logan School

SDG No.: Q3614
RunNo.: LB137882

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/03/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/03/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/03/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/03/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/03/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/03/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/03/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/12/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/12/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/12/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/12/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/12/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/12/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/12/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/12/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/12/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/12/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/12/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/12/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/12/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/12/2025
Sample ID: CCB3							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/12/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/12/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/12/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/12/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/12/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/12/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/12/2025

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

RunNo.: LB137886

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	11/13/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	11/13/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	11/13/2025
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	11/13/2025

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

RunNo.: LB137922

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025

Preparation Blank Summary

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137882BLS							
Bromide	mg/Kg	< 20.0000	20.0000	U	7	40	11/12/2025
Chloride	mg/Kg	< 6.0000	6.0000	U	3.5	12	11/12/2025
Fluoride	mg/Kg	< 4.0000	4.0000	U	1.8	8	11/12/2025
Nitrite	mg/Kg	< 6.0000	6.0000	U	1.5	12	11/12/2025
Nitrate	mg/Kg	< 5.0000	5.0000	U	1.8	10	11/12/2025
Sulfate	mg/Kg	< 30.0000	30.0000	U	8.8	60	11/12/2025
Orthophosphate as P	mg/Kg	< 10.0000	10.0000	U	6.7	20	11/12/2025
Sample ID: PB170537BL							
Hexavalent Chromium	mg/Kg	0.08	0.2000	J	0.07	0.4	11/13/2025
Sample ID: PB170569BL							
Ammonia as N	mg/Kg	< 2.5000	2.5000	U	2.2	5	11/17/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3609-01
Client ID:	AU-713-COMP-01MS	Percent Solids for Spike Sample:	81.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	1520		0.095	J	1570	40	97		11/13/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3609-01
Client ID:	AU-713-COMP-01MS	Percent Solids for Spike Sample:	81.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	85-115	42.2		0.095	J	49.0	2	86		11/13/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3609-01
Client ID:	AU-713-COMP-01MS	Percent Solids for Spike Sample:	81.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	40.1		0.095	J	49.0	2	82		11/13/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3609-04
Client ID:	AU-713-COMP-02MS	Percent Solids for Spike Sample:	84.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/Kg	80-120	234		8.20	U	230	1	102		11/12/2025
Chloride	mg/Kg	80-120	81.1		11.7	J	70.1	1	99		11/12/2025
Fluoride	mg/Kg	80-120	42.8		6.30	J	46.7	1	78	*	11/12/2025
Nitrite	mg/Kg	80-120	69.8		1.80	U	70.1	1	100		11/12/2025
Nitrate	mg/Kg	80-120	61.2		7.80	J	58.4	1	91		11/12/2025
Sulfate	mg/Kg	80-120	359		23.0	J	350	1	96		11/12/2025
Orthophosphate as P	mg/Kg	80-120	82.6		7.80	U	120	1	69	*	11/12/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3609-04
Client ID:	AU-713-COMP-02MSD	Percent Solids for Spike Sample:	84.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/Kg	80-120	236		8.20	U	240	1	98		11/12/2025
Chloride	mg/Kg	80-120	81.5		11.7	J	70.6	1	99		11/12/2025
Fluoride	mg/Kg	80-120	43.1		6.30	J	47.1	1	78	*	11/12/2025
Nitrite	mg/Kg	80-120	70.5		1.80	U	70.6	1	100		11/12/2025
Nitrate	mg/Kg	80-120	61.7		7.80	J	58.9	1	92		11/12/2025
Sulfate	mg/Kg	80-120	358		23.0	J	350	1	96		11/12/2025
Orthophosphate as P	mg/Kg	80-120	82.6		7.80	U	120	1	69	*	11/12/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3614-01
Client ID:	COMP-1MS	Percent Solids for Spike Sample:	82.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	57.5		3.20	J	58.4	1	93		11/17/2025

Matrix Spike Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3614-01
Client ID:	COMP-1MSD	Percent Solids for Spike Sample:	82.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	56.1		3.20	J	59.6	1	89		11/17/2025

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3609-01
Client ID:	AU-713-COMP-01DUP	Percent Solids for Spike Sample:	81.6

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	+/-20	0.095	J	0.084	U	1	200	*	11/13/2025

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3609-04
Client ID:	AU-713-COMP-02MSD	Percent Solids for Spike Sample:	84.6

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Chloride	mg/Kg	+/-15	81.1		81.5		1	0		11/12/2025
Orthophosphate as P	mg/Kg	+/-15	82.6		82.6		1	0		11/12/2025
Sulfate	mg/Kg	+/-15	359		358		1	0		11/12/2025
Bromide	mg/Kg	+/-15	234		236		1	1		11/12/2025
Fluoride	mg/Kg	+/-15	42.8		43.1		1	1		11/12/2025
Nitrate	mg/Kg	+/-15	61.2		61.7		1	1		11/12/2025
Nitrite	mg/Kg	+/-15	69.8		70.5		1	1		11/12/2025

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3614-01
Client ID:	COMP-1DUP	Percent Solids for Spike Sample:	82.3

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	3.20	J	2.60	U	1	200	*	11/17/2025

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3614-01
Client ID:	COMP-1MSD	Percent Solids for Spike Sample:	82.3

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	57.5		56.1		1	2		11/17/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

Run No.: LB137882

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137882BSS							
Bromide	mg/Kg	200	210		105	1	90-110	11/12/2025
Chloride	mg/Kg	60	63.1		105	1	90-110	11/12/2025
Fluoride	mg/Kg	40	42.1		105	1	90-110	11/12/2025
Nitrite	mg/Kg	60	62.6		104	1	90-110	11/12/2025
Nitrate	mg/Kg	50	52.0		104	1	90-110	11/12/2025
Sulfate	mg/Kg	300	312		104	1	90-110	11/12/2025
Orthophosphate as P	mg/Kg	100	99.1		99	1	90-110	11/12/2025

Laboratory Control Sample Summary

Client: Kleinfelder
Project: Logan School

SDG No.: Q3614
Run No.: LB137886

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170537BS							
Hexavalent Chromium	mg/Kg	20	20.6		103	1	84-110	11/13/2025

Laboratory Control Sample Summary

Client: Kleinfelder
Project: Logan School

SDG No.: Q3614
Run No.: LB137922

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170569BS							
Ammonia as N	mg/Kg	50	50.3		101	1	90-110	11/17/2025