

Initial and Continuing Calibration Verification

Client: Kleinfelder
Project: Logan School

SDG No.: Q3614
RunNo.: LB137882

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.7	10	97	90-110	11/03/2025
Chloride	mg/L	2.9	3	97	90-110	11/03/2025
Fluoride	mg/L	2	2	100	90-110	11/03/2025
Nitrite	mg/L	2.9	3	97	90-110	11/03/2025
Nitrate	mg/L	2.4	2.5	96	90-110	11/03/2025
Sulfate	mg/L	14.5	15	97	90-110	11/03/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	11/03/2025
Sample ID: CCV1						
Bromide	mg/L	10.5	10	105	90-110	11/12/2025
Chloride	mg/L	3.2	3	107	90-110	11/12/2025
Fluoride	mg/L	2.1	2	105	90-110	11/12/2025
Nitrite	mg/L	3.1	3	103	90-110	11/12/2025
Nitrate	mg/L	2.6	2.5	104	90-110	11/12/2025
Sulfate	mg/L	15.4	15	103	90-110	11/12/2025
Orthophosphate as P	mg/L	4.7	5	94	90-110	11/12/2025
Sample ID: CCV2						
Bromide	mg/L	10.5	10	105	90-110	11/12/2025
Chloride	mg/L	3.2	3	107	90-110	11/12/2025
Fluoride	mg/L	2.1	2	105	90-110	11/12/2025
Nitrite	mg/L	3.1	3	103	90-110	11/12/2025
Nitrate	mg/L	2.6	2.5	104	90-110	11/12/2025
Sulfate	mg/L	15.5	15	103	90-110	11/12/2025
Orthophosphate as P	mg/L	5	5	100	90-110	11/12/2025
Sample ID: CCV3						
Bromide	mg/L	10.5	10	105	90-110	11/12/2025
Chloride	mg/L	3.2	3	107	90-110	11/12/2025
Fluoride	mg/L	2.1	2	105	90-110	11/12/2025
Nitrite	mg/L	3.1	3	103	90-110	11/12/2025
Nitrate	mg/L	2.6	2.5	104	90-110	11/12/2025
Sulfate	mg/L	15.8	15	105	90-110	11/12/2025
Orthophosphate as P	mg/L	5.5	5	110	90-110	11/12/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

RunNo.: LB137886

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.488	0.5	98	90-110	11/13/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/13/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.491	0.5	98	90-110	11/13/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.485	0.5	97	90-110	11/13/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

RunNo.: LB137922

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	11/17/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.95	1	95	90-110	11/17/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.97	1	97	90-110	11/17/2025