

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder
Project: Logan School

SDG No.: Q3614
RunNo.: LB137882

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/03/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/03/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/03/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/03/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/03/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/03/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/03/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/12/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/12/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/12/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/12/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/12/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/12/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/12/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/12/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/12/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/12/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/12/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/12/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/12/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/12/2025
Sample ID: CCB3							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/12/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/12/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/12/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/12/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/12/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/12/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/12/2025

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

RunNo.: LB137886

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	11/13/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	11/13/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	11/13/2025
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	11/13/2025

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

RunNo.: LB137922

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/17/2025