

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q3614
Project:	Logan School	Sample ID:	Q3609-01
Client ID:	AU-713-COMP-01DUP	Percent Solids for Spike Sample:	81.6

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	+/-20	0.095	J	0.084	U	1	200	*	11/13/2025

Duplicate Sample Summary

Client: Kleinfelder	SDG No.: Q3614
Project: Logan School	Sample ID: Q3609-04
Client ID: AU-713-COMP-02MSD	Percent Solids for Spike Sample: 84.6

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Chloride	mg/Kg	+/-15	81.1		81.5		1	0		11/12/2025
Orthophosphate as P	mg/Kg	+/-15	82.6		82.6		1	0		11/12/2025
Sulfate	mg/Kg	+/-15	359		358		1	0		11/12/2025
Bromide	mg/Kg	+/-15	234		236		1	1		11/12/2025
Fluoride	mg/Kg	+/-15	42.8		43.1		1	1		11/12/2025
Nitrate	mg/Kg	+/-15	61.2		61.7		1	1		11/12/2025
Nitrite	mg/Kg	+/-15	69.8		70.5		1	1		11/12/2025

Duplicate Sample Summary

Client: Kleinfelder Project: Logan School Client ID: COMP-1DUP	SDG No.: Q3614 Sample ID: Q3614-01 Percent Solids for Spike Sample: 82.3
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	3.20	J	2.60	U	1	200	*	11/17/2025

Duplicate Sample Summary

Client: Kleinfelder Project: Logan School Client ID: COMP-1MSD	SDG No.: Q3614 Sample ID: Q3614-01 Percent Solids for Spike Sample: 82.3
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	57.5		56.1		1	2		11/17/2025