

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder

SDG No.: Q3614

Contract: POWE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170522BS							
Aluminum	mg/Kg	100	97.5		98	80 - 120	P
Antimony	mg/Kg	40.0	39.0		98	80 - 120	P
Arsenic	mg/Kg	40.0	37.9		95	80 - 120	P
Barium	mg/Kg	10.0	9.51		95	80 - 120	P
Beryllium	mg/Kg	10.0	9.94		99	80 - 120	P
Boron	mg/Kg	15.0	14.5		97	80 - 120	P
Cadmium	mg/Kg	10.0	9.37		94	80 - 120	P
Chromium	mg/Kg	20.0	19.8		99	80 - 120	P
Cobalt	mg/Kg	10.0	9.62		96	80 - 120	P
Copper	mg/Kg	15.0	15.0		100	80 - 120	P
Iron	mg/Kg	150	153		102	80 - 120	P
Lead	mg/Kg	50.0	46.2		92	80 - 120	P
Manganese	mg/Kg	10.0	10.1		100	80 - 120	P
Molybdenum	mg/Kg	20.0	19.4		97	80 - 120	P
Nickel	mg/Kg	25.0	24.1		96	80 - 120	P
Selenium	mg/Kg	100	95.0		95	80 - 120	P
Silver	mg/Kg	3.8	3.65		96	80 - 120	P
Thallium	mg/Kg	100	95.4		95	80 - 120	P
Vanadium	mg/Kg	15.0	14.7		98	80 - 120	P
Zinc	mg/Kg	10.0	9.97		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder

SDG No.: Q3614

Contract: POWE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170524BS Mercury	mg/Kg	0.26	0.22		83	80 - 120	CV