

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 16:50

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	4.64	4.65	4.55	4.75	0.01
Decachlorobiphenyl	10.41	10.42	10.32	10.52	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 16:50

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl	8.77	8.79	8.69	8.89	0.02