

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3614
 GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2025 10/21/2025

Client Sample No.: CCAL04 Date Analyzed: 11/13/2025
 Lab Sample No.: AR1660CCC500 Data File : PP076498.D Time Analyzed: 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.789	5.695	5.895	530.640	500.000	6.1
Aroclor-1016-2	5.811	5.717	5.917	538.090	500.000	7.6
Aroclor-1016-3	5.873	5.780	5.980	554.420	500.000	10.9
Aroclor-1016-4	5.970	5.877	6.077	563.300	500.000	12.7
Aroclor-1016-5	6.263	6.169	6.369	554.730	500.000	10.9
Aroclor-1260-1	7.381	7.287	7.487	570.430	500.000	14.1
Aroclor-1260-2	7.634	7.540	7.740	567.830	500.000	13.6
Aroclor-1260-3	7.993	7.899	8.099	588.380	500.000	17.7
Aroclor-1260-4	8.220	8.127	8.327	579.720	500.000	15.9
Aroclor-1260-5	8.545	8.452	8.652	545.360	500.000	9.1
Decachlorobiphenyl	10.408	10.319	10.519	49.930	50.000	-0.1
Tetrachloro-m-xylene	4.636	4.542	4.742	52.430	50.000	4.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076498.D **Time Analyzed:** 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.870	4.778	4.978	505.390	500.000	1.1
Aroclor-1016-2	4.927	4.836	5.036	500.680	500.000	0.1
Aroclor-1016-3	5.047	4.954	5.154	500.710	500.000	0.1
Aroclor-1016-4	5.088	4.996	5.196	509.680	500.000	1.9
Aroclor-1016-5	5.301	5.210	5.410	507.050	500.000	1.4
Aroclor-1260-1	6.329	6.238	6.438	584.480	500.000	16.9
Aroclor-1260-2	6.517	6.425	6.625	566.140	500.000	13.2
Aroclor-1260-3	6.670	6.578	6.778	548.930	500.000	9.8
Aroclor-1260-4	7.138	7.047	7.247	578.240	500.000	15.6
Aroclor-1260-5	7.379	7.287	7.487	535.110	500.000	7.0
Decachlorobiphenyl	8.770	8.682	8.882	45.660	50.000	-8.7
Tetrachloro-m-xylene	3.776	3.681	3.881	51.520	50.000	3.0