

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: Kleinfelder
Project: Logan School
Client Sample ID: COMP-1
Lab Sample ID: Q3614-01
Analytical Method: 8082A

Sample Wt/Vol: 30.09 g Final Vol: 10000 uL Prep Method: SW3541B Prep Date: 11/13/25 Date Collected: 11/11/25
Date Received: 11/12/25
SDG No.: Q3614
Matrix: SOIL
% Solid: 82.3

Test: PCB Group1

| CAS Number | · Parameter | Conc. | Qua. | DF | MDL | LOQ / CRQL | Units | Date Ana. | BatchID |
|------------|----------------------|-------|------|----|----------|------------|---------|----------------|----------|
| TARGETS | | | | | | | | | |
| 12674-11-2 | Aroclor-1016 | 4.80 | U | 1 | 4.80 | 20.6 | ug/kg | 11/13/25 14:55 | PB170526 |
| 11097-69-1 | Aroclor-1254 | 3.90 | U | 1 | 3.90 | 20.6 | ug/kg | 11/13/25 14:55 | PB170526 |
| 11096-82-5 | Aroclor-1260 | 3.90 | U | 1 | 3.90 | 20.6 | ug/kg | 11/13/25 14:55 | PB170526 |
| SURROGATES | | | | | | | | | |
| 877-09-8 | Tetrachloro-m-xylene | 20.5 | | | 21 - 165 | 103% | SPK: 20 | | |
| 2051-24-3 | Decachlorobiphenyl | 19.3 | | | 10 - 170 | 97% | SPK: 20 | | |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit