

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/17/2025 10/17/2025

Continuing Calib Time: 16:32

Initial Calibration Time(s): 11:34 12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.07       | 9.06      | 8.96                   | 9.16 | -0.01      |
| Tetrachloro-m-xylene | 3.54       | 3.55      | 3.45                   | 3.65 | 0.01       |
| Aldrin               | 5.26       | 5.26      | 5.16                   | 5.36 | 0.00       |
| Dieldrin             | 6.34       | 6.34      | 6.24                   | 6.44 | 0.00       |
| 4,4'-DDE             | 6.19       | 6.19      | 6.09                   | 6.29 | 0.00       |
| 4,4'-DDD             | 6.70       | 6.70      | 6.60                   | 6.80 | 0.00       |
| 4,4'-DDT             | 7.02       | 7.02      | 6.92                   | 7.12 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/17/2025 10/17/2025

Continuing Calib Time: 16:32

Initial Calibration Time(s): 11:34 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 8.06       | 8.06      | 7.96                   | 8.16 | 0.00       |
| Tetrachloro-m-xylene | 2.87       | 2.88      | 2.78                   | 2.98 | 0.01       |
| Aldrin               | 4.36       | 4.36      | 4.26                   | 4.46 | 0.00       |
| Dieldrin             | 5.50       | 5.50      | 5.40                   | 5.60 | 0.00       |
| 4,4'-DDE             | 5.36       | 5.37      | 5.27                   | 5.47 | 0.01       |
| 4,4'-DDD             | 5.92       | 5.92      | 5.82                   | 6.02 | 0.00       |
| 4,4'-DDT             | 6.17       | 6.17      | 6.07                   | 6.27 | 0.00       |