

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091143.D **Time Analyzed:** 09:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.699	6.600	6.800	52.030	50.000	4.1
4,4'-DDE	6.189	6.090	6.290	45.810	50.000	-8.4
4,4'-DDT	7.014	6.915	7.115	42.630	50.000	-14.7
Aldrin	5.263	5.164	5.364	47.360	50.000	-5.3
Decachlorobiphenyl	9.064	8.964	9.164	44.110	50.000	-11.8
Dieldrin	6.340	6.240	6.440	46.880	50.000	-6.2
Tetrachloro-m-xylene	3.544	3.445	3.645	47.340	50.000	-5.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091143.D **Time Analyzed:** 09:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	5.918	5.821	6.021	56.260	50.000	12.5
4,4'-DDE	5.363	5.266	5.466	52.520	50.000	5.0
4,4'-DDT	6.171	6.074	6.274	48.030	50.000	-3.9
Aldrin	4.358	4.261	4.461	52.420	50.000	4.8
Decachlorobiphenyl	8.059	7.962	8.162	54.170	50.000	8.3
Dieldrin	5.501	5.404	5.604	52.360	50.000	4.7
Tetrachloro-m-xylene	2.874	2.775	2.975	52.510	50.000	5.0