

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: POWE02
Lab Code: ACE
SDG NO.: Q3614
GC Column: ZB-MR1
ID: 0.32 (mm)
 Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No.: CCAL02
Date Analyzed: 11/13/2025
Lab Sample No.: PSTDCCC050
Data File : PD091161.D
Time Analyzed: 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	50.860	50.000	1.7
4,4'-DDE	6.190	6.090	6.290	44.260	50.000	-11.5
4,4'-DDT	7.015	6.915	7.115	41.060	50.000	-17.9
Aldrin	5.263	5.164	5.364	46.500	50.000	-7.0
Decachlorobiphenyl	9.066	8.964	9.164	44.150	50.000	-11.7
Dieldrin	6.340	6.240	6.440	45.330	50.000	-9.3
Tetrachloro-m-xylene	3.544	3.445	3.645	47.040	50.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091161.D **Time Analyzed:** 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.919	5.821	6.021	53.110	50.000	6.2
4,4'-DDE	5.363	5.266	5.466	49.670	50.000	-0.7
4,4'-DDT	6.172	6.074	6.274	44.260	50.000	-11.5
Aldrin	4.359	4.261	4.461	50.560	50.000	1.1
Decachlorobiphenyl	8.059	7.962	8.162	48.030	50.000	-3.9
Dieldrin	5.502	5.404	5.604	49.540	50.000	-0.9
Tetrachloro-m-xylene	2.874	2.775	2.975	51.260	50.000	2.5