

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091220.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.700	6.599	6.799	46.860	50.000	-6.3
4,4'-DDE	6.189	6.089	6.289	48.040	50.000	-3.9
4,4'-DDT	7.016	6.914	7.114	48.850	50.000	-2.3
Aldrin	5.263	5.163	5.363	48.120	50.000	-3.8
Decachlorobiphenyl	9.065	8.963	9.163	45.920	50.000	-8.2
Dieldrin	6.340	6.239	6.439	47.780	50.000	-4.4
Tetrachloro-m-xylene	3.544	3.444	3.644	48.000	50.000	-4.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091220.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.918	5.817	6.017	49.260	50.000	-1.5
4,4'-DDE	5.363	5.263	5.463	51.190	50.000	2.4
4,4'-DDT	6.172	6.071	6.271	52.440	50.000	4.9
Aldrin	4.358	4.258	4.458	52.180	50.000	4.4
Decachlorobiphenyl	8.059	7.958	8.158	48.700	50.000	-2.6
Dieldrin	5.502	5.401	5.601	51.270	50.000	2.5
Tetrachloro-m-xylene	2.873	2.774	2.974	51.820	50.000	3.6