

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091236.D **Time Analyzed:** 20:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.697	6.599	6.799	44.500	50.000	-11.0
4,4'-DDE	6.187	6.089	6.289	49.510	50.000	-1.0
4,4'-DDT	7.013	6.914	7.114	58.790	50.000	17.6
Aldrin	5.261	5.163	5.363	48.730	50.000	-2.5
Decachlorobiphenyl	9.062	8.963	9.163	46.500	50.000	-7.0
Dieldrin	6.337	6.239	6.439	48.980	50.000	-2.0
Tetrachloro-m-xylene	3.542	3.444	3.644	48.160	50.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091236.D **Time Analyzed:** 20:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	45.100	50.000	-9.8
4,4'-DDE	5.362	5.263	5.463	49.970	50.000	-0.1
4,4'-DDT	6.169	6.071	6.271	59.320	50.000	18.6
Aldrin	4.356	4.258	4.458	49.970	50.000	-0.1
Decachlorobiphenyl	8.057	7.958	8.158	44.480	50.000	-11.0
Dieldrin	5.500	5.401	5.601	49.880	50.000	-0.2
Tetrachloro-m-xylene	2.872	2.774	2.974	49.870	50.000	-0.3