

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q3614	OrderDate:	11/12/2025 11:51:00 AM
Client:	Kleinfelder	Project:	Logan School
Contact:	Mark Warchol	Location:	J21,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3614-01	COMP-1	SOIL			11/11/25			11/12/25
			PCB Group1	8082A		11/13/25	11/13/25	
			PESTICIDE Group1	8081B		11/13/25	11/13/25	
Q3614-02	COMP-2	SOIL			11/11/25			11/12/25
			PCB Group1	8082A		11/13/25	11/13/25	
			PESTICIDE Group1	8081B		11/13/25	11/13/25	
Q3614-03	COMP-3	SOIL			11/11/25			11/12/25
			PCB Group1	8082A		11/13/25	11/13/25	
			PESTICIDE Group1	8081B		11/13/25	11/13/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3614

Order ID: Q3614

Client: Kleinfelder

Project ID: Logan School

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: **Q3614**

Client: **Kleinfelder**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		31	149
		Tetrachloro-m-xyl	1	20	19.4	97		41	134
		Decachlorobiphen	2	20	20.3	102		31	149
		Tetrachloro-m-xyl	2	20	20.0	100		41	134
I.BLK-PD091141.D	PIBLK-PD091141.D	Decachlorobiphen	1	20	21.3	107		31	149
		Tetrachloro-m-xyl	1	20	21.9	109		41	134
		Decachlorobiphen	2	20	24.7	123		31	149
		Tetrachloro-m-xyl	2	20	24.7	123		41	134
PB170527BL	PB170527BL	Decachlorobiphen	1	20	18.8	94		10	143
		Tetrachloro-m-xyl	1	20	17.8	89		10	131
		Decachlorobiphen	2	20	21.6	108		10	143
		Tetrachloro-m-xyl	2	20	20.9	104		10	131
Q3609-13MS	AU-713-COMP-05MS	Decachlorobiphen	1	20	16.6	83		10	143
		Tetrachloro-m-xyl	1	20	17.1	86		10	131
		Decachlorobiphen	2	20	19.3	96		10	143
		Tetrachloro-m-xyl	2	20	20.7	104		10	131
Q3609-13MSD	AU-713-COMP-05MSD	Decachlorobiphen	1	20	15.3	76		10	143
		Tetrachloro-m-xyl	1	20	16.5	83		10	131
		Decachlorobiphen	2	20	18.6	93		10	143
		Tetrachloro-m-xyl	2	20	19.6	98		10	131
Q3614-01	COMP-1	Decachlorobiphen	1	20	16.5	82		10	143
		Tetrachloro-m-xyl	1	20	17.1	85		10	131
		Decachlorobiphen	2	20	19.4	97		10	143
		Tetrachloro-m-xyl	2	20	19.6	98		10	131
Q3614-02	COMP-2	Decachlorobiphen	1	20	16.4	82		10	143
		Tetrachloro-m-xyl	1	20	15.7	78		10	131
		Decachlorobiphen	2	20	17.2	86		10	143
		Tetrachloro-m-xyl	2	20	18.7	94		10	131
Q3614-03	COMP-3	Decachlorobiphen	1	20	16.1	80		10	143
		Tetrachloro-m-xyl	1	20	17.5	87		10	131
		Decachlorobiphen	2	20	18.3	91		10	143
		Tetrachloro-m-xyl	2	20	20.4	102		10	131
I.BLK-PD091160.D	PIBLK-PD091160.D	Decachlorobiphen	1	20	20.4	102		31	149
		Tetrachloro-m-xyl	1	20	20.6	103		41	134
		Decachlorobiphen	2	20	21.1	105		31	149
		Tetrachloro-m-xyl	2	20	23.4	117		41	134
I.BLK-PD091193.D	PIBLK-PD091193.D	Decachlorobiphen	1	20	20.8	104		31	149
		Tetrachloro-m-xyl	1	20	19.5	98		41	134
		Decachlorobiphen	2	20	20.1	100		31	149
		Tetrachloro-m-xyl	2	20	20.0	100		41	134
I.BLK-PD091218.D	PIBLK-PD091218.D	Decachlorobiphen	1	20	20.0	100		31	149

Surrogate Summary

SDG No.: Q3614

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD091218.D	PIBLK-PD091218.D	Tetrachloro-m-xyl	1	20	19.6	98		41	134
		Decachlorobiphen	2	20	19.3	96		31	149
PB170527BS	PB170527BS	Tetrachloro-m-xyl	2	20	22.6	113		41	134
		Decachlorobiphen	1	20	23.5	117		10	143
		Tetrachloro-m-xyl	1	20	21.5	108		10	131
		Decachlorobiphen	2	20	24.9	124		10	143
		Tetrachloro-m-xyl	2	20	24.1	120		10	131
I.BLK-PD091234.D	PIBLK-PD091234.D	Decachlorobiphen	1	20	20.8	104		31	149
		Tetrachloro-m-xyl	1	20	20.0	100		41	134
		Decachlorobiphen	2	20	18.9	94		31	149
		Tetrachloro-m-xyl	2	20	21.4	107		41	134

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3614</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	DataFile :	<u>PD091151.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result				Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3609-13MS (Column 1)	Client Sample ID:		AU-713-COMP-05MS								
	Aldrin	19.61	0	20.6	ug/kg	105				38	139	
	Dieldrin	19.61	0	20.5	ug/kg	105				32	140	
	4,4'-DDE	19.61	0	20.2	ug/kg	103				20	156	
	4,4'-DDD	19.61	0	22.0	ug/kg	112				47	163	
	4,4'-DDT	19.61	0	19.8	ug/kg	101				25	149	
Lab Sample ID:	Q3609-13MS (Column 2)	Client Sample ID:		AU-713-COMP-05MS								
	Aldrin	19.61	0	21.9	ug/kg	112				38	139	
	Dieldrin	19.61	0	22.6	ug/kg	115				32	140	
	4,4'-DDE	19.61	0	22.5	ug/kg	115				20	156	
	4,4'-DDD	19.61	0	23.1	ug/kg	118				47	163	
	4,4'-DDT	19.61	0	22.4	ug/kg	114				25	149	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3614</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	DataFile :	<u>PD091152.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3609-13MSD (Column 1)	Client Sample ID:		AU-713-COMP-05MSD								
	Aldrin	19.62	0	19.5	ug/kg	99		6		38	139	30
	Dieldrin	19.62	0	19.5	ug/kg	99		6		32	140	20
	4,4'-DDE	19.62	0	19.1	ug/kg	97		6		20	156	16
	4,4'-DDD	19.62	0	20.8	ug/kg	106		6		47	163	20
	4,4'-DDT	19.62	0	18.8	ug/kg	96		5		25	149	15
Lab Sample ID:	Q3609-13MSD (Column 2)	Client Sample ID:		AU-713-COMP-05MSD								
	Aldrin	19.62	0	20.6	ug/kg	105		6		38	139	30
	Dieldrin	19.62	0	21.3	ug/kg	109		5		32	140	20
	4,4'-DDE	19.62	0	21.3	ug/kg	109		5		20	156	16
	4,4'-DDD	19.62	0	21.8	ug/kg	111		6		47	163	20
	4,4'-DDT	19.62	0	21.1	ug/kg	108		5		25	149	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3614</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PD091221.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB170527BS (Column 1)	Aldrin	16.65	16.2	ug/kg	97				82	124		
	Dieldrin	16.65	16.3	ug/kg	98				85	121		
	4,4'-DDE	16.65	16.3	ug/kg	98				81	123		
	4,4'-DDD	16.65	16.0	ug/kg	96				80	131		
	4,4'-DDT	16.65	16.5	ug/kg	99				70	129		
PB170527BS (Column 2)	Aldrin	16.65	17.2	ug/kg	103				82	124		
	Dieldrin	16.65	17.1	ug/kg	103				85	121		
	4,4'-DDE	16.65	17.1	ug/kg	103				81	123		
	4,4'-DDD	16.65	16.6	ug/kg	100				80	131		
	4,4'-DDT	16.65	17.6	ug/kg	106				70	129		

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170527BL

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Lab Sample ID: PB170527BL

Lab File ID: PD091144.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 11/13/2025

Date Analyzed (1): 11/13/2025

Date Analyzed (2): 11/13/2025

Time Analyzed (1): 12:17

Time Analyzed (2): 12:17

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
AU-713-COMP-05MS	Q3609-13MS	PD091151.D	11/13/2025	11/13/2025
AU-713-COMP-05MSD	Q3609-13MSD	PD091152.D	11/13/2025	11/13/2025
COMP-1	Q3614-01	PD091153.D	11/13/2025	11/13/2025
COMP-2	Q3614-02	PD091154.D	11/13/2025	11/13/2025
COMP-3	Q3614-03	PD091155.D	11/13/2025	11/13/2025
PB170527BS	PB170527BS	PD091221.D	11/17/2025	11/17/2025

COMMENTS: _____



SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder	Date Collected:	11/11/25
Project:	Logan School	Date Received:	11/12/25
Client Sample ID:	COMP-1	SDG No.:	Q3614
Lab Sample ID:	Q3614-01	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	82.3
Sample Wt/Vol:	30.09 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	PESTICIDE Group1
	Prep Date:		11/13/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.15	U	1	0.15	2.10	ug/kg	11/13/25 14:20	PB170527
60-57-1	Dieldrin	0.17	U	1	0.17	2.10	ug/kg	11/13/25 14:20	PB170527
72-55-9	4,4-DDE	0.17	U	1	0.17	2.10	ug/kg	11/13/25 14:20	PB170527
72-54-8	4,4-DDD	0.18	U	1	0.18	2.10	ug/kg	11/13/25 14:20	PB170527
50-29-3	4,4-DDT	0.17	U	1	0.17	2.10	ug/kg	11/13/25 14:20	PB170527
SURROGATES									
2051-24-3	Decachlorobiphenyl	19.4			10 - 143	97%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.6			10 - 131	98%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091153.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 14:20
 Operator : AR\AJ
 Sample : Q3614-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-1

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:03:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.542	2.872	53086150	585.2E6	17.047m	19.653m
28) SA Decachlor...	9.065	8.059	77000122	586.5E6	16.460	19.363

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091153.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 14:20
Operator : AR\AJ
Sample : Q3614-01
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
COMP-1


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:03:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: -0.003 min
Response: 53086150
Conc: 17.05 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-1


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 585201889
Conc: 19.65 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 77000122
Conc: 16.46 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.002 min
Response: 586516442
Conc: 19.36 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder	Date Collected:	11/11/25
Project:	Logan School	Date Received:	11/12/25
Client Sample ID:	COMP-2	SDG No.:	Q3614
Lab Sample ID:	Q3614-02	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	81.7
Sample Wt/Vol:	30.05 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	PESTICIDE Group1
	Prep Date:		11/13/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.15	U	1	0.15	2.10	ug/kg	11/13/25 14:33	PB170527
60-57-1	Dieldrin	0.17	U	1	0.17	2.10	ug/kg	11/13/25 14:33	PB170527
72-55-9	4,4-DDE	0.17	U	1	0.17	2.10	ug/kg	11/13/25 14:33	PB170527
72-54-8	4,4-DDD	0.18	U	1	0.18	2.10	ug/kg	11/13/25 14:33	PB170527
50-29-3	4,4-DDT	0.17	U	1	0.17	2.10	ug/kg	11/13/25 14:33	PB170527
SURROGATES									
2051-24-3	Decachlorobiphenyl	17.2			10 - 143	86%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	18.7			10 - 131	94%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091154.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 14:33
 Operator : AR\AJ
 Sample : Q3614-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-2

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:04:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	48783270	558.2E6	15.665m	18.744m
28)	SA Decachlor...	9.065	8.059	76628686	520.5E6	16.381	17.182

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 14:33
Operator : AR\AJ
Sample : Q3614-02
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
COMP-2

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:04:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD091154.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.542 min
Delta R.T.: -0.003 min
Response: 48783270
Conc: 15.66 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-2

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025


Response_ Signal: PD091154.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.003 min
Response: 558152711
Conc: 18.74 ng/ml m


Response_ Signal: PD091154.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 76628686
Conc: 16.38 ng/ml

Response_ Signal: PD091154.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.003 min
Response: 520470596
Conc: 17.18 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder	Date Collected:	11/11/25
Project:	Logan School	Date Received:	11/12/25
Client Sample ID:	COMP-3	SDG No.:	Q3614
Lab Sample ID:	Q3614-03	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	83.5
Sample Wt/Vol:	30.03 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	PESTICIDE Group1
	Prep Date:		11/13/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.14	U	1	0.14	2.00	ug/kg	11/13/25 14:47	PB170527
60-57-1	Dieldrin	0.17	U	1	0.17	2.00	ug/kg	11/13/25 14:47	PB170527
72-55-9	4,4-DDE	0.17	U	1	0.17	2.00	ug/kg	11/13/25 14:47	PB170527
72-54-8	4,4-DDD	0.18	U	1	0.18	2.00	ug/kg	11/13/25 14:47	PB170527
50-29-3	4,4-DDT	0.17	U	1	0.17	2.00	ug/kg	11/13/25 14:47	PB170527
SURROGATES									
2051-24-3	Decachlorobiphenyl	18.3			10 - 143	91%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.4			10 - 131	102%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091155.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 14:47
 Operator : AR\AJ
 Sample : Q3614-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-3

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:04:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.542	2.874	54455406	609.0E6	17.486m	20.452
2) SA Decachlor...	9.065	8.060	75166225	554.0E6	16.068	18.290

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 14:47
Operator : AR\AJ
Sample : Q3614-03
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
COMP-3


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:04:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: -0.003 min
Response: 54455406
Conc: 17.49 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-3


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: -0.001 min
Response: 609015422
Conc: 20.45 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 75166225
Conc: 16.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.060 min
Delta R.T.: -0.002 min
Response: 554043616
Conc: 18.29 ng/ml



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>POWE02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3614</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>10/17/2025</u> <u>10/17/2025</u>
		Calibration Times:	<u>11:34</u> <u>12:28</u>

 GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD090650.D</u>	RT 075 = <u>PD090651.D</u>
RT 050 = <u>PD090652.D</u>	RT 025 = <u>PD090653.D</u>	RT 005 = <u>PD090654.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	6.70	6.70	6.70	6.70	6.70	6.70	6.60	6.80
4,4'-DDE	6.19	6.19	6.19	6.19	6.19	6.19	6.09	6.29
4,4'-DDT	7.02	7.02	7.02	7.02	7.02	7.02	6.92	7.12
Aldrin	5.26	5.27	5.26	5.26	5.26	5.26	5.16	5.36
Decachlorobiphenyl	9.07	9.07	9.06	9.07	9.07	9.07	8.97	9.17
Dieldrin	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
Tetrachloro-m-xylene	3.54	3.55	3.55	3.54	3.54	3.54	3.44	3.64

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>POWE02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3614</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>10/17/2025</u> <u>10/17/2025</u>
		Calibration Times:	<u>11:34</u> <u>12:28</u>

 GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD090650.D</u>	RT 075 = <u>PD090651.D</u>
RT 050 = <u>PD090652.D</u>	RT 025 = <u>PD090653.D</u>	RT 005 = <u>PD090654.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	5.92	5.92	5.92	5.92	5.92	5.92	5.82	6.02
4,4'-DDE	5.37	5.37	5.37	5.37	5.37	5.37	5.27	5.47
4,4'-DDT	6.17	6.18	6.17	6.17	6.18	6.17	6.07	6.27
Aldrin	4.36	4.36	4.36	4.36	4.36	4.36	4.26	4.46
Decachlorobiphenyl	8.06	8.06	8.06	8.06	8.06	8.06	7.96	8.16
Dieldrin	5.50	5.50	5.50	5.50	5.50	5.50	5.40	5.60
Tetrachloro-m-xylene	2.87	2.88	2.88	2.87	2.87	2.87	2.77	2.97



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Instrument ID: ECD_D

Calibration Date(s): 10/17/2025 10/17/2025

Calibration Times: 11:34 12:28

GC Column: ZB-MR1

ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4150980000	4075270000	4055470000	3840760000	4148570000	4054210000	3
4,4'-DDE	5435610000	5370750000	5317610000	5053060000	5456160000	5326640000	3
4,4'-DDT	4256810000	4152600000	4117930000	3927880000	4097130000	4110470000	3
Aldrin	6694230000	6562230000	6516940000	6238140000	6724820000	6547270000	3
Decachlorobiphenyl	4361180000	4315270000	4506230000	4707060000	5500070000	4677960000	10
Dieldrin	5929890000	5841940000	5818180000	5550330000	6016400000	5831350000	3
Tetrachloro-m-xylene	3010860000	2951180000	3030100000	3124300000	3454390000	3114160000	6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Instrument ID: ECD_D

Calibration Date(s): 10/17/2025 10/17/2025

Calibration Times: 11:34 12:28

GC Column: ZB-MR2

ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	33708900000	34309400000	35560900000	36237500000	42373300000	36438000000	10
4,4'-DDE	40500800000	41191000000	42548600000	43577200000	51373700000	43838300000	10
4,4'-DDT	33888000000	33982600000	34868200000	34959100000	37667600000	35073100000	4
Aldrin	43083800000	43574800000	45033800000	46172400000	53729700000	46318900000	9
Decachlorobiphenyl	28856800000	28247200000	29308000000	30057200000	34987600000	30291400000	9
Dieldrin	40601100000	41425500000	43096200000	44203200000	51593900000	44184000000	10
Tetrachloro-m-xylene	27903400000	27689100000	28617400000	29726500000	34949400000	29777100000	10



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Instrument ID: _____

Date(s) Analyzed: _____

GC Column: _____

ID: _____ (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
		1				
		2				
		3				
		4				
		5				

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090650.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:34
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:33:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	301.1E6	2790.3E6	99.682	98.737
2)	SA Decachlor...	9.066	8.061	436.1E6	2885.7E6	98.364	99.224
Target Compounds							
2)	A alpha-BHC	3.994	3.386	775.3E6	4911.8E6	103.334	99.427
3)	MA gamma-BHC...	4.324	3.722	718.6E6	4496.5E6	102.762	98.892
4)	MA Heptachlor	4.922	4.075	575.8E6	4127.2E6	101.879	97.673
5)	MB Aldrin	5.264	4.360	669.4E6	4308.4E6	101.342	97.787
6)	B beta-BHC	4.511	4.019	251.5E6	1829.7E6	99.502	97.902
7)	B delta-BHC	4.760	4.255	716.9E6	4532.5E6	102.853	98.756
8)	B Heptachlo...	5.685	4.864	579.6E6	3820.3E6	100.668	96.869
9)	A Endosulfan I	6.068	5.238	545.2E6	3572.9E6	100.030	96.676
10)	B gamma-Chl...	5.940	5.117	597.2E6	4192.1E6	100.988	97.867
11)	B alpha-Chl...	6.020	5.181	596.6E6	4000.4E6	100.230	97.693
12)	B 4,4'-DDE	6.190	5.366	543.6E6	4050.1E6	101.097	97.534
13)	MA Dieldrin	6.341	5.504	593.0E6	4060.1E6	100.951	97.019
14)	MA Endrin	6.568	5.780	503.6E6	3690.2E6	100.979	96.725
15)	B Endosulfa...	6.781	6.072	487.3E6	3427.0E6	99.879	96.941
16)	A 4,4'-DDD	6.700	5.920	415.1E6	3370.9E6	101.164	97.326
17)	MA 4,4'-DDT	7.016	6.174	425.7E6	3388.8E6	101.658	98.574
18)	B Endrin al...	6.910	6.250	350.6E6	2511.1E6	99.070	96.573
19)	B Endosulfa...	7.145	6.474	455.7E6	3296.3E6	100.010	97.059
20)	A Methoxychlor	7.489	6.745	205.0E6	1611.7E6	99.308	96.752
21)	B Endrin ke...	7.625	6.982	482.0E6	3446.0E6	100.390	97.056
22)	Mirex	8.107	7.175	295.5E6	2331.6E6	97.723	97.238



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090650.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:34
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:33:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090651.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:48
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:35:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.875	221.3E6	2076.7E6	73.844	73.982
28)	SA Decachlor...	9.065	8.062	323.6E6	2118.5E6	73.652	73.550
Target Compounds							
2)	A alpha-BHC	3.994	3.387	555.2E6	3629.4E6	74.331	73.972
3)	MA gamma-BHC...	4.325	3.723	517.9E6	3345.4E6	74.376	74.044
4)	MA Heptachlor	4.923	4.076	419.8E6	3135.8E6	74.519	74.471
5)	MB Aldrin	5.265	4.361	492.2E6	3268.1E6	74.671	74.449
6)	B beta-BHC	4.512	4.019	186.7E6	1377.4E6	74.235	74.130
7)	B delta-BHC	4.760	4.255	517.9E6	3391.5E6	74.535	74.261
8)	B Heptachlo...	5.685	4.865	428.1E6	2916.9E6	74.574	74.304
9)	A Endosulfan I	6.068	5.239	405.3E6	2744.0E6	74.575	74.497
10)	B gamma-Chl...	5.940	5.117	442.1E6	3181.7E6	74.831	74.517
11)	B alpha-Chl...	6.021	5.182	442.9E6	3041.3E6	74.605	74.512
12)	B 4,4'-DDE	6.191	5.366	402.8E6	3089.3E6	74.945	74.597
13)	MA Dieldrin	6.341	5.504	438.1E6	3106.9E6	74.726	74.493
14)	MA Endrin	6.568	5.781	372.2E6	2837.1E6	74.757	74.574
15)	B Endosulfa...	6.781	6.072	362.7E6	2622.6E6	74.556	74.457
16)	A 4,4'-DDD	6.700	5.921	305.6E6	2573.2E6	74.659	74.529
17)	MA 4,4'-DDT	7.016	6.175	311.4E6	2548.7E6	74.584	74.422
18)	B Endrin al...	6.910	6.251	263.1E6	1930.5E6	74.563	74.493
19)	B Endosulfa...	7.145	6.474	337.9E6	2515.4E6	74.430	74.374
20)	A Methoxychlor	7.488	6.745	151.2E6	1223.7E6	73.804	73.968
21)	B Endrin ke...	7.625	6.983	355.2E6	2626.5E6	74.325	74.314
22)	Mirex	8.108	7.176	222.5E6	1780.4E6	74.051	74.497



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090651.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:48
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:35:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 12:01
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:30:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.875	151.5E6	1430.9E6	50.000	50.000
28)	SA Decachlor...	9.064	8.062	225.3E6	1465.4E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.994	3.387	362.6E6	2484.2E6	50.000	50.000
3)	MA gamma-BHC...	4.325	3.723	340.0E6	2298.6E6	50.000	50.000
4)	MA Heptachlor	4.923	4.075	277.3E6	2162.0E6	50.000	50.000
5)	MB Aldrin	5.264	4.361	325.8E6	2251.7E6	50.000	50.000
6)	B beta-BHC	4.512	4.019	127.0E6	954.0E6	50.000	50.000
7)	B delta-BHC	4.760	4.255	338.6E6	2323.3E6	50.000	50.000
8)	B Heptachlo...	5.684	4.864	285.9E6	2033.6E6	50.000	50.000
9)	A Endosulfan I	6.068	5.238	272.5E6	1909.3E6	50.000	50.000
10)	B gamma-Chl...	5.940	5.117	292.8E6	2187.4E6	50.000	50.000
11)	B alpha-Chl...	6.021	5.181	296.9E6	2094.7E6	50.000	50.000
12)	B 4,4'-DDE	6.190	5.366	265.9E6	2127.4E6	50.000	50.000
13)	MA Dieldrin	6.340	5.504	290.9E6	2154.8E6	50.000	50.000
14)	MA Endrin	6.568	5.781	246.9E6	1970.0E6	50.000	50.000
15)	B Endosulfa...	6.781	6.072	244.3E6	1821.6E6	50.000	50.000
16)	A 4,4'-DDD	6.700	5.921	202.8E6	1778.0E6	50.000	50.000
17)	MA 4,4'-DDT	7.015	6.174	205.9E6	1743.4E6	50.000	50.000
18)	B Endrin al...	6.910	6.250	178.6E6	1344.7E6	50.000	50.000
19)	B Endosulfa...	7.145	6.474	227.8E6	1748.0E6	50.000	50.000
20)	A Methoxychlor	7.488	6.745	103.9E6	859.9E6	50.000	50.000
21)	B Endrin ke...	7.625	6.983	239.1E6	1827.5E6	50.000	50.000
22)	Mirex	8.107	7.175	154.6E6	1232.0E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090652.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:01
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:30:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090653.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 12:15
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:38:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	78107465	743.2E6	25.786	26.090
28)	SA Decachlor...	9.065	8.061	117.7E6	751.4E6	26.311	25.807
Target Compounds							
2)	A alpha-BHC	3.993	3.386	168.6E6	1262.5E6	23.132	25.545
3)	MA gamma-BHC...	4.324	3.722	160.4E6	1170.2E6	23.501	25.669
4)	MA Heptachlor	4.922	4.074	132.2E6	1112.6E6	23.835	26.052
5)	MB Aldrin	5.263	4.360	156.0E6	1154.3E6	23.982	25.959
6)	B beta-BHC	4.511	4.018	64015475	495.1E6	25.339	26.214
7)	B delta-BHC	4.759	4.255	158.6E6	1178.9E6	23.327	25.605
8)	B Heptachlo...	5.684	4.864	139.6E6	1059.4E6	24.488	26.460
9)	A Endosulfan I	6.067	5.238	133.7E6	987.0E6	24.700	26.323
10)	B gamma-Chl...	5.939	5.116	139.8E6	1108.7E6	23.988	25.717
11)	B alpha-Chl...	6.020	5.181	154.3E6	1067.7E6	25.731	25.860
12)	B 4,4'-DDE	6.190	5.365	126.3E6	1089.4E6	23.861	25.967
13)	MA Dieldrin	6.340	5.503	138.8E6	1105.1E6	23.985	26.105
14)	MA Endrin	6.567	5.780	117.8E6	1017.3E6	23.978	26.284
15)	B Endosulfa...	6.781	6.071	119.7E6	938.0E6	24.710	26.203
16)	A 4,4'-DDD	6.700	5.920	96019034	905.9E6	23.822	25.918
17)	MA 4,4'-DDT	7.016	6.173	98197092	874.0E6	23.870	25.388
18)	B Endrin al...	6.909	6.250	89492653	694.2E6	25.271	26.317
19)	B Endosulfa...	7.144	6.473	112.0E6	896.5E6	24.754	26.114
20)	A Methoxychlor	7.488	6.744	52364279	443.4E6	25.424	26.326
21)	B Endrin ke...	7.625	6.982	117.4E6	945.4E6	24.668	26.289
22)	Mirex	8.108	7.175	79715192	640.8E6	26.131	26.336



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:15
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:38:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090654.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 12:28
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/19/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:41:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.874	17271955	174.7E6	5.550m	5.868
28) SA Decachlor...	9.066	8.062	27500355	174.9E6	5.879	5.775
Target Compounds						
2) A alpha-BHC	3.993	3.386	33677596	288.9E6	4.692	5.655
3) MA gamma-BHC...	4.324	3.723	33423739	269.5E6	4.916	5.703
4) MA Heptachlor	4.922	4.075	28789856	260.9E6	5.151	5.850
5) MB Aldrin	5.264	4.360	33624118	268.6E6	5.136	5.800
6) B beta-BHC	4.512	4.019	14963256	118.0E6	5.712	5.951
7) B delta-BHC	4.760	4.255	33098385	272.1E6	4.895	5.702
8) B Heptachlo...	5.684	4.864	30689599	265.3E6	5.301	6.280
9) A Endosulfan I	6.067	5.238	30143871	234.9E6	5.444	5.962
10) B gamma-Chl...	5.940	5.117	31705442	261.5E6	5.346	5.818
11) B alpha-Chl...	6.019	5.181	34603140	252.7E6	5.576m	5.857
12) B 4,4'-DDE	6.190	5.366	27280775	256.9E6	5.122	5.859
13) MA Dieldrin	6.341	5.504	30082015	258.0E6	5.159	5.839
14) MA Endrin	6.568	5.781	25601278	238.2E6	5.168	5.883
15) B Endosulfa...	6.781	6.072	28078989	222.9E6	5.616	5.936
16) A 4,4'-DDD	6.700	5.921	20742842	211.9E6	5.116	5.814
17) MA 4,4'-DDT	7.016	6.175	20485670	188.3E6	4.984	5.370
18) B Endrin al...	6.910	6.250	20741409	164.8E6	5.663	5.950
19) B Endosulfa...	7.144	6.474	25464611	209.5E6	5.490	5.845
20) A Methoxychlor	7.489	6.745	11946535	98858490	5.620	5.672
21) B Endrin ke...	7.624	6.983	26302012	223.7E6	5.414m	5.931
22) Mirex	8.108	7.175	19009608	155.6E6	5.939	6.058

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090654.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:28
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC005

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:41:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090657.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 13:09
 Operator : AR\AJ
 Sample : PCHLORICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:49:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:48:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.874	108.8E6	1215.0E6	50.000	50.000
28)	SA Decachlor...	9.066	8.062	167.8E6	1105.6E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.708	3.898	92608674	605.2E6	500.000	500.000
24)	Chlordane-2	5.235	4.479	92865500	631.9E6	500.000	500.000
25)	Chlordane-3	5.941	5.117	379.1E6	1885.9E6	500.000	500.000
26)	Chlordane-4	6.026	5.181	458.8E6	1625.9E6	500.000	500.000
27)	Chlordane-5	6.865	6.081	74238996	730.3E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090657.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 13:09
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:49:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090662.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 14:17
 Operator : AR\AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:36:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:35:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.875	149.0E6	1451.5E6	50.000	50.000
7)	SA Decachlor...	9.066	8.062	225.6E6	1489.1E6	50.000	50.000
Target Compounds							
2)	Toxaphene-1	6.236	5.466	16282219	131.9E6	500.000	500.000
3)	Toxaphene-2	6.436	5.638	24310877	88280081	500.000	500.000
4)	Toxaphene-3	7.143	6.745	42855293	329.4E6	500.000	500.000
5)	Toxaphene-4	7.558	7.189	47868684	247.6E6	500.000	500.000
6)	Toxaphene-5	7.923	7.320	30240231	123.0E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090662.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 14:17
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:36:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:35:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090665.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 16:46
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD101725

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 17:27:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	153.7E6	1464.9E6	49.347	49.195
28)	SA Decachlor...	9.065	8.062	232.3E6	1556.2E6	49.651	51.373
Target Compounds							
2)	A alpha-BHC	3.993	3.386	368.9E6	2547.7E6	51.395	49.860
3)	MA gamma-BHC...	4.325	3.722	346.1E6	2351.4E6	50.913	49.766
4)	MA Heptachlor	4.923	4.074	269.7E6	2210.7E6	48.243	49.564
5)	MB Aldrin	5.264	4.360	332.4E6	2302.8E6	50.770	49.716
6)	B beta-BHC	4.512	4.018	128.8E6	974.3E6	49.166	49.134
7)	B delta-BHC	4.760	4.255	343.7E6	2372.2E6	50.832	49.714
8)	B Heptachlo...	5.684	4.864	292.1E6	2073.7E6	50.455	48.633
9)	A Endosulfan I	6.068	5.238	276.7E6	1960.5E6	49.982	49.772
10)	B gamma-Chl...	5.940	5.117	297.4E6	2236.0E6	50.146	49.745
11)	B alpha-Chl...	6.020	5.181	299.6E6	2144.4E6	48.473	49.711
12)	B 4,4'-DDE	6.190	5.366	269.1E6	2179.5E6	50.526	49.716
13)	MA Dieldrin	6.341	5.503	295.7E6	2209.3E6	50.705	50.003
14)	MA Endrin	6.568	5.780	251.8E6	2041.8E6	50.824	50.421
15)	B Endosulfa...	6.781	6.072	251.4E6	1876.1E6	50.275	49.956
16)	A 4,4'-DDD	6.701	5.920	206.7E6	1821.0E6	50.990	49.977
17)	MA 4,4'-DDT	7.016	6.174	206.8E6	1798.7E6	50.320	51.285
18)	B Endrin al...	6.910	6.250	184.2E6	1401.3E6	50.283	50.600
19)	B Endosulfa...	7.144	6.474	231.2E6	1809.0E6	49.837	50.467
20)	A Methoxychlor	7.489	6.745	104.2E6	890.1E6	48.999	51.076
21)	B Endrin ke...	7.625	6.983	242.9E6	1911.1E6	49.982	50.668
22)	Mirex	8.108	7.176	158.1E6	1294.5E6	49.394	50.389



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090665.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 16:46
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 17:27:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090666.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 17:50
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD101725CHLOR

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 18:03:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.873	151.2E6	1640.5E6	48.564	55.094
28)	SA Decachlor...	9.069	8.063	245.8E6	1635.1E6	52.541	53.979
Target Compounds							
23)	Chlordane-1	4.711	3.897	81712455	547.0E6	432.404	438.923
24)	Chlordane-2	5.238	4.479	83577133	567.3E6	432.751	434.873
25)	Chlordane-3	5.944	5.117	339.4E6	1711.3E6	444.973	439.739
26)	Chlordane-4	6.029	5.182	443.9E6	1471.1E6	458.160	440.289
27)	Chlordane-5	6.868	6.082	67446272	668.4E6	448.868	452.832



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090666.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 17:50
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD101725CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 18:03:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090667.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 18:42
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD101725TOX

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/21/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 09:07:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 16:43:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.874	144.8E6	1414.7E6	47.850	47.865
7)	SA Decachlor...	9.073	8.063	222.4E6	1485.6E6	48.104	48.572
Target Compounds							
2)	Toxaphene-1	6.243	5.468	16199211	128.4E6	486.326	472.241
3)	Toxaphene-2	6.442	5.639	23160349	87664602	469.047	488.530
4)	Toxaphene-3	7.149	6.748	41739789	339.6E6	481.161	526.232
5)	Toxaphene-4	7.565	7.190	46882437	247.3E6	496.038	505.430
6)	Toxaphene-5	7.930	7.320	29219348	133.4E6	485.054	589.259m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090667.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 18:42
Operator : AR\AJ
Sample : PTOXICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 09:07:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 16:43:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Miza 10/21/2025
Supervised By: mohammad ahmed 10/24/2025

Instrument :
ECD_D
ClientSample :
ICVPD101725TOX

Calibration Times: **20:56** **21:51**

LAB FILE ID:	RT 100 =	<u>PD091196.D</u>	RT 075 =	<u>PD091197.D</u>
	RT 050 =	PD091198.D	RT 025 =	PD091199.D
			RT 005 =	PD091200.D

[illegible]

Calibration Times: **20:56** **21:51**

LAB FILE ID:	RT 100 =	<u>PD091196.D</u>	RT 075 =	<u>PD091197.D</u>
	RT 050 =	PD091198.D	RT 025 =	PD091199.D
			RT 005 =	PD091200.D

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Instrument ID: ECD_D

Calibration Date(s): 11/14/2025 11/14/2025

Calibration Times: 20:56 21:51

GC Column: ZB-MR1

ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091196.D</u>	CF 075 =	<u>PD091197.D</u>		
CF 050 =		<u>PD091198.D</u>	CF 025 =	<u>PD091199.D</u>	CF 005 =	<u>PD091200.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4592720000	4750410000	4706670000	4762630000	5745110000	4911510000	10
4,4'-DDE	5537410000	5724420000	5594750000	5596930000	6543500000	5799400000	7
4,4'-DDT	4033900000	4103640000	4023370000	3973800000	4542840000	4135510000	6
Aldrin	6913630000	7108360000	6994820000	7033330000	8231840000	7256400000	8
Decachlorobiphenyl	4266550000	4520270000	4633000000	4951710000	6537810000	4981870000	18
Dieldrin	6112350000	6323390000	6220240000	6264150000	7404460000	6464920000	8
Tetrachloro-m-xylene	3100550000	3224630000	3262160000	3454950000	4200860000	3448630000	13



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: POWE02
SDG NO.: Q3614

Calibration Date(s): 11/14/2025 11/14/2025
Calibration Times: 20:56 21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091196.D</u>	CF 075 =	<u>PD091197.D</u>		
CF 050 =		<u>PD091198.D</u>	CF 025 =	<u>PD091199.D</u>	CF 005 =	<u>PD091200.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	42169800000	44652200000	46098100000	48914200000	62508300000	48868500000	16
4,4'-DDE	46784900000	49153600000	50413800000	53041500000	67816900000	53442100000	16
4,4'-DDT	36565400000	37844700000	38225900000	38748200000	44278500000	39132500000	8
Aldrin	49756900000	52224800000	53642000000	56183400000	70485000000	56458400000	14
Decachlorobiphenyl	35391500000	36530600000	37125400000	38755300000	50071600000	39574900000	15
Dieldrin	47342900000	49974500000	51581700000	54390600000	68265900000	54311100000	15
Tetrachloro-m-xylene	32274400000	33875900000	34577100000	36418900000	46004800000	36630200000	15



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Instrument ID: _____

Date(s) Analyzed: _____

GC Column: _____

ID: _____ (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
		1				
		2				
		3				
		4				
		5				

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 20:56
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:13:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	310.1E6	3227.4E6	95.046	93.341
28)	SA Decachlor...	9.063	8.059	426.7E6	3539.1E6	92.090	95.330
Target Compounds							
2)	A alpha-BHC	3.993	3.385	799.1E6	5708.1E6	101.539	95.126
3)	MA gamma-BHC...	4.324	3.721	735.0E6	5203.7E6	100.431	94.689
4)	MA Heptachlor	4.921	4.073	687.3E6	4845.3E6	99.281	92.571
5)	MB Aldrin	5.262	4.358	691.4E6	4975.7E6	98.839	92.757
6)	B beta-BHC	4.511	4.017	258.4E6	2085.6E6	94.961	92.161
7)	B delta-BHC	4.759	4.253	734.3E6	5195.2E6	100.875	93.888
8)	B Heptachlo...	5.683	4.862	604.6E6	4379.5E6	97.550	91.159
9)	A Endosulfan I	6.066	5.236	562.2E6	4121.9E6	96.781	90.923
10)	B gamma-Chl...	5.938	5.114	612.3E6	4854.2E6	98.587	93.402
11)	B alpha-Chl...	6.019	5.179	616.9E6	4616.2E6	98.511	92.892
12)	B 4,4'-DDE	6.189	5.363	553.7E6	4678.5E6	98.975	92.802
13)	MA Dieldrin	6.339	5.501	611.2E6	4734.3E6	98.265	91.782
14)	MA Endrin	6.566	5.778	476.3E6	3972.3E6	99.746	92.358
15)	B Endosulfa...	6.779	6.069	502.2E6	4023.4E6	95.589	91.118
16)	A 4,4'-DDD	6.699	5.918	459.3E6	4217.0E6	97.579	91.478
17)	MA 4,4'-DDT	7.014	6.171	403.4E6	3656.5E6	100.262	95.656
18)	B Endrin al...	6.908	6.248	364.3E6	2990.5E6	93.829	90.185
19)	B Endosulfa...	7.143	6.471	463.5E6	3880.7E6	95.233	91.193
20)	A Methoxychlor	7.487	6.742	196.8E6	1787.5E6	96.081	92.289
21)	B Endrin ke...	7.624	6.980	513.1E6	4374.3E6	96.367	90.729
22)	Mirex	8.106	7.172	302.9E6	2859.2E6	92.228	91.815



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:56
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:13:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091197.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 21:10
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:14:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	241.8E6	2540.7E6	74.137	73.479
28)	SA Decachlor...	9.064	8.059	339.0E6	2739.8E6	73.175	73.798
Target Compounds							
2)	A alpha-BHC	3.993	3.385	609.1E6	4455.1E6	77.386	74.245
3)	MA gamma-BHC...	4.324	3.721	561.9E6	4062.0E6	76.788	73.915
4)	MA Heptachlor	4.921	4.073	528.9E6	3832.2E6	76.395	73.217
5)	MB Aldrin	5.263	4.358	533.1E6	3916.9E6	76.217	73.019
6)	B beta-BHC	4.511	4.017	202.6E6	1647.3E6	74.433	72.793
7)	B delta-BHC	4.760	4.253	560.8E6	4089.0E6	77.046	73.898
8)	B Heptachlo...	5.683	4.862	464.9E6	3475.8E6	75.011	72.350
9)	A Endosulfan I	6.067	5.236	438.5E6	3285.7E6	75.488	72.478
10)	B gamma-Chl...	5.939	5.114	475.7E6	3811.6E6	76.593	73.343
11)	B alpha-Chl...	6.019	5.179	474.0E6	3637.2E6	75.698	73.191
12)	B 4,4'-DDE	6.189	5.364	429.3E6	3686.5E6	76.738	73.125
13)	MA Dieldrin	6.339	5.501	474.3E6	3748.1E6	76.244	72.663
14)	MA Endrin	6.567	5.778	364.6E6	3131.1E6	76.360	72.799
15)	B Endosulfa...	6.780	6.070	391.6E6	3197.3E6	74.544	72.410
16)	A 4,4'-DDD	6.699	5.918	356.3E6	3348.9E6	75.697	72.648
17)	MA 4,4'-DDT	7.015	6.172	307.8E6	2838.4E6	76.496	74.252
18)	B Endrin al...	6.909	6.248	286.5E6	2380.4E6	73.801	71.786
19)	B Endosulfa...	7.144	6.472	361.8E6	3074.4E6	74.333	72.245
20)	A Methoxychlor	7.488	6.742	152.5E6	1407.3E6	74.482	72.657
21)	B Endrin ke...	7.625	6.980	398.8E6	3478.6E6	74.898	72.150
22)	Mirex	8.107	7.173	239.3E6	2256.0E6	72.881	72.445



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091197.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 21:10
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:14:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 21:24
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:14:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	163.1E6	1728.9E6	50.000	50.000
28)	SA Decachlor...	9.063	8.058	231.7E6	1856.3E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.993	3.386	393.5E6	3000.3E6	50.000	50.000
3)	MA gamma-BHC...	4.324	3.721	365.9E6	2747.8E6	50.000	50.000
4)	MA Heptachlor	4.921	4.073	346.1E6	2617.1E6	50.000	50.000
5)	MB Aldrin	5.263	4.358	349.7E6	2682.1E6	50.000	50.000
6)	B beta-BHC	4.511	4.018	136.1E6	1131.5E6	50.000	50.000
7)	B delta-BHC	4.760	4.253	364.0E6	2766.7E6	50.000	50.000
8)	B Heptachlo...	5.682	4.862	309.9E6	2402.1E6	50.000	50.000
9)	A Endosulfan I	6.066	5.235	290.5E6	2266.7E6	50.000	50.000
10)	B gamma-Chl...	5.938	5.114	310.6E6	2598.5E6	50.000	50.000
11)	B alpha-Chl...	6.019	5.179	313.1E6	2484.7E6	50.000	50.000
12)	B 4,4'-DDE	6.189	5.363	279.7E6	2520.7E6	50.000	50.000
13)	MA Dieldrin	6.339	5.501	311.0E6	2579.1E6	50.000	50.000
14)	MA Endrin	6.566	5.778	238.8E6	2150.5E6	50.000	50.000
15)	B Endosulfa...	6.779	6.069	262.7E6	2207.8E6	50.000	50.000
16)	A 4,4'-DDD	6.699	5.917	235.3E6	2304.9E6	50.000	50.000
17)	MA 4,4'-DDT	7.014	6.171	201.2E6	1911.3E6	50.000	50.000
18)	B Endrin al...	6.908	6.248	194.1E6	1658.0E6	50.000	50.000
19)	B Endosulfa...	7.143	6.471	243.4E6	2127.8E6	50.000	50.000
20)	A Methoxychlor	7.487	6.741	102.4E6	968.4E6	50.000	50.000
21)	B Endrin ke...	7.624	6.980	266.2E6	2410.7E6	50.000	50.000
22)	Mirex	8.106	7.173	164.2E6	1557.0E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 21:24
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:14:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091199.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 21:37
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:15:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	86373657	910.5E6	26.477	26.332
28)	SA Decachlor...	9.065	8.060	123.8E6	968.9E6	26.720	26.098
Target Compounds							
2)	A alpha-BHC	3.993	3.385	190.3E6	1552.3E6	24.183	25.870
3)	MA gamma-BHC...	4.324	3.721	179.1E6	1424.8E6	24.470	25.927
4)	MA Heptachlor	4.921	4.073	172.6E6	1376.5E6	24.928	26.299
5)	MB Aldrin	5.263	4.358	175.8E6	1404.6E6	25.138	26.184
6)	B beta-BHC	4.511	4.018	71261837	597.0E6	26.184	26.382
7)	B delta-BHC	4.759	4.253	177.1E6	1438.1E6	24.324	25.989
8)	B Heptachlo...	5.683	4.862	158.2E6	1272.1E6	25.525	26.479
9)	A Endosulfan I	6.066	5.236	149.8E6	1202.9E6	25.783	26.533
10)	B gamma-Chl...	5.938	5.115	157.8E6	1355.6E6	25.412	26.083
11)	B alpha-Chl...	6.019	5.179	163.3E6	1301.1E6	26.071	26.181
12)	B 4,4'-DDE	6.189	5.363	139.9E6	1326.0E6	25.010	26.303
13)	MA Dieldrin	6.340	5.502	156.6E6	1359.8E6	25.176	26.361
14)	MA Endrin	6.567	5.778	119.1E6	1124.9E6	24.935	26.153
15)	B Endosulfa...	6.780	6.070	135.5E6	1168.5E6	25.790	26.463
16)	A 4,4'-DDD	6.699	5.918	119.1E6	1222.9E6	25.297	26.527
17)	MA 4,4'-DDT	7.015	6.172	99345097	968.7E6	24.692	25.342
18)	B Endrin al...	6.909	6.248	102.2E6	882.3E6	26.325	26.609
19)	B Endosulfa...	7.144	6.472	125.6E6	1118.0E6	25.803	26.271
20)	A Methoxychlor	7.488	6.743	53585924	505.3E6	26.164	26.091
21)	B Endrin ke...	7.624	6.981	137.7E6	1278.6E6	25.857	26.519
22)	Mirex	8.106	7.173	88650675	829.2E6	26.994	26.629



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091199.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 21:37
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:15:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091200.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 21:51
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC005

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/17/2025
 Supervised By : Sohil Jodhani 11/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:15:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	21004307	230.0E6	6.439m	6.652m
28)	SA Decachlor...	9.063	8.059	32689037	250.4E6	7.056	6.744
Target Compounds							
2)	A alpha-BHC	3.992	3.385	41908732	387.2E6	5.325	6.452
3)	MA gamma-BHC...	4.323	3.721	40207275	354.8E6	5.494	6.456
4)	MA Heptachlor	4.919	4.073	40291748	345.5E6	5.820m	6.600
5)	MB Aldrin	5.262	4.358	41159200	352.4E6	5.884	6.570
6)	B beta-BHC	4.511	4.017	18349309	154.0E6	6.742	6.806
7)	B delta-BHC	4.757	4.253	39887495	365.3E6	5.480m	6.601
8)	B Heptachlo...	5.681	4.862	38568147	328.6E6	6.223m	6.841
9)	A Endosulfan I	6.067	5.236	37227156	303.9E6	6.408	6.704
10)	B gamma-Chl...	5.939	5.114	37561958	337.9E6	6.047	6.502
11)	B alpha-Chl...	6.019	5.179	45659405	326.3E6	7.291	6.566
12)	B 4,4'-DDE	6.188	5.363	32717510	339.1E6	5.848	6.726
13)	MA Dieldrin	6.339	5.502	37022298	341.3E6	5.952	6.617
14)	MA Endrin	6.566	5.778	28273962	282.4E6	5.921	6.567
15)	B Endosulfa...	6.780	6.070	34393712	298.2E6	6.546	6.753
16)	A 4,4'-DDD	6.699	5.918	28725548	312.5E6	6.103	6.780
17)	MA 4,4'-DDT	7.014	6.172	22714196	221.4E6	5.646	5.792
18)	B Endrin al...	6.908	6.248	26987398	231.8E6	6.952	6.989
19)	B Endosulfa...	7.143	6.471	31867215	285.5E6	6.547	6.709
20)	A Methoxychlor	7.487	6.742	13392299	120.2E6	6.539	6.205
21)	B Endrin ke...	7.624	6.980	34435592	326.5E6	6.467	6.772
22)	Mirex	8.105	7.173	23674832	213.6E6	7.209	6.859

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 21:51
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :

ECD_D

Client Sample Id :

PSTDICC005

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 11/17/2025

Supervised By :Sohil Jodhani 11/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:15:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091203.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 22:32
 Operator : AR\AJ
 Sample : PCHLORICC500
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:29:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:28:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.873	143.9E6	1902.7E6	50.000	50.000
28)	SA Decachlor...	9.063	8.058	209.3E6	1703.5E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.707	3.896	113.3E6	788.6E6	500.000	500.000
24)	Chlordane-2	5.233	4.476	110.4E6	798.4E6	500.000	500.000
25)	Chlordane-3	5.939	5.114	448.7E6	2625.7E6	500.000	500.000
26)	Chlordane-4	6.024	5.178	539.1E6	2177.3E6	500.000	500.000
27)	Chlordane-5	6.864	6.078	88781325	938.4E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 22:32
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:29:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:28:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091208.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 23:40
 Operator : AR\AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 05:12:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 05:11:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.874	163.5E6	1762.4E6	50.000	50.000
7)	SA Decachlor...	9.063	8.058	236.5E6	1944.6E6	50.000	50.000
Target Compounds							
2)	Toxaphene-1	6.234	5.463	21406071	174.2E6	500.000	500.000
3)	Toxaphene-2	6.435	5.635	34489739	115.5E6	500.000	500.000
4)	Toxaphene-3	7.141	6.743	56311870	545.5E6	500.000	500.000
5)	Toxaphene-4	7.556	7.186	65356075	352.5E6	500.000	500.000
6)	Toxaphene-5	7.922	7.317	39645850	221.3E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091208.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 23:40
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 66 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 05:12:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 05:11:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091211.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2025 00:21
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD111525

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:49:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	162.5E6	1718.4E6	47.132	46.911
28)	SA Decachlor...	9.063	8.058	232.3E6	1905.9E6	46.634	48.160
Target Compounds							
2)	A alpha-BHC	3.993	3.385	390.8E6	2984.9E6	48.872	47.228
3)	MA gamma-BHC...	4.324	3.721	361.9E6	2728.5E6	48.435	47.188
4)	MA Heptachlor	4.921	4.073	342.0E6	2581.9E6	47.755	46.765
5)	MB Aldrin	5.262	4.358	348.3E6	2666.1E6	47.995	47.222
6)	B beta-BHC	4.511	4.017	135.0E6	1124.4E6	46.461	46.799
7)	B delta-BHC	4.760	4.253	363.6E6	2741.9E6	48.917	46.888
8)	B Heptachlo...	5.683	4.862	313.5E6	2381.7E6	48.259	46.737
9)	A Endosulfan I	6.067	5.235	288.8E6	2250.9E6	47.010	47.039
10)	B gamma-Chl...	5.938	5.114	308.9E6	2585.4E6	47.516	47.328
11)	B alpha-Chl...	6.019	5.179	316.9E6	2460.3E6	46.047	47.013
12)	B 4,4'-DDE	6.188	5.363	276.9E6	2496.0E6	47.745	46.704
13)	MA Dieldrin	6.339	5.501	309.0E6	2552.1E6	47.803	46.991
14)	MA Endrin	6.566	5.778	234.4E6	2114.9E6	47.229	46.798
15)	B Endosulfa...	6.779	6.070	260.8E6	2189.8E6	46.913	46.910
16)	A 4,4'-DDD	6.698	5.917	235.1E6	2292.3E6	47.858	46.907
17)	MA 4,4'-DDT	7.014	6.171	197.9E6	1882.1E6	47.844	48.095
18)	B Endrin al...	6.908	6.248	193.5E6	1652.0E6	46.452	46.813
19)	B Endosulfa...	7.143	6.471	241.2E6	2113.4E6	46.888	47.137
20)	A Methoxychlor	7.487	6.742	101.6E6	965.3E6	46.720	48.141
21)	B Endrin ke...	7.623	6.980	268.5E6	2432.9E6	47.664	47.746
22)	Mirex	8.105	7.172	165.8E6	1568.2E6	46.599	47.322



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091211.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2025 00:21
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD111525

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:49:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091212.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2025 00:49
 Operator : AR\AJ
 Sample : PCHLORICV050
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD111525CHLOR

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:41:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:40:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	144.0E6	1897.0E6	45.917	45.899
28)	SA Decachlor...	9.063	8.059	208.1E6	1722.5E6	44.473	45.857
Target Compounds							
23)	Chlordane-1	4.707	3.896	114.2E6	783.3E6	464.144	455.309
24)	Chlordane-2	5.234	4.477	110.1E6	794.6E6	445.736	453.786
25)	Chlordane-3	5.939	5.115	447.1E6	2621.5E6	463.407	455.626
26)	Chlordane-4	6.024	5.179	540.3E6	2175.2E6	449.814	454.716
27)	Chlordane-5	6.863	6.079	88911954	936.5E6	450.023	461.495



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2025 00:49
Operator : AR\AJ
Sample : PCHLORICV050
Misc :
ALS Vial : 71 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD111525CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:41:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:40:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091213.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2025 01:30
 Operator : AR\AJ
 Sample : PTOXICV050
 Misc :
 ALS Vial : 74 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD111525TOX

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 05:21:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 05:17:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	163.1E6	1761.6E6	49.939	50.116
7) SA Decachlor...	9.063	8.059	235.5E6	1961.7E6	49.446	50.407
Target Compounds						
2) Toxaphene-1	6.234	5.464	21366491	177.4E6	506.837	508.980
3) Toxaphene-2	6.435	5.635	34816331	114.1E6	493.754	505.708
4) Toxaphene-3	7.141	6.744	56004022	537.8E6	503.381	513.961
5) Toxaphene-4	7.557	7.186	66555590	352.7E6	512.569	521.193
6) Toxaphene-5	7.923	7.317	39466192	222.5E6	507.805	531.712

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091213.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2025 01:30
Operator : AR\AJ
Sample : PTOXICV050
Misc :
ALS Vial : 74 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 05:21:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 05:17:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/17/2025 10/17/2025

Continuing Calib Time: 09:47

Initial Calibration Time(s): 11:34 12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
Aldrin	5.26	5.26	5.16	5.36	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
4,4'-DDT	7.01	7.02	6.92	7.12	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
Continuing Calib Date: 11/13/2025 **Initial Calibration Date(s):** 10/17/2025 10/17/2025
Continuing Calib Time: 09:47 **Initial Calibration Time(s):** 11:34 12:28

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
Aldrin	4.36	4.36	4.26	4.46	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.37	5.27	5.47	0.01
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3614
 GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No.: CCAL01 Date Analyzed: 11/13/2025
 Lab Sample No.: PSTDCCC050 Data File : PD091143.D Time Analyzed: 09:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.699	6.600	6.800	52.030	50.000	4.1
4,4'-DDE	6.189	6.090	6.290	45.810	50.000	-8.4
4,4'-DDT	7.014	6.915	7.115	42.630	50.000	-14.7
Aldrin	5.263	5.164	5.364	47.360	50.000	-5.3
Decachlorobiphenyl	9.064	8.964	9.164	44.110	50.000	-11.8
Dieldrin	6.340	6.240	6.440	46.880	50.000	-6.2
Tetrachloro-m-xylene	3.544	3.445	3.645	47.340	50.000	-5.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091143.D **Time Analyzed:** 09:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.918	5.821	6.021	56.260	50.000	12.5
4,4'-DDE	5.363	5.266	5.466	52.520	50.000	5.0
4,4'-DDT	6.171	6.074	6.274	48.030	50.000	-3.9
Aldrin	4.358	4.261	4.461	52.420	50.000	4.8
Decachlorobiphenyl	8.059	7.962	8.162	54.170	50.000	8.3
Dieldrin	5.501	5.404	5.604	52.360	50.000	4.7
Tetrachloro-m-xylene	2.874	2.775	2.975	52.510	50.000	5.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091143.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 09:47
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:02:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	147.4E6	1563.6E6	47.341	52.511
28)	SA Decachlor...	9.064	8.059	206.3E6	1640.8E6	44.109	54.166
Target Compounds							
2)	A alpha-BHC	3.993	3.385	351.8E6	2706.2E6	49.011	52.964
3)	MA gamma-BHC...	4.324	3.721	325.7E6	2476.6E6	47.906	52.416
4)	MA Heptachlor	4.922	4.073	311.9E6	2360.4E6	55.800	52.920
5)	MB Aldrin	5.263	4.358	310.1E6	2428.0E6	47.362	52.420
6)	B beta-BHC	4.512	4.017	122.2E6	1018.1E6	46.633	51.343
7)	B delta-BHC	4.760	4.253	320.8E6	2493.2E6	47.440	52.249
8)	B Heptachlo...	5.683	4.862	271.8E6	2156.8E6	46.958	50.583
9)	A Endosulfan I	6.067	5.236	257.4E6	1969.0E6	46.488	49.987
10)	B gamma-Chl...	5.939	5.114	274.6E6	2326.9E6	46.298	51.768
11)	B alpha-Chl...	6.019	5.179	277.1E6	2227.2E6	44.828	51.631
12)	B 4,4'-DDE	6.189	5.363	244.0E6	2302.5E6	45.806	52.524
13)	MA Dieldrin	6.340	5.501	273.3E6	2313.6E6	46.875	52.363
14)	MA Endrin	6.567	5.778	216.8E6	2025.4E6	43.768	50.017
15)	B Endosulfa...	6.780	6.069	228.5E6	1961.9E6	45.699	52.239
16)	A 4,4'-DDD	6.699	5.918	210.9E6	2050.2E6	52.030	56.264
17)	MA 4,4'-DDT	7.014	6.171	175.2E6	1684.5E6	42.628	48.028
18)	B Endrin al...	6.909	6.248	169.6E6	1458.6E6	46.308	52.672
19)	B Endosulfa...	7.144	6.471	211.8E6	1883.9E6	45.674	52.558
20)	A Methoxychlor	7.488	6.742	92525728	862.8E6	43.530	49.507
21)	B Endrin ke...	7.625	6.980	236.2E6	2101.7E6	48.616	55.722
22)	Mirex	8.107	7.173	146.5E6	1391.2E6	45.755	54.152



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091143.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 09:47
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:02:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 16:32

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
Aldrin	5.26	5.26	5.16	5.36	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
Continuing Calib Date: 11/13/2025 **Initial Calibration Date(s):** 10/17/2025 10/17/2025
Continuing Calib Time: 16:32 **Initial Calibration Time(s):** 11:34 12:28

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
Aldrin	4.36	4.36	4.26	4.46	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.37	5.27	5.47	0.01
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091161.D **Time Analyzed:** 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	50.860	50.000	1.7
4,4'-DDE	6.190	6.090	6.290	44.260	50.000	-11.5
4,4'-DDT	7.015	6.915	7.115	41.060	50.000	-17.9
Aldrin	5.263	5.164	5.364	46.500	50.000	-7.0
Decachlorobiphenyl	9.066	8.964	9.164	44.150	50.000	-11.7
Dieldrin	6.340	6.240	6.440	45.330	50.000	-9.3
Tetrachloro-m-xylene	3.544	3.445	3.645	47.040	50.000	-5.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3614
 GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No.: CCAL02 Date Analyzed: 11/13/2025
 Lab Sample No.: PSTDCCC050 Data File : PD091161.D Time Analyzed: 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.919	5.821	6.021	53.110	50.000	6.2
4,4'-DDE	5.363	5.266	5.466	49.670	50.000	-0.7
4,4'-DDT	6.172	6.074	6.274	44.260	50.000	-11.5
Aldrin	4.359	4.261	4.461	50.560	50.000	1.1
Decachlorobiphenyl	8.059	7.962	8.162	48.030	50.000	-3.9
Dieldrin	5.502	5.404	5.604	49.540	50.000	-0.9
Tetrachloro-m-xylene	2.874	2.775	2.975	51.260	50.000	2.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091161.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 16:32
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:05:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	146.5E6	1526.3E6	47.036	51.256
28)	SA Decachlor...	9.066	8.059	206.5E6	1454.8E6	44.153	48.028
Target Compounds							
2)	A alpha-BHC	3.994	3.385	347.4E6	2630.9E6	48.406	51.490
3)	MA gamma-BHC...	4.324	3.721	321.5E6	2401.9E6	47.288	50.835
4)	MA Heptachlor	4.922	4.073	304.5E6	2291.7E6	54.484	51.380
5)	MB Aldrin	5.263	4.359	304.4E6	2341.8E6	46.497	50.558
6)	B beta-BHC	4.512	4.018	121.1E6	983.0E6	46.224	49.575
7)	B delta-BHC	4.760	4.254	316.5E6	2400.5E6	46.810	50.307
8)	B Heptachlo...	5.684	4.862	265.6E6	2063.9E6	45.877	48.403
9)	A Endosulfan I	6.068	5.236	250.0E6	1828.0E6	45.154	46.408
10)	B gamma-Chl...	5.939	5.115	268.2E6	2216.8E6	45.224	49.318
11)	B alpha-Chl...	6.020	5.180	268.9E6	2121.0E6	43.502	49.168
12)	B 4,4'-DDE	6.190	5.363	235.7E6	2177.4E6	44.259	49.668
13)	MA Dieldrin	6.340	5.502	264.3E6	2189.1E6	45.330	49.544
14)	MA Endrin	6.568	5.779	209.5E6	1904.6E6	42.297	47.034
15)	B Endosulfa...	6.781	6.070	225.5E6	1833.8E6	45.103	48.830
16)	A 4,4'-DDD	6.700	5.919	206.2E6	1935.0E6	50.860	53.105
17)	MA 4,4'-DDT	7.015	6.172	168.8E6	1552.3E6	41.065	44.258
18)	B Endrin al...	6.910	6.248	166.7E6	1347.6E6	45.514	48.664
19)	B Endosulfa...	7.144	6.472	204.6E6	1718.7E6	44.115	47.949
20)	A Methoxychlor	7.488	6.743	88950083	783.3E6	41.847	44.945
21)	B Endrin ke...	7.625	6.981	228.0E6	1880.7E6	46.933	49.864
22)	Mirex	8.108	7.174	140.1E6	1226.6E6	43.759	47.746



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 16:32
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:05:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025 11/14/2025

Continuing Calib Time: 10:18

Initial Calibration Time(s): 20:56 21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
4,4'-DDT	7.02	7.01	6.91	7.11	-0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025 11/14/2025

Continuing Calib Time: 10:18

Initial Calibration Time(s): 20:56 21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3614
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091220.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.700	6.599	6.799	46.860	50.000	-6.3
4,4'-DDE	6.189	6.089	6.289	48.040	50.000	-3.9
4,4'-DDT	7.016	6.914	7.114	48.850	50.000	-2.3
Aldrin	5.263	5.163	5.363	48.120	50.000	-3.8
Decachlorobiphenyl	9.065	8.963	9.163	45.920	50.000	-8.2
Dieldrin	6.340	6.239	6.439	47.780	50.000	-4.4
Tetrachloro-m-xylene	3.544	3.444	3.644	48.000	50.000	-4.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3614
 GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No.: CCAL03 Date Analyzed: 11/17/2025
 Lab Sample No.: PSTDCCC050 Data File : PD091220.D Time Analyzed: 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.918	5.817	6.017	49.260	50.000	-1.5
4,4'-DDE	5.363	5.263	5.463	51.190	50.000	2.4
4,4'-DDT	6.172	6.071	6.271	52.440	50.000	4.9
Aldrin	4.358	4.258	4.458	52.180	50.000	4.4
Decachlorobiphenyl	8.059	7.958	8.158	48.700	50.000	-2.6
Dieldrin	5.502	5.401	5.601	51.270	50.000	2.5
Tetrachloro-m-xylene	2.873	2.774	2.974	51.820	50.000	3.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091220.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 10:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:08:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.873	165.5E6	1898.3E6	47.999	51.823
28)	SA Decachlor...	9.065	8.059	228.8E6	1927.4E6	45.917	48.702
Target Compounds							
2)	A alpha-BHC	3.993	3.385	398.7E6	3308.5E6	49.860	52.348
3)	MA gamma-BHC...	4.324	3.721	370.7E6	3023.0E6	49.608	52.282
4)	MA Heptachlor	4.922	4.073	350.5E6	2841.7E6	48.938	51.471
5)	MB Aldrin	5.263	4.358	349.2E6	2946.0E6	48.119	52.180
6)	B beta-BHC	4.511	4.017	138.1E6	1239.7E6	47.527	51.595
7)	B delta-BHC	4.760	4.254	366.7E6	3022.3E6	49.343	51.684
8)	B Heptachlo...	5.684	4.862	309.4E6	2623.9E6	47.620	51.491
9)	A Endosulfan I	6.067	5.236	289.7E6	2451.5E6	47.154	51.230
10)	B gamma-Chl...	5.939	5.115	310.1E6	2843.8E6	47.706	52.057
11)	B alpha-Chl...	6.020	5.179	313.2E6	2706.0E6	45.505	51.709
12)	B 4,4'-DDE	6.189	5.363	278.6E6	2735.9E6	48.036	51.194
13)	MA Dieldrin	6.340	5.502	308.9E6	2784.6E6	47.782	51.272
14)	MA Endrin	6.567	5.779	234.4E6	2265.9E6	47.232	50.139
15)	B Endosulfa...	6.781	6.070	257.1E6	2343.2E6	46.249	50.197
16)	A 4,4'-DDD	6.700	5.918	230.1E6	2407.4E6	46.858	49.263
17)	MA 4,4'-DDT	7.016	6.172	202.0E6	2052.3E6	48.845	52.445
18)	B Endrin al...	6.910	6.249	192.9E6	1760.0E6	46.315	49.875
19)	B Endosulfa...	7.145	6.472	239.0E6	2241.5E6	46.454	49.995
20)	A Methoxychlor	7.488	6.743	102.8E6	983.7E6	47.288	49.061
21)	B Endrin ke...	7.626	6.981	267.6E6	2547.5E6	47.505	49.994
22)	Mirex	8.108	7.173	165.1E6	1635.1E6	46.416	49.340



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091220.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 10:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:08:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025 11/14/2025

Continuing Calib Time: 20:02

Initial Calibration Time(s): 20:56 21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
4,4'-DDT	7.01	7.01	6.91	7.11	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025 11/14/2025

Continuing Calib Time: 20:02

Initial Calibration Time(s): 20:56 21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3614
 GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No.: CCAL04 Date Analyzed: 11/17/2025
 Lab Sample No.: PSTDCCC050 Data File : PD091236.D Time Analyzed: 20:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.697	6.599	6.799	44.500	50.000	-11.0
4,4'-DDE	6.187	6.089	6.289	49.510	50.000	-1.0
4,4'-DDT	7.013	6.914	7.114	58.790	50.000	17.6
Aldrin	5.261	5.163	5.363	48.730	50.000	-2.5
Decachlorobiphenyl	9.062	8.963	9.163	46.500	50.000	-7.0
Dieldrin	6.337	6.239	6.439	48.980	50.000	-2.0
Tetrachloro-m-xylene	3.542	3.444	3.644	48.160	50.000	-3.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3614
 GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No.: CCAL04 Date Analyzed: 11/17/2025
 Lab Sample No.: PSTDCCC050 Data File : PD091236.D Time Analyzed: 20:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	45.100	50.000	-9.8
4,4'-DDE	5.362	5.263	5.463	49.970	50.000	-0.1
4,4'-DDT	6.169	6.071	6.271	59.320	50.000	18.6
Aldrin	4.356	4.258	4.458	49.970	50.000	-0.1
Decachlorobiphenyl	8.057	7.958	8.158	44.480	50.000	-11.0
Dieldrin	5.500	5.401	5.601	49.880	50.000	-0.2
Tetrachloro-m-xylene	2.872	2.774	2.974	49.870	50.000	-0.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091236.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 20:02
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:11:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	166.1E6	1826.6E6	48.160	49.865
2)	SA Decachlor...	9.062	8.057	231.7E6	1760.3E6	46.499	44.481
Target Compounds							
2)	A alpha-BHC	3.991	3.383	402.8E6	3193.7E6	50.374	50.531
3)	MA gamma-BHC...	4.322	3.719	374.8E6	2869.0E6	50.158	49.618
4)	MA Heptachlor	4.919	4.071	369.0E6	2875.3E6	51.519	52.080
5)	MB Aldrin	5.261	4.356	353.6E6	2821.5E6	48.729	49.974
6)	B beta-BHC	4.509	4.015	139.2E6	1176.1E6	47.904	48.948
7)	B delta-BHC	4.757	4.252	338.6E6	2921.4E6	45.560	49.958
8)	B Heptachlo...	5.681	4.860	312.2E6	2512.4E6	48.057	49.303
9)	A Endosulfan I	6.065	5.234	297.1E6	2371.0E6	48.362	49.550
10)	B gamma-Chl...	5.937	5.113	316.9E6	2754.3E6	48.753	50.420
11)	B alpha-Chl...	6.017	5.178	319.8E6	2625.3E6	46.465	50.167
12)	B 4,4'-DDE	6.187	5.362	287.1E6	2670.7E6	49.505	49.973
13)	MA Dieldrin	6.337	5.500	316.6E6	2709.1E6	48.976	49.880
14)	MA Endrin	6.565	5.776	268.1E6	2497.0E6	54.017	55.253
15)	B Endosulfa...	6.778	6.068	276.1E6	2245.3E6	49.658	48.100
16)	A 4,4'-DDD	6.697	5.916	218.6E6	2203.8E6	44.504	45.096
17)	MA 4,4'-DDT	7.013	6.169	243.1E6	2321.3E6	58.785	59.318
18)	B Endrin al...	6.907	6.246	193.5E6	1675.1E6	46.458	47.468
19)	B Endosulfa...	7.142	6.470	243.6E6	2168.6E6	47.340	48.369
20)	A Methoxychlor	7.486	6.741	120.5E6	1140.9E6	55.416	56.901
21)	B Endrin ke...	7.623	6.978	254.2E6	2294.0E6	45.127	45.020
22)	Mirex	8.105	7.171	162.8E6	1485.3E6	45.776	44.818



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091236.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 20:02
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:11:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.066	8.970	9.170	22.860	20.000	14.3
Tetrachloro-m-xylene	3.544	3.490	3.590	21.710	20.000	8.6
alpha-BHC	3.994	3.940	4.040	9.050	10.000	-9.5
beta-BHC	4.512	4.460	4.560	10.070	10.000	0.7
gamma-BHC (Lindane)	4.324	4.270	4.370	9.330	10.000	-6.7
Endrin	6.568	6.500	6.640	49.200	50.000	-1.6
4,4'-DDT	7.016	6.950	7.090	100.270	100.000	0.3
Methoxychlor	7.489	7.420	7.560	236.810	250.000	-5.3

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.960	8.160	21.400	20.000	7.0
Tetrachloro-m-xylene	2.875	2.820	2.930	21.220	20.000	6.1
alpha-BHC	3.386	3.340	3.440	10.300	10.000	3.0
beta-BHC	4.019	3.970	4.070	10.460	10.000	4.6
gamma-BHC (Lindane)	3.723	3.670	3.770	10.340	10.000	3.4
Endrin	5.781	5.710	5.850	47.940	50.000	-4.1
4,4'-DDT	6.174	6.100	6.240	95.670	100.000	-4.3
Methoxychlor	6.745	6.670	6.820	201.830	250.000	-19.3

PEM
 Data File: PD090648.D Date Acquired 10/17/2025 11:07
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	243728295.7	249054181.9	5325886.17	2.14
Endrin aldehyde	6.91	1372426.584			
Endrin ketone	7.63	3953459.588			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	1941282940	1991060515	49777575.7	2.50
Endrin aldehyde #2	6.25	19940698.04			
Endrin ketone #2	6.98	29836877.65			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	412144601.3	415215115.8	3070514.5	0.74
4,4'-DDE	0.00	0			
4,4'-DDD	6.70	3070514.503			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3355351424	3394188860	38837436.1	1.14
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.92	38837436.09			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:07
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/19/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:04:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.875	67593250	632.0E6	21.705	21.224
28)	SA Decachlor...	9.066	8.062	106.9E6	648.1E6	22.859	21.395
Target Compounds							
2)	A alpha-BHC	3.994	3.386	64971308	526.5E6	9.053	10.304
3)	MA gamma-BHC...	4.324	3.723	63399907	488.4E6	9.325	10.337
6)	B beta-BHC	4.512	4.019	26386693	207.4E6	10.073	10.459
14)	MA Endrin	6.568	5.781	243.7E6	1941.3E6	49.201	47.940
16)	A 4,4'-DDD	6.701	5.923	3070515	38837436	0.757	1.066 #
17)	MA 4,4'-DDT	7.016	6.174	412.1E6	3355.4E6	100.267	95.667
18)	B Endrin al...	6.913	6.248	1372427	19940698	0.375m	0.720 #
20)	A Methoxychlor	7.489	6.745	503.3E6	3517.5E6	236.805	201.834
21)	B Endrin ke...	7.625	6.981	3953460	29836878	0.814	0.791m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:07
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:04:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD091142.D Date Analyzed: 11/13/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:34

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.064	8.960	9.160	21.760	20.000	8.8
Tetrachloro-m-xylene	3.544	3.490	3.590	21.470	20.000	7.4
alpha-BHC	3.993	3.940	4.040	9.150	10.000	-8.5
beta-BHC	4.511	4.460	4.560	10.130	10.000	1.3
gamma-BHC (Lindane)	4.323	4.270	4.370	9.220	10.000	-7.8
Endrin	6.567	6.500	6.640	44.260	50.000	-11.5
4,4'-DDT	7.015	6.940	7.090	87.100	100.000	-12.9
Methoxychlor	7.488	7.420	7.560	210.950	250.000	-15.6

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD091142.D Date Analyzed: 11/13/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:34

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.059	7.960	8.160	24.860	20.000	24.3
Tetrachloro-m-xylene	2.874	2.820	2.920	23.700	20.000	18.5
alpha-BHC	3.385	3.330	3.440	11.460	10.000	14.6
beta-BHC	4.017	3.970	4.070	11.600	10.000	16.0
gamma-BHC (Lindane)	3.721	3.670	3.770	11.490	10.000	14.9
Endrin	5.778	5.710	5.850	49.900	50.000	-0.2
4,4'-DDT	6.171	6.100	6.240	95.770	100.000	-4.2
Methoxychlor	6.742	6.670	6.810	210.840	250.000	-15.7

Data File: PEM
 PD091142.D Date Acquired 11/13/2025 9:34
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	219263452.1	240024357.3	20760905.2	8.65
Endrin aldehyde	6.91	5537180.139			
Endrin ketone	7.62	15223725.04			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	2020680787	2227596246	206915458	9.29
Endrin aldehyde #2	6.25	53922570.5			
Endrin ketone #2	6.98	152992888			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.01	358017564.4	393099657	35082092.7	8.92
4,4'-DDE	6.19	891389.46			
4,4'-DDD	6.70	34190703.22			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3358938544	3733736788	374798243	10.04
4,4'-DDE #2	5.36	8616195.174			
4,4'-DDD #2	5.92	366182048.2			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091142.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 09:34
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:02:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	66846269	705.8E6	21.465	23.702
28)	SA Decachlor...	9.064	8.059	101.8E6	753.1E6	21.756	24.861
Target Compounds							
2)	A alpha-BHC	3.993	3.385	65669256	585.5E6	9.150	11.458 #
3)	MA gamma-BHC...	4.323	3.721	62697745	542.7E6	9.222	11.486
6)	B beta-BHC	4.511	4.017	26535252	230.0E6	10.129	11.599
12)	B 4,4'-DDE	6.189	5.363	891389	8616195	0.167	0.197m
14)	MA Endrin	6.567	5.778	219.3E6	2020.7E6	44.263	49.901
16)	A 4,4'-DDD	6.699	5.918	34190703	366.2E6	8.433	10.049
17)	MA 4,4'-DDT	7.015	6.171	358.0E6	3358.9E6	87.099	95.770
18)	B Endrin al...	6.910	6.247	5537180	53922570	1.512	1.947 #
20)	A Methoxychlor	7.488	6.742	448.4E6	3674.4E6	210.946	210.836
21)	B Endrin ke...	7.624	6.980	15223725	153.0E6	3.133	4.056 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 09:34
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:02:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091194.D Date Analyzed: 11/14/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:15

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.064	8.960	9.160	20.170	20.000	0.9
Tetrachloro-m-xylene	3.544	3.490	3.590	19.500	20.000	-2.5
alpha-BHC	3.993	3.940	4.040	8.250	10.000	-17.5
beta-BHC	4.511	4.460	4.560	9.240	10.000	-7.6
gamma-BHC (Lindane)	4.324	4.270	4.370	8.450	10.000	-15.5
Endrin	6.567	6.500	6.640	44.390	50.000	-11.2
4,4'-DDT	7.015	6.940	7.090	88.290	100.000	-11.7
Methoxychlor	7.487	7.420	7.560	207.040	250.000	-17.2

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091194.D Date Analyzed: 11/14/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:15

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.059	7.960	8.160	19.620	20.000	-1.9
Tetrachloro-m-xylene	2.874	2.820	2.920	19.370	20.000	-3.2
alpha-BHC	3.385	3.330	3.440	9.380	10.000	-6.2
beta-BHC	4.018	3.970	4.070	9.530	10.000	-4.7
gamma-BHC (Lindane)	3.721	3.670	3.770	9.460	10.000	-5.4
Endrin	5.779	5.710	5.850	45.450	50.000	-9.1
4,4'-DDT	6.172	6.100	6.240	89.340	100.000	-10.7
Methoxychlor	6.742	6.670	6.810	186.630	250.000	-25.3

PEM
 Data File: PD091194.D Date Acquired 11/14/2025 20:15
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down Down
Endrin	6.57	220342338.4	243232929.5	22890591.1	9.41
Endrin aldehyde	6.91	5179237.826			
Endrin ketone	7.62	17711353.28			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	2053826744	2298027068	244200325	10.63
Endrin aldehyde #2	6.25	61443615.46			
Endrin ketone #2	6.98	182756709.2			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.01	365114982.7	402071544.1	36956561.3	9.19
4,4'-DDE	6.19	1068437.233			
4,4'-DDD	6.70	35888124.09			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3496021709	3894922521	398900812	10.24
4,4'-DDE #2	5.36	8351296.931			
4,4'-DDD #2	5.92	390549515.4			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091194.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 20:15
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/17/2025
 Supervised By : Sohil Jodhani 11/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:48:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	67260764	709.4E6	19.504	19.367
28)	SA Decachlor...	9.064	8.059	100.5E6	776.5E6	20.172	19.621
Target Compounds							
2)	A alpha-BHC	3.993	3.385	65992882	592.9E6	8.254	9.381
3)	MA gamma-BHC...	4.324	3.721	63139958	546.9E6	8.449	9.458
6)	B beta-BHC	4.511	4.018	26858838	229.1E6	9.244	9.533
12)	B 4,4'-DDE	6.188	5.362	1068437	8351297	0.184	0.156m
14)	MA Endrin	6.567	5.779	220.3E6	2053.8E6	44.393	45.446
16)	A 4,4'-DDD	6.698	5.918	35888124	390.5E6	7.307	7.992
17)	MA 4,4'-DDT	7.015	6.172	365.1E6	3496.0E6	88.288	89.338
18)	B Endrin al...	6.908	6.246	5179238	61443615	1.243m	1.741m#
20)	A Methoxychlor	7.487	6.742	450.2E6	3742.1E6	207.044	186.626
21)	B Endrin ke...	7.624	6.980	17711353	182.8E6	3.144	3.587

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091194.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:15
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/17/2025
Supervised By : Sohil Jodhani 11/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:48:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091219.D Date Analyzed: 11/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.067	8.970	9.170	20.390	20.000	2.0
Tetrachloro-m-xylene	3.544	3.490	3.590	21.040	20.000	5.2
alpha-BHC	3.993	3.940	4.040	8.920	10.000	-10.8
beta-BHC	4.511	4.460	4.560	9.940	10.000	-0.6
gamma-BHC (Lindane)	4.324	4.270	4.370	9.120	10.000	-8.8
Endrin	6.567	6.500	6.640	44.720	50.000	-10.6
4,4'-DDT	7.016	6.950	7.090	91.520	100.000	-8.5
Methoxychlor	7.489	7.420	7.560	211.580	250.000	-15.4

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091219.D Date Analyzed: 11/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.060	7.960	8.160	20.690	20.000	3.5
Tetrachloro-m-xylene	2.873	2.820	2.920	22.610	20.000	13.1
alpha-BHC	3.385	3.330	3.440	11.010	10.000	10.1
beta-BHC	4.017	3.970	4.070	11.240	10.000	12.4
gamma-BHC (Lindane)	3.721	3.670	3.770	11.030	10.000	10.3
Endrin	5.777	5.710	5.850	47.210	50.000	-5.6
4,4'-DDT	6.171	6.100	6.240	98.490	100.000	-1.5
Methoxychlor	6.741	6.670	6.810	191.760	250.000	-23.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091219.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 10:04
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:07:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.873	72576102	828.3E6	21.045	22.611
28)	SA Decachlor...	9.067	8.060	101.6E6	818.9E6	20.391	20.693
Target Compounds							
2)	A alpha-BHC	3.993	3.385	71357424	696.0E6	8.925	11.012
3)	MA gamma-BHC...	4.324	3.721	68153737	637.8E6	9.120	11.031
6)	B beta-BHC	4.511	4.017	28881769	270.2E6	9.940	11.245
12)	B 4,4'-DDE	6.189	5.364	1587125	18405683	0.274	0.344 #
14)	MA Endrin	6.567	5.777	222.0E6	2133.5E6	44.724	47.209
16)	A 4,4'-DDD	6.700	5.918	34824362	411.6E6	7.090	8.422
17)	MA 4,4'-DDT	7.016	6.171	378.5E6	3854.1E6	91.516	98.487
18)	B Endrin al...	6.910	6.247	7568926	84362914	1.817	2.391 #
20)	A Methoxychlor	7.489	6.741	460.1E6	3845.0E6	211.583	191.757
21)	B Endrin ke...	7.626	6.980	19517946	235.0E6	3.465	4.612 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 10:04
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:07:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090649.D
Acq On : 17 Oct 2025 11:20
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Title : GC Extractables
Last Update : Fri Oct 17 17:43:28 2025
Integrator: ChemStation

RT#1	RT#2	Resolution
3.544	5.940	100.00%
5.940	6.068	100.00%
6.068	6.190	100.00%
6.190	6.341	100.00%
6.341	7.144	100.00%
7.144	7.488	100.00%
7.488	7.625	100.00%
7.625	9.066	100.00%

Signal #2

2.874	5.116	100.00%
5.116	5.238	100.00%
5.238	5.366	100.00%
5.366	5.503	100.00%
5.503	6.474	100.00%
6.474	6.745	100.00%
6.745	6.983	100.00%
6.983	8.062	100.00%

PD101725.M Sat Oct 25 06:10:53 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090649.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:20
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESCHK

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/19/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:04:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.874	55125186	545.1E6	17.701m	18.307
28)	SA Decachlor...	9.066	8.062	88858238	547.9E6	18.995	18.088
Target Compounds							
9)	A Endosulfan I	6.068	5.238	43926319	330.7E6	7.934	8.397
10)	B gamma-Chl...	5.940	5.116	47696248	387.3E6	8.042	8.616
12)	B 4,4'-DDE	6.190	5.366	86477351	754.1E6	16.235	17.202
13)	MA Dieldrin	6.341	5.503	92552800	748.0E6	15.872	16.930
19)	B Endosulfa...	7.144	6.474	76754913	612.1E6	16.548	17.078
20)	A Methoxychlor	7.488	6.745	169.7E6	1331.6E6	79.844	76.409
21)	B Endrin ke...	7.625	6.983	80182457	638.1E6	16.502	16.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090649.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:20
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
RESCHK

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/19/2025
Supervised By : mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:04:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091195.D
Acq On : 14 Nov 2025 20:43
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 51 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Title : GC Extractables
Last Update : Sat Nov 15 04:46:13 2025
Integrator: ChemStation

RT#1	RT#2	Resolution
3.544	5.939	100.00%
5.939	6.067	100.00%
6.067	6.189	100.00%
6.189	6.339	100.00%
6.339	7.143	100.00%
7.143	7.487	100.00%
7.487	7.624	100.00%
7.624	9.063	100.00%

Signal #2

2.874	5.114	100.00%
5.114	5.235	100.00%
5.235	5.363	100.00%
5.363	5.501	100.00%
5.501	6.471	100.00%
6.471	6.742	100.00%
6.742	6.980	100.00%
6.980	8.059	100.00%

PD111525.M Sat Nov 15 05:00:31 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091195.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 20:43
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:48:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	59572378	631.9E6	17.274	17.250
28)	SA Decachlor...	9.063	8.059	86552368	667.6E6	17.373	16.869
Target Compounds							
9)	A Endosulfan I	6.067	5.235	45261732	379.7E6	7.368	7.934
10)	B gamma-Chl...	5.939	5.114	49362700	455.8E6	7.593	8.344
12)	B 4,4'-DDE	6.189	5.363	87544052	881.3E6	15.095	16.491
13)	MA Dieldrin	6.339	5.501	96362551	874.4E6	14.905	16.100
19)	B Endosulfa...	7.143	6.471	78876117	723.4E6	15.331	16.134
20)	A Methoxychlor	7.487	6.742	154.5E6	1426.0E6	71.049	71.120
21)	B Endrin ke...	7.624	6.980	80568543	766.5E6	14.302	15.043



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091195.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:43
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:48:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

Instrument ID: ECD_D

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 10/17/2025

10/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
LBLK	LBLK	10/17/2025	10:53	PD090647.D	9.07	3.55
PEM	PEM	10/17/2025	11:07	PD090648.D	9.07	3.54
RESCHK	RESCHK	10/17/2025	11:20	PD090649.D	9.07	3.54
PSTDICC100	PSTDICC100	10/17/2025	11:34	PD090650.D	9.07	3.54
PSTDICC075	PSTDICC075	10/17/2025	11:48	PD090651.D	9.07	3.55
PSTDICC050	PSTDICC050	10/17/2025	12:01	PD090652.D	9.06	3.55
PSTDICC025	PSTDICC025	10/17/2025	12:15	PD090653.D	9.07	3.54
PSTDICC005	PSTDICC005	10/17/2025	12:28	PD090654.D	9.07	3.54
PCHLORICC500	PCHLORICC500	10/17/2025	13:09	PD090657.D	9.07	3.55
PTOXICC500	PTOXICC500	10/17/2025	14:17	PD090662.D	9.07	3.55
LBLK	LBLK	11/13/2025	09:20	PD091141.D	9.07	3.54
PEM	PEM	11/13/2025	09:34	PD091142.D	9.06	3.54
PSTDCCC050	PSTDCCC050	11/13/2025	09:47	PD091143.D	9.06	3.54
PB170527BL	PB170527BL	11/13/2025	12:17	PD091144.D	9.08	3.55
AU-713-COMP-05MS	Q3609-13MS	11/13/2025	13:52	PD091151.D	9.06	3.54
AU-713-COMP-05MSD	Q3609-13MSD	11/13/2025	14:06	PD091152.D	9.06	3.54
COMP-1	Q3614-01	11/13/2025	14:20	PD091153.D	9.07	3.54
COMP-2	Q3614-02	11/13/2025	14:33	PD091154.D	9.07	3.54
COMP-3	Q3614-03	11/13/2025	14:47	PD091155.D	9.07	3.54
LBLK	LBLK	11/13/2025	16:19	PD091160.D	9.08	3.55
PSTDCCC050	PSTDCCC050	11/13/2025	16:32	PD091161.D	9.07	3.54
LBLK	LBLK	11/14/2025	20:01	PD091193.D	9.06	3.54
PEM	PEM	11/14/2025	20:15	PD091194.D	9.06	3.54
RESCHK	RESCHK	11/14/2025	20:43	PD091195.D	9.06	3.54
PSTDICC100	PSTDICC100	11/14/2025	20:56	PD091196.D	9.06	3.54
PSTDICC075	PSTDICC075	11/14/2025	21:10	PD091197.D	9.06	3.54
PSTDICC050	PSTDICC050	11/14/2025	21:24	PD091198.D	9.06	3.54
PSTDICC025	PSTDICC025	11/14/2025	21:37	PD091199.D	9.07	3.54
PSTDICC005	PSTDICC005	11/14/2025	21:51	PD091200.D	9.06	3.54
PCHLORICC500	PCHLORICC500	11/14/2025	22:32	PD091203.D	9.06	3.54
PTOXICC500	PTOXICC500	11/14/2025	23:40	PD091208.D	9.06	3.54
LBLK	LBLK	11/17/2025	09:51	PD091218.D	9.07	3.55
PEM	PEM	11/17/2025	10:04	PD091219.D	9.07	3.54
PSTDCCC050	PSTDCCC050	11/17/2025	10:18	PD091220.D	9.07	3.54
PB170527BS	PB170527BS	11/17/2025	12:54	PD091221.D	9.07	3.55
LBLK	LBLK	11/17/2025	19:21	PD091234.D	9.06	3.54
PSTDCCC050	PSTDCCC050	11/17/2025	20:02	PD091236.D	9.06	3.54

Analytical Sequence

Client: Kleinfelder

SDG No.: Q3614

Project: Logan School

Instrument ID: ECD_D

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 10/17/2025

10/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
LBLK	LBLK	10/17/2025	10:53	PD090647.D	8.06	2.88
PEM	PEM	10/17/2025	11:07	PD090648.D	8.06	2.88
RESCHK	RESCHK	10/17/2025	11:20	PD090649.D	8.06	2.87
PSTDICC100	PSTDICC100	10/17/2025	11:34	PD090650.D	8.06	2.87
PSTDICC075	PSTDICC075	10/17/2025	11:48	PD090651.D	8.06	2.88
PSTDICC050	PSTDICC050	10/17/2025	12:01	PD090652.D	8.06	2.88
PSTDICC025	PSTDICC025	10/17/2025	12:15	PD090653.D	8.06	2.87
PSTDICC005	PSTDICC005	10/17/2025	12:28	PD090654.D	8.06	2.87
PCHLORICC500	PCHLORICC500	10/17/2025	13:09	PD090657.D	8.06	2.87
PTOXICC500	PTOXICC500	10/17/2025	14:17	PD090662.D	8.06	2.88
LBLK	LBLK	11/13/2025	09:20	PD091141.D	8.06	2.87
PEM	PEM	11/13/2025	09:34	PD091142.D	8.06	2.87
PSTDCCC050	PSTDCCC050	11/13/2025	09:47	PD091143.D	8.06	2.87
PB170527BL	PB170527BL	11/13/2025	12:17	PD091144.D	8.06	2.87
AU-713-COMP-05MS	Q3609-13MS	11/13/2025	13:52	PD091151.D	8.06	2.87
AU-713-COMP-05MSD	Q3609-13MSD	11/13/2025	14:06	PD091152.D	8.06	2.87
COMP-1	Q3614-01	11/13/2025	14:20	PD091153.D	8.06	2.87
COMP-2	Q3614-02	11/13/2025	14:33	PD091154.D	8.06	2.87
COMP-3	Q3614-03	11/13/2025	14:47	PD091155.D	8.06	2.87
LBLK	LBLK	11/13/2025	16:19	PD091160.D	8.06	2.87
PSTDCCC050	PSTDCCC050	11/13/2025	16:32	PD091161.D	8.06	2.87
LBLK	LBLK	11/14/2025	20:01	PD091193.D	8.06	2.88
PEM	PEM	11/14/2025	20:15	PD091194.D	8.06	2.87
RESCHK	RESCHK	11/14/2025	20:43	PD091195.D	8.06	2.87
PSTDICC100	PSTDICC100	11/14/2025	20:56	PD091196.D	8.06	2.87
PSTDICC075	PSTDICC075	11/14/2025	21:10	PD091197.D	8.06	2.87
PSTDICC050	PSTDICC050	11/14/2025	21:24	PD091198.D	8.06	2.87
PSTDICC025	PSTDICC025	11/14/2025	21:37	PD091199.D	8.06	2.87
PSTDICC005	PSTDICC005	11/14/2025	21:51	PD091200.D	8.06	2.87
PCHLORICC500	PCHLORICC500	11/14/2025	22:32	PD091203.D	8.06	2.87
PTOXICC500	PTOXICC500	11/14/2025	23:40	PD091208.D	8.06	2.87
LBLK	LBLK	11/17/2025	09:51	PD091218.D	8.06	2.87
PEM	PEM	11/17/2025	10:04	PD091219.D	8.06	2.87
PSTDCCC050	PSTDCCC050	11/17/2025	10:18	PD091220.D	8.06	2.87
PB170527BS	PB170527BS	11/17/2025	12:54	PD091221.D	8.06	2.87
LBLK	LBLK	11/17/2025	19:21	PD091234.D	8.06	2.87
PSTDCCC050	PSTDCCC050	11/17/2025	20:02	PD091236.D	8.06	2.87

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

AU-713-COMP-05MS

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Lab Sample ID: Q3609-13MS

Date(s) Analyzed: 11/13/2025 11/13/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	22.0	4.9
	2	5.92	5.87	5.97	23.1	
4,4'-DDT	1	7.02	6.97	7.07	19.8	12.3
	2	6.17	6.12	6.22	22.4	
Aldrin	1	5.26	5.21	5.31	20.6	6.1
	2	4.36	4.31	4.41	21.9	
4,4'-DDE	1	6.19	6.14	6.24	20.2	10.8
	2	5.36	5.31	5.41	22.5	
Dieldrin	1	6.34	6.29	6.39	20.5	9.7
	2	5.50	5.45	5.55	22.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

AU-713-COMP-05MSD

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Lab Sample ID: Q3609-13MSD

Date(s) Analyzed: 11/13/2025 11/13/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	20.8	4.7
	2	5.92	5.87	5.97	21.8	
4,4'-DDT	1	7.01	6.96	7.06	18.8	11.5
	2	6.17	6.12	6.22	21.1	
Aldrin	1	5.26	5.21	5.31	19.5	5.5
	2	4.36	4.31	4.41	20.6	
4,4'-DDE	1	6.19	6.14	6.24	19.1	10.9
	2	5.36	5.31	5.41	21.3	
Dieldrin	1	6.34	6.29	6.39	19.5	8.8
	2	5.50	5.45	5.55	21.3	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170527BS

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3614

Lab Sample ID: PB170527BS

Date(s) Analyzed: 11/17/2025 11/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.71	6.66	6.76	16.0	3.7
	2	5.92	5.87	5.97	16.6	
4,4'-DDT	1	7.02	6.97	7.07	16.5	6.5
	2	6.17	6.12	6.22	17.6	
Aldrin	1	5.27	5.22	5.32	16.2	6
	2	4.36	4.31	4.41	17.2	
4,4'-DDE	1	6.20	6.15	6.25	16.3	4.8
	2	5.36	5.31	5.41	17.1	
Dieldrin	1	6.35	6.30	6.40	16.3	4.8
	2	5.50	5.45	5.55	17.1	



QC SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder		Date Collected:	
Project:	Logan School		Date Received:	
Client Sample ID:	PB170527BL		SDG No.:	Q3614
Lab Sample ID:	PB170527BL		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100
Sample Wt/Vol:	30.02 g	Final Vol: 10000 uL	Test:	PESTICIDE Group1
Prep Method:	SW3541B		Prep Date:	11/13/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.12	U	1	0.12	1.70	ug/kg	11/13/25 12:17	PB170527
60-57-1	Dieldrin	0.14	U	1	0.14	1.70	ug/kg	11/13/25 12:17	PB170527
72-55-9	4,4-DDE	0.14	U	1	0.14	1.70	ug/kg	11/13/25 12:17	PB170527
72-54-8	4,4-DDD	0.15	U	1	0.15	1.70	ug/kg	11/13/25 12:17	PB170527
50-29-3	4,4-DDT	0.14	U	1	0.14	1.70	ug/kg	11/13/25 12:17	PB170527
SURROGATES									
2051-24-3	Decachlorobiphenyl	21.6			10 - 143	108%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.9			10 - 131	104%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091144.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 12:17
 Operator : AR\AJ
 Sample : PB170527BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB170527BL

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:02:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.552	2.873	55275000	620.7E6	17.750	20.846
2) SA Decachlor...	9.075	8.061	88119380	653.6E6	18.837m	21.578

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091144.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 12:17
Operator : AR\AJ
Sample : PB170527BL
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PB170527BL


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:02:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.552 min
Delta R.T.: 0.007 min
Response: 55275000
Conc: 17.75 ng/ml

Instrument :

ECD_D


ClientSampleId :

PB170527BL

Manual Integrations


APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.002 min
Response: 620748906
Conc: 20.85 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.075 min
Delta R.T.: 0.011 min
Response: 88119380
Conc: 18.84 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 653637586
Conc: 21.58 ng/ml

Report of Analysis

Client: Kleinfelder	Date Collected: 10/17/25	
Project: Logan School	Date Received: 10/17/25	
Client Sample ID: PIBLK-PD090647.D	SDG No.: Q3614	
Lab Sample ID: I.BLK-PD090647.D	Matrix: WATER	
Analytical Method: 8081B	% Solid: 0	
Sample Wt/Vol: 1000 mL	Test: PESTICIDE Group1	Final Vol: 10000 uL
Prep Method: 3510C		Prep Date:

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	10/17/25	pd101725
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	10/17/25	pd101725
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	10/17/25	pd101725
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	10/17/25	pd101725
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	10/17/25	pd101725
SURROGATES									
2051-24-3	Decachlorobiphenyl	21.3			31 - 149	106%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.0			41 - 134	100%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 10:53
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:03:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.875	60526543	594.5E6	19.436	19.964
28)	SA Decachlor...	9.066	8.062	99514776	616.2E6	21.273	20.342

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 10:53
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:03:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 60526543
Conc: 19.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 594475365
Conc: 19.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 99514776
Conc: 21.27 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 616188932
Conc: 20.34 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder		Date Collected:	11/13/25
Project:	Logan School		Date Received:	11/13/25
Client Sample ID:	PIBLK-PD091141.D		SDG No.:	Q3614
Lab Sample ID:	I.BLK-PD091141.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PESTICIDE Group1
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/13/25	PD111325
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/13/25	PD111325
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/13/25	PD111325
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/13/25	PD111325
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/13/25	PD111325
SURROGATES									
2051-24-3	Decachlorobiphenyl	24.7			31 - 149	123%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	24.7			41 - 134	123%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 09:20
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:02:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.873	68060701	734.7E6	21.855	24.675
28)	SA Decachlor...	9.065	8.059	99696988	747.9E6	21.312	24.691

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 09:20
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:02:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 68060701
Conc: 21.86 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.001 min
Response: 734742826
Conc: 24.67 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 99696988
Conc: 21.31 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.003 min
Response: 747929551
Conc: 24.69 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder	Date Collected:	11/13/25
Project:	Logan School	Date Received:	11/13/25
Client Sample ID:	PIBLK-PD091160.D	SDG No.:	Q3614
Lab Sample ID:	I.BLK-PD091160.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL
Prep Method:	3510C	Test:	PESTICIDE Group1
	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/13/25	pd111325
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/13/25	pd111325
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/13/25	pd111325
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/13/25	pd111325
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/13/25	pd111325
SURROGATES									
2051-24-3	Decachlorobiphenyl	21.1			31 - 149	105%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	23.4			41 - 134	117%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091160.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 16:19
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:05:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.874	64086617	696.0E6	20.579	23.372
28)	SA Decachlor...	9.076	8.061	95449334	638.0E6	20.404	21.063

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091160.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 16:19
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:05:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: 0.006 min
Response: 64086617
Conc: 20.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: -0.001 min
Response: 695955834
Conc: 23.37 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.076 min
Delta R.T.: 0.011 min
Response: 95449334
Conc: 20.40 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 638023268
Conc: 21.06 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder	Date Collected:	11/14/25
Project:	Logan School	Date Received:	11/14/25
Client Sample ID:	PIBLK-PD091193.D	SDG No.:	Q3614
Lab Sample ID:	I.BLK-PD091193.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL
Prep Method:	3510C	Test:	PESTICIDE Group1
	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/14/25	PD111525
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/14/25	PD111525
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/14/25	PD111525
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/14/25	PD111525
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/14/25	PD111525
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.8			31 - 149	104%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.0			41 - 134	100%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091193.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:01
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:47:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.875	67367446	734.0E6	19.535	20.038
28)	SA Decachlor...	9.064	8.059	103.5E6	793.9E6	20.767	20.060

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091193.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:01
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:47:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 67367446
Conc: 19.53 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 734006061
Conc: 20.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.064 min
Delta R.T.: 0.000 min
Response: 103459386
Conc: 20.77 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: 0.000 min
Response: 793870994
Conc: 20.06 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder		Date Collected:	11/17/25
Project:	Logan School		Date Received:	11/17/25
Client Sample ID:	PIBLK-PD091218.D		SDG No.:	Q3614
Lab Sample ID:	I.BLK-PD091218.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PESTICIDE Group1
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/17/25	pd111725
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/17/25	pd111725
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.0			31 - 149	100%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	22.6			41 - 134	113%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091218.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 09:51
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:07:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.873	67500173	827.2E6	19.573	22.584
28)	SA Decachlor...	9.071	8.061	99597433	763.2E6	19.992	19.286

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 09:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:07:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.548 min
Delta R.T.: 0.004 min
Response: 67500173
Conc: 19.57 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: 0.000 min
Response: 827248235
Conc: 22.58 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.071 min
Delta R.T.: 0.007 min
Response: 99597433
Conc: 19.99 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.003 min
Response: 763248619
Conc: 19.29 ng/ml

Report of Analysis

Client: Kleinfelder	Date Collected: 11/17/25	
Project: Logan School	Date Received: 11/17/25	
Client Sample ID: PIBLK-PD091234.D	SDG No.: Q3614	
Lab Sample ID: I.BLK-PD091234.D	Matrix: WATER	
Analytical Method: 8081B	% Solid: 0	
Sample Wt/Vol: 1000 mL	Test: PESTICIDE Group1	Final Vol: 10000 uL
Prep Method: 3510C		Prep Date:

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/17/25	pd111725
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/17/25	pd111725
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.8			31 - 149	104%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.4			41 - 134	107%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091234.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 19:21
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:11:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	68845015	784.5E6	19.963	21.417
28)	SA Decachlor...	9.062	8.057	103.4E6	746.6E6	20.751	18.866

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091234.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 19:21
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:11:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.542 min
Delta R.T.: -0.002 min
Response: 68845015
Conc: 19.96 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 784503654
Conc: 21.42 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.062 min
Delta R.T.: -0.001 min
Response: 103377420
Conc: 20.75 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.001 min
Response: 746618568
Conc: 18.87 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Logan School	Date Received:	
Client Sample ID:	PB170527BS	SDG No.:	Q3614
Lab Sample ID:	PB170527BS	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	100
Sample Wt/Vol:	30.03 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	PESTICIDE Group1
	Prep Date:	11/13/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	17.2		1	0.12	1.70	ug/kg	11/17/25 12:54	PB170527
60-57-1	Dieldrin	17.1		1	0.14	1.70	ug/kg	11/17/25 12:54	PB170527
72-55-9	4,4-DDE	17.1		1	0.14	1.70	ug/kg	11/17/25 12:54	PB170527
72-54-8	4,4-DDD	16.6		1	0.15	1.70	ug/kg	11/17/25 12:54	PB170527
50-29-3	4,4-DDT	17.6		1	0.14	1.70	ug/kg	11/17/25 12:54	PB170527
SURROGATES									
2051-24-3	Decachlorobiphenyl	24.9			10 - 143	124%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	24.1			10 - 131	120%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091221.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 12:54
 Operator : AR\AJ
 Sample : PB170527BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB170527BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:08:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.873	74181191	880.8E6	21.510	24.046
28)	SA Decachlor...	9.073	8.061	117.0E6	983.7E6	23.487	24.857
Target Compounds							
2)	A alpha-BHC	3.999	3.385	392.5E6	3246.6E6	49.094	51.368
3)	MA gamma-BHC...	4.330	3.721	366.1E6	2979.4E6	48.984	51.528
4)	MA Heptachlor	4.928	4.073	350.5E6	2806.9E6	48.936	50.842
5)	MB Aldrin	5.269	4.359	353.3E6	2916.6E6	48.689	51.660
6)	B beta-BHC	4.517	4.017	136.6E6	1219.4E6	47.021	50.750
7)	B delta-BHC	4.766	4.254	366.6E6	2995.4E6	49.327	51.224
8)	B Heptachlo...	5.689	4.862	312.5E6	2637.4E6	48.100	51.755
9)	A Endosulfan I	6.074	5.237	296.6E6	2467.1E6	48.282	51.557
10)	B gamma-Chl...	5.946	5.115	318.5E6	2840.9E6	48.993	52.005
11)	B alpha-Chl...	6.026	5.180	317.7E6	2702.3E6	46.164	51.639
12)	B 4,4'-DDE	6.196	5.364	283.4E6	2737.5E6	48.860	51.224
13)	MA Dieldrin	6.346	5.502	315.8E6	2782.5E6	48.846	51.233
14)	MA Endrin	6.574	5.779	221.0E6	2083.0E6	44.523	46.092
15)	B Endosulfa...	6.787	6.071	269.3E6	2372.0E6	48.434	50.814
16)	A 4,4'-DDD	6.706	5.919	236.3E6	2438.6E6	48.110	49.902
17)	MA 4,4'-DDT	7.022	6.173	205.0E6	2071.8E6	49.578	52.942
18)	B Endrin al...	6.915	6.250	206.1E6	1869.5E6	49.479	52.976
19)	B Endosulfa...	7.151	6.473	245.5E6	2278.4E6	47.713	50.818
20)	A Methoxychlor	7.495	6.743	100.5E6	979.1E6	46.209	48.829
21)	B Endrin ke...	7.632	6.982	281.3E6	2679.1E6	49.929	52.578
22)	Mirex	8.114	7.175	165.6E6	1647.6E6	46.562	49.715



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091221.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 12:54
Operator : AR\AJ
Sample : PB170527BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170527BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:08:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder	Date Collected:	11/11/25
Project:	Logan School	Date Received:	11/12/25
Client Sample ID:	AU-713-COMP-05MS	SDG No.:	Q3614
Lab Sample ID:	Q3609-13MS	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	84.9
Sample Wt/Vol:	30.03 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	PESTICIDE Group1
	Prep Date:		11/13/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	21.9		1	0.14	2.00	ug/kg	11/13/25 13:52	PB170527
60-57-1	Dieldrin	22.6		1	0.16	2.00	ug/kg	11/13/25 13:52	PB170527
72-55-9	4,4-DDE	22.5		1	0.16	2.00	ug/kg	11/13/25 13:52	PB170527
72-54-8	4,4-DDD	23.1		1	0.18	2.00	ug/kg	11/13/25 13:52	PB170527
50-29-3	4,4-DDT	22.4		1	0.16	2.00	ug/kg	11/13/25 13:52	PB170527
SURROGATES									
2051-24-3	Decachlorobiphenyl	19.3			10 - 143	96%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.7			10 - 131	104%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091151.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 13:52
 Operator : AR\AJ
 Sample : Q3609-13MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 AU-713-COMP-05MS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:03:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.874	53277660	616.5E6	17.108m	20.703
28)	SA Decachlor...	9.064	8.059	77448426	583.1E6	16.556	19.250
Target Compounds							
2)	A alpha-BHC	3.993	3.385	387.7E6	2977.7E6	54.017	58.277
3)	MA gamma-BHC...	4.323	3.721	356.5E6	2729.2E6	52.435	57.763
4)	MA Heptachlor	4.921	4.073	346.2E6	2639.2E6	61.944	59.171
5)	MB Aldrin	5.262	4.358	343.2E6	2580.7E6	52.419	55.716
6)	B beta-BHC	4.511	4.017	137.2E6	1125.6E6	52.389	56.762
7)	B delta-BHC	4.759	4.253	360.3E6	2750.8E6	53.280	57.649
8)	B Heptachlo...	5.682	4.862	303.9E6	2404.8E6	52.501m	56.398
9)	A Endosulfan I	6.067	5.236	286.6E6	2263.0E6	51.760	57.451
10)	B gamma-Chl...	5.939	5.114	305.5E6	2563.7E6	51.504	57.037
11)	B alpha-Chl...	6.019	5.179	305.8E6	2480.6E6	49.482	57.503
12)	B 4,4'-DDE	6.189	5.364	274.7E6	2520.0E6	51.574	57.485
13)	MA Dieldrin	6.339	5.502	305.2E6	2543.1E6	52.346	57.558
14)	MA Endrin	6.567	5.778	239.8E6	2130.5E6	48.399	52.613
15)	B Endosulfa...	6.780	6.070	256.4E6	2189.8E6	51.277	58.309
16)	A 4,4'-DDD	6.699	5.918	227.0E6	2141.7E6	55.996	58.775
17)	MA 4,4'-DDT	7.015	6.172	207.2E6	2006.2E6	50.417	57.200
18)	B Endrin al...	6.909	6.248	191.0E6	1647.4E6	52.135	59.489
19)	B Endosulfa...	7.143	6.472	235.9E6	2113.1E6	50.868	58.950
20)	A Methoxychlor	7.488	6.742	109.3E6	1009.4E6	51.422	57.921
21)	B Endrin ke...	7.624	6.981	266.0E6	2393.5E6	54.736	63.458
22)	Mirex	8.106	7.173	159.2E6	1508.7E6	49.730	58.725

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091151.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 13:52
Operator : AR\AJ
Sample : Q3609-13MS
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
AU-713-COMP-05MS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:03:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Kleinfelder	Date Collected:	11/11/25
Project:	Logan School	Date Received:	11/12/25
Client Sample ID:	AU-713-COMP-05MSD	SDG No.:	Q3614
Lab Sample ID:	Q3609-13MSD	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	84.9
Sample Wt/Vol:	30.02 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	PESTICIDE Group1
	Prep Date:		11/13/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
309-00-2	Aldrin	20.6		1	0.14	2.00	ug/kg	11/13/25 14:06	PB170527
60-57-1	Dieldrin	21.3		1	0.16	2.00	ug/kg	11/13/25 14:06	PB170527
72-55-9	4,4-DDE	21.3		1	0.16	2.00	ug/kg	11/13/25 14:06	PB170527
72-54-8	4,4-DDD	21.8		1	0.18	2.00	ug/kg	11/13/25 14:06	PB170527
50-29-3	4,4-DDT	21.1		1	0.16	2.00	ug/kg	11/13/25 14:06	PB170527
SURROGATES									
2051-24-3	Decachlorobiphenyl	18.6			10 - 143	93%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.6			10 - 131	98%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
 Data File : PD091152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2025 14:06
 Operator : AR\AJ
 Sample : Q3609-13MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 AU-713-COMP-05MSD

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/14/2025
 Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 22:03:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.874	51500650	584.0E6	16.538m	19.612
28)	SA Decachlor...	9.064	8.059	71486078	561.8E6	15.281	18.546
Target Compounds							
2)	A alpha-BHC	3.993	3.385	364.1E6	2799.3E6	50.727	54.786
3)	MA gamma-BHC...	4.324	3.721	336.3E6	2564.6E6	49.467	54.279
4)	MA Heptachlor	4.921	4.073	327.9E6	2482.0E6	58.664	55.646
5)	MB Aldrin	5.263	4.358	324.7E6	2433.2E6	49.593	52.531
6)	B beta-BHC	4.511	4.017	129.6E6	1056.8E6	49.492	53.296
7)	B delta-BHC	4.759	4.253	337.4E6	2578.9E6	49.898	54.047
8)	B Heptachlo...	5.683	4.862	286.6E6	2263.5E6	49.505	53.085
9)	A Endosulfan I	6.067	5.236	272.2E6	2125.7E6	49.165	53.965
10)	B gamma-Chl...	5.938	5.115	289.7E6	2410.2E6	48.843	53.620
11)	B alpha-Chl...	6.019	5.179	294.7E6	2332.5E6	47.684	54.070
12)	B 4,4'-DDE	6.189	5.363	259.2E6	2374.7E6	48.661	54.169
13)	MA Dieldrin	6.339	5.501	290.1E6	2396.1E6	49.742	54.230
14)	MA Endrin	6.567	5.778	225.7E6	2003.5E6	45.552	49.476
15)	B Endosulfa...	6.780	6.070	242.9E6	2070.6E6	48.581	55.134
16)	A 4,4'-DDD	6.699	5.918	214.5E6	2028.8E6	52.903	55.677
17)	MA 4,4'-DDT	7.014	6.171	196.9E6	1884.2E6	47.899	53.723
18)	B Endrin al...	6.909	6.248	180.9E6	1551.9E6	49.378	56.041
19)	B Endosulfa...	7.143	6.471	222.8E6	1998.7E6	48.037	55.760
20)	A Methoxychlor	7.488	6.741	106.5E6	948.2E6	50.114	54.409
21)	B Endrin ke...	7.624	6.980	252.0E6	2270.8E6	51.854	60.204
22)	Mirex	8.106	7.173	150.3E6	1425.7E6	46.956	55.494

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111325\
Data File : PD091152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2025 14:06
Operator : AR\AJ
Sample : Q3609-13MSD
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
AU-713-COMP-05MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/14/2025
Supervised By :Sohil Jodhani 11/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 22:03:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	pd101725	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD090648.D	Endrin aldehyde	Abdul	10/19/2025 5:20:46 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090648.D	Endrin ketone #2	Abdul	10/19/2025 5:20:46 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
RESCHK	PD090649.D	Tetrachloro-m-xylene	Abdul	10/19/2025 5:20:37 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	alpha-Chlordane	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	Endrin ketone	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	Tetrachloro-m-xylene	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PTOXICV500	PD090667.D	Toxaphene-5 #2	Abdul	10/21/2025 8:49:16 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	4,4"-DDD	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	4,4"-DDD #2	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	Endrin aldehyde	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software

Manual Integration Report

Sequence:	pd111325	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD091142.D	4,4"-DDE #2	rahul	11/14/2025 8:20:27 AM	Sohil	11/14/2025 4:23:20	Peak Integrated by Software
PB170527BL	PD091144.D	Decachlorobiphenyl	rahul	11/14/2025 8:20:33 AM	Sohil	11/14/2025 4:23:26	Peak Integrated by Software
Q3609-13MS	PD091151.D	Heptachlor epoxide	rahul	11/14/2025 8:20:49 AM	Sohil	11/14/2025 4:23:56	Peak Integrated by Software
Q3609-13MS	PD091151.D	Tetrachloro-m-xylene	rahul	11/14/2025 8:20:49 AM	Sohil	11/14/2025 4:23:56	Peak Integrated by Software
Q3609-13MSD	PD091152.D	Tetrachloro-m-xylene	rahul	11/14/2025 8:20:52 AM	Sohil	11/14/2025 4:24:02	Peak Integrated by Software
Q3614-01	PD091153.D	Tetrachloro-m-xylene	rahul	11/14/2025 8:20:57 AM	Sohil	11/14/2025 4:24:07	Peak Integrated by Software
Q3614-01	PD091153.D	Tetrachloro-m-xylene #2	rahul	11/14/2025 8:20:57 AM	Sohil	11/14/2025 4:24:07	Peak Integrated by Software
Q3614-02	PD091154.D	Tetrachloro-m-xylene	rahul	11/14/2025 8:21:01 AM	Sohil	11/14/2025 4:24:13	Peak Integrated by Software
Q3614-02	PD091154.D	Tetrachloro-m-xylene #2	rahul	11/14/2025 8:21:01 AM	Sohil	11/14/2025 4:24:13	Peak Integrated by Software
Q3614-03	PD091155.D	Tetrachloro-m-xylene	rahul	11/14/2025 8:21:04 AM	Sohil	11/14/2025 4:24:20	Peak Integrated by Software

Manual Integration Report

Sequence:	PD111525	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD091194.D	4,4"-DDE #2	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PEM	PD091194.D	Endrin aldehyde	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PEM	PD091194.D	Endrin aldehyde #2	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PSTDICC005	PD091200.D	delta-BHC	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Heptachlor	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Heptachlor epoxide	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Tetrachloro-m-xylene	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Tetrachloro-m-xylene #2	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PEM	PD091215.D	4,4"-DDE #2	Abdul	11/17/2025 8:53:18 AM	Sohil	11/17/2025 2:14:03	Peak Integrated by Software

Manual Integration Report

Sequence:

pd111725

Instrument

ECD_d

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD091235.D	4,4"-DDD #2	Abdul	11/18/2025 11:03:29 AM	 	 	Peak Integrated by Software
PEM	PD091235.D	Endrin aldehyde	Abdul	11/18/2025 11:03:29 AM	 	 	Peak Integrated by Software
PEM	PD091235.D	Endrin ketone #2	Abdul	11/18/2025 11:03:29 AM	 	 	Peak Integrated by Software

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090646.D	17 Oct 2025 10:40	AR\AJ	Ok
2	I.BLK	PD090647.D	17 Oct 2025 10:53	AR\AJ	Ok
3	PEM	PD090648.D	17 Oct 2025 11:07	AR\AJ	Ok,M
4	RESCHK	PD090649.D	17 Oct 2025 11:20	AR\AJ	Ok,M
5	PSTDICC100	PD090650.D	17 Oct 2025 11:34	AR\AJ	Ok
6	PSTDICC075	PD090651.D	17 Oct 2025 11:48	AR\AJ	Ok
7	PSTDICC050	PD090652.D	17 Oct 2025 12:01	AR\AJ	Ok
8	PSTDICC025	PD090653.D	17 Oct 2025 12:15	AR\AJ	Ok
9	PSTDICC005	PD090654.D	17 Oct 2025 12:28	AR\AJ	Ok,M
10	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42	AR\AJ	Ok
11	PCHLORICC750	PD090656.D	17 Oct 2025 12:56	AR\AJ	Ok
12	PCHLORICC500	PD090657.D	17 Oct 2025 13:09	AR\AJ	Ok
13	PCHLORICC250	PD090658.D	17 Oct 2025 13:23	AR\AJ	Ok
14	PCHLORICC050	PD090659.D	17 Oct 2025 13:36	AR\AJ	Ok,M
15	PTOXICC1000	PD090660.D	17 Oct 2025 13:50	AR\AJ	Ok
16	PTOXICC750	PD090661.D	17 Oct 2025 14:03	AR\AJ	Ok
17	PTOXICC500	PD090662.D	17 Oct 2025 14:17	AR\AJ	Ok
18	PTOXICC250	PD090663.D	17 Oct 2025 14:30	AR\AJ	Ok
19	PTOXICC100	PD090664.D	17 Oct 2025 16:18	AR\AJ	Ok
20	PSTDICV050	PD090665.D	17 Oct 2025 16:46	AR\AJ	Ok
21	PCHLORICV500	PD090666.D	17 Oct 2025 17:50	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD090667.D	17 Oct 2025 18:42	AR\AJ	Ok,M
23	I.BLK	PD090668.D	17 Oct 2025 19:50	AR\AJ	Ok
24	PEM	PD090669.D	17 Oct 2025 20:04	AR\AJ	Ok,M
25	PSTDCCC050	PD090670.D	17 Oct 2025 20:18	AR\AJ	Ok
26	PB170142BL	PD090671.D	17 Oct 2025 20:31	AR\AJ	Ok
27	PB170142BS	PD090672.D	17 Oct 2025 20:45	AR\AJ	Ok
28	Q3366-01	PD090673.D	17 Oct 2025 20:59	AR\AJ	Ok,M
29	Q3366-03	PD090674.D	17 Oct 2025 21:12	AR\AJ	Ok,M
30	Q3367-01	PD090675.D	17 Oct 2025 21:26	AR\AJ	Ok,M
31	Q3369-01	PD090676.D	17 Oct 2025 21:40	AR\AJ	Ok,M
32	Q3369-05	PD090677.D	17 Oct 2025 21:53	AR\AJ	Ok,M
33	Q3363-01	PD090678.D	17 Oct 2025 22:48	AR\AJ	Ok,M
34	Q3363-02	PD090679.D	17 Oct 2025 23:02	AR\AJ	Ok,M
35	Q3363-03	PD090680.D	17 Oct 2025 23:15	AR\AJ	Ok,M
36	Q3363-04	PD090681.D	17 Oct 2025 23:29	AR\AJ	Ok,M
37	Q3363-04MS	PD090682.D	17 Oct 2025 23:43	AR\AJ	Ok,M
38	Q3363-04MSD	PD090683.D	17 Oct 2025 23:56	AR\AJ	Ok,M
39	Q3363-05	PD090684.D	18 Oct 2025 00:10	AR\AJ	Ok,M
40	Q3343-04MS	PD090685.D	18 Oct 2025 00:24	AR\AJ	Ok,M
41	Q3343-04MSD	PD090686.D	18 Oct 2025 00:37	AR\AJ	Ok
42	I.BLK	PD090687.D	18 Oct 2025 00:51	AR\AJ	Ok
43	PSTDCCC050	PD090688.D	18 Oct 2025 02:13	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111325

Review By	rahul	Review On	11/13/2025 1:53:42 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:24:51 PM
SubDirectory	PD111325	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD091140.D	13 Nov 2025 09:07	AR\AJ	Ok
2	I.BLK	PD091141.D	13 Nov 2025 09:20	AR\AJ	Ok
3	PEM	PD091142.D	13 Nov 2025 09:34	AR\AJ	Ok,M
4	PSTDCCC050	PD091143.D	13 Nov 2025 09:47	AR\AJ	Ok
5	PB170527BL	PD091144.D	13 Nov 2025 12:17	AR\AJ	Ok,M
6	PB170527BS	PD091145.D	13 Nov 2025 12:30	AR\AJ	Not Ok
7	Q3609-01	PD091146.D	13 Nov 2025 12:44	AR\AJ	Ok
8	Q3609-04	PD091147.D	13 Nov 2025 12:58	AR\AJ	Ok,M
9	Q3609-07	PD091148.D	13 Nov 2025 13:11	AR\AJ	Dilution
10	Q3609-10	PD091149.D	13 Nov 2025 13:25	AR\AJ	Ok,M
11	Q3609-13	PD091150.D	13 Nov 2025 13:39	AR\AJ	Ok,M
12	Q3609-13MS	PD091151.D	13 Nov 2025 13:52	AR\AJ	Ok,M
13	Q3609-13MSD	PD091152.D	13 Nov 2025 14:06	AR\AJ	Ok,M
14	Q3614-01	PD091153.D	13 Nov 2025 14:20	AR\AJ	Ok,M
15	Q3614-02	PD091154.D	13 Nov 2025 14:33	AR\AJ	Ok,M
16	Q3614-03	PD091155.D	13 Nov 2025 14:47	AR\AJ	Ok,M
17	Q3618-01	PD091156.D	13 Nov 2025 15:01	AR\AJ	Ok,M
18	Q3618-02	PD091157.D	13 Nov 2025 15:14	AR\AJ	Ok,M
19	Q3621-02	PD091158.D	13 Nov 2025 15:28	AR\AJ	Dilution
20	Q3624-01	PD091159.D	13 Nov 2025 15:42	AR\AJ	ReRun
21	I.BLK	PD091160.D	13 Nov 2025 16:19	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111325

Review By	rahul	Review On	11/13/2025 1:53:42 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:24:51 PM
SubDirectory	PD111325	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PSTDCCC050	PD091161.D	13 Nov 2025 16:32	AR\AJ	Ok
----	------------	------------	-------------------	-------	----

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	I.BLK	PD091193.D	14 Nov 2025 20:01	AR\AJ	Ok
2	PEM	PD091194.D	14 Nov 2025 20:15	AR\AJ	Ok,M
3	RESCHK	PD091195.D	14 Nov 2025 20:43	AR\AJ	Ok
4	PSTDICC100	PD091196.D	14 Nov 2025 20:56	AR\AJ	Ok
5	PSTDICC075	PD091197.D	14 Nov 2025 21:10	AR\AJ	Ok
6	PSTDICC050	PD091198.D	14 Nov 2025 21:24	AR\AJ	Ok
7	PSTDICC025	PD091199.D	14 Nov 2025 21:37	AR\AJ	Ok
8	PSTDICC005	PD091200.D	14 Nov 2025 21:51	AR\AJ	Ok,M
9	PCHLORICC1000	PD091201.D	14 Nov 2025 22:04	AR\AJ	Ok
10	PCHLORICC750	PD091202.D	14 Nov 2025 22:18	AR\AJ	Ok
11	PCHLORICC500	PD091203.D	14 Nov 2025 22:32	AR\AJ	Ok
12	PCHLORICC250	PD091204.D	14 Nov 2025 22:46	AR\AJ	Ok,M
13	PCHLORICC50	PD091205.D	14 Nov 2025 22:59	AR\AJ	Ok,M
14	PTOXICC1000	PD091206.D	14 Nov 2025 23:13	AR\AJ	Ok
15	PTOXICC750	PD091207.D	14 Nov 2025 23:27	AR\AJ	Ok
16	PTOXICC500	PD091208.D	14 Nov 2025 23:40	AR\AJ	Ok
17	PTOXICC250	PD091209.D	14 Nov 2025 23:54	AR\AJ	Ok
18	PTOXICC100	PD091210.D	15 Nov 2025 00:08	AR\AJ	Ok
19	PSTDICV050	PD091211.D	15 Nov 2025 00:21	AR\AJ	Ok
20	PCHLORICV050	PD091212.D	15 Nov 2025 00:49	AR\AJ	Ok
21	PTOXICV050	PD091213.D	15 Nov 2025 01:30	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	I.BLK	PD091214.D	15 Nov 2025 01:44	AR\AJ	Ok
23	PEM	PD091215.D	15 Nov 2025 01:57	AR\AJ	Ok,M
24	PSTDCCC050	PD091216.D	15 Nov 2025 02:11	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111725

Review By	Abdul	Review On	11/18/2025 11:05:20 AM
Supervise By		Supervise On	
SubDirectory	PD111725	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD091217.D	17 Nov 2025 09:35	AR\AJ	Ok
2	I.BLK	PD091218.D	17 Nov 2025 09:51	AR\AJ	Ok
3	PEM	PD091219.D	17 Nov 2025 10:04	AR\AJ	Ok
4	PSTDCCC050	PD091220.D	17 Nov 2025 10:18	AR\AJ	Ok
5	PB170527BS	PD091221.D	17 Nov 2025 12:54	AR\AJ	Ok
6	PB170506BS	PD091222.D	17 Nov 2025 13:33	AR\AJ	Ok
7	PB170575BL	PD091223.D	17 Nov 2025 13:46	AR\AJ	Ok
8	PB170575BS	PD091224.D	17 Nov 2025 14:00	AR\AJ	Ok
9	Q3638-01	PD091225.D	17 Nov 2025 14:14	AR\AJ	Ok,NS
10	Q3638-03	PD091226.D	17 Nov 2025 14:27	AR\AJ	Ok,NS
11	Q3640-02	PD091227.D	17 Nov 2025 14:41	AR\AJ	Ok,NS
12	Q3640-03	PD091228.D	17 Nov 2025 14:55	AR\AJ	Ok,NS
13	Q3641-01	PD091229.D	17 Nov 2025 15:08	AR\AJ	Ok,NS
14	Q3645-01	PD091230.D	17 Nov 2025 15:49	AR\AJ	Ok
15	Q3646-01	PD091231.D	17 Nov 2025 16:02	AR\AJ	Ok,NS
16	Q3647-01	PD091232.D	17 Nov 2025 16:16	AR\AJ	Ok,NS
17	Q3648-01	PD091233.D	17 Nov 2025 16:30	AR\AJ	Ok,NS
18	I.BLK	PD091234.D	17 Nov 2025 19:21	AR\AJ	Ok
19	PEM	PD091235.D	17 Nov 2025 19:35	AR\AJ	Ok,NS
20	PSTDCCC050	PD091236.D	17 Nov 2025 20:02	AR\AJ	Ok
21	Q3649-01	PD091237.D	17 Nov 2025 21:11	AR\AJ	Ok,NS

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111725

Review By	Abdul	Review On	11/18/2025 11:05:20 AM
Supervise By		Supervise On	
SubDirectory	PD111725	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3649-03	PD091238.D	17 Nov 2025 21:24	AR\AJ	Ok,NS
23	Q3649-03MS	PD091239.D	17 Nov 2025 21:38	AR\AJ	Ok,NS
24	Q3649-03MSD	PD091240.D	17 Nov 2025 21:52	AR\AJ	Ok,NS
25	Q3649-06	PD091241.D	17 Nov 2025 22:05	AR\AJ	Ok,NS
26	Q3649-08	PD091242.D	17 Nov 2025 22:19	AR\AJ	Ok,NS
27	I.BLK	PD091243.D	17 Nov 2025 22:33	AR\AJ	Ok
28	PSTDCCC050	PD091244.D	17 Nov 2025 23:00	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090646.D	17 Oct 2025 10:40		AR\AJ	Ok
2	I.BLK	I.BLK	PD090647.D	17 Oct 2025 10:53		AR\AJ	Ok
3	PEM	PEM	PD090648.D	17 Oct 2025 11:07		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD090649.D	17 Oct 2025 11:20		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PD090650.D	17 Oct 2025 11:34		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD090651.D	17 Oct 2025 11:48		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD090652.D	17 Oct 2025 12:01		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD090653.D	17 Oct 2025 12:15		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD090654.D	17 Oct 2025 12:28		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD090656.D	17 Oct 2025 12:56		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD090657.D	17 Oct 2025 13:09		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD090658.D	17 Oct 2025 13:23		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD090659.D	17 Oct 2025 13:36		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD090660.D	17 Oct 2025 13:50		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD090661.D	17 Oct 2025 14:03		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD090662.D	17 Oct 2025 14:17		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD090663.D	17 Oct 2025 14:30		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD090664.D	17 Oct 2025 16:18		AR\AJ	Ok
20	PSTDICV050	ICVPD101725	PD090665.D	17 Oct 2025 16:46		AR\AJ	Ok
21	PCHLORICV500	ICVPD101725CHLOR	PD090666.D	17 Oct 2025 17:50		AR\AJ	Ok
22	PTOXICV500	ICVPD101725TOX	PD090667.D	17 Oct 2025 18:42		AR\AJ	Ok,M
23	I.BLK	I.BLK	PD090668.D	17 Oct 2025 19:50		AR\AJ	Ok
24	PEM	PEM	PD090669.D	17 Oct 2025 20:04		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD090670.D	17 Oct 2025 20:18		AR\AJ	Ok
26	PB170142BL	PB170142BL	PD090671.D	17 Oct 2025 20:31		AR\AJ	Ok
27	PB170142BS	PB170142BS	PD090672.D	17 Oct 2025 20:45		AR\AJ	Ok
28	Q3366-01	TR-06-101625	PD090673.D	17 Oct 2025 20:59		AR\AJ	Ok,M
29	Q3366-03	TR-1-101625	PD090674.D	17 Oct 2025 21:12		AR\AJ	Ok,M
30	Q3367-01	BU-03-101625	PD090675.D	17 Oct 2025 21:26		AR\AJ	Ok,M
31	Q3369-01	MH-22	PD090676.D	17 Oct 2025 21:40		AR\AJ	Ok,M
32	Q3369-05	MH-21	PD090677.D	17 Oct 2025 21:53		AR\AJ	Ok,M
33	Q3363-01	PR132-SO1-085011-20	PD090678.D	17 Oct 2025 22:48		AR\AJ	Ok,M
34	Q3363-02	PR132-SO1-039085-20	PD090679.D	17 Oct 2025 23:02		AR\AJ	Ok,M
35	Q3363-03	PR132-SO3-028010-20	PD090680.D	17 Oct 2025 23:15		AR\AJ	Ok,M
36	Q3363-04	PR132-SO3-010013-20	PD090681.D	17 Oct 2025 23:29		AR\AJ	Ok,M
37	Q3363-04MS	PR132-SO3-010013-20	PD090682.D	17 Oct 2025 23:43		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q3363-04MSD	PR132-SO3-010013-20	PD090683.D	17 Oct 2025 23:56		AR\AJ	Ok,M
39	Q3363-05	COMP01-20251015	PD090684.D	18 Oct 2025 00:10		AR\AJ	Ok,M
40	Q3343-04MS	SOIL-ROLLOFF-WC-11	PD090685.D	18 Oct 2025 00:24		AR\AJ	Ok,M
41	Q3343-04MSD	SOIL-ROLLOFF-WC-11	PD090686.D	18 Oct 2025 00:37		AR\AJ	Ok
42	I.BLK	I.BLK	PD090687.D	18 Oct 2025 00:51		AR\AJ	Ok
43	PSTDCCC050	PSTDCCC050	PD090688.D	18 Oct 2025 02:13		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111325

Review By	rahul	Review On	11/13/2025 1:53:42 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:24:51 PM
SubDirectory	PD111325	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091140.D	13 Nov 2025 09:07		AR\AJ	Ok
2	I.BLK	I.BLK	PD091141.D	13 Nov 2025 09:20		AR\AJ	Ok
3	PEM	PEM	PD091142.D	13 Nov 2025 09:34		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD091143.D	13 Nov 2025 09:47		AR\AJ	Ok
5	PB170527BL	PB170527BL	PD091144.D	13 Nov 2025 12:17		AR\AJ	Ok,M
6	PB170527BS	PB170527BS	PD091145.D	13 Nov 2025 12:30	Endosulfan recovery fail	AR\AJ	Not Ok
7	Q3609-01	AU-713-COMP-01	PD091146.D	13 Nov 2025 12:44		AR\AJ	Ok
8	Q3609-04	AU-713-COMP-02	PD091147.D	13 Nov 2025 12:58		AR\AJ	Ok,M
9	Q3609-07	AU-713-COMP-03	PD091148.D	13 Nov 2025 13:11	Need 20X	AR\AJ	Dilution
10	Q3609-10	AU-713-COMP-04	PD091149.D	13 Nov 2025 13:25		AR\AJ	Ok,M
11	Q3609-13	AU-713-COMP-05	PD091150.D	13 Nov 2025 13:39		AR\AJ	Ok,M
12	Q3609-13MS	AU-713-COMP-05MS	PD091151.D	13 Nov 2025 13:52		AR\AJ	Ok,M
13	Q3609-13MSD	AU-713-COMP-05MSD	PD091152.D	13 Nov 2025 14:06		AR\AJ	Ok,M
14	Q3614-01	COMP-1	PD091153.D	13 Nov 2025 14:20		AR\AJ	Ok,M
15	Q3614-02	COMP-2	PD091154.D	13 Nov 2025 14:33		AR\AJ	Ok,M
16	Q3614-03	COMP-3	PD091155.D	13 Nov 2025 14:47		AR\AJ	Ok,M
17	Q3618-01	EME-TOP-SOIL	PD091156.D	13 Nov 2025 15:01		AR\AJ	Ok,M
18	Q3618-02	EME-GENERAL-FILL	PD091157.D	13 Nov 2025 15:14		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111325

Review By	rahul	Review On	11/13/2025 1:53:42 PM
Supervise By	Sohil	Supervise On	11/14/2025 4:24:51 PM
SubDirectory	PD111325	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3621-02	VNJ-231	PD091158.D	13 Nov 2025 15:28	Need 10X	AR\AJ	Dilution
20	Q3624-01	OK-02-11122025	PD091159.D	13 Nov 2025 15:42	DCB high in both column	AR\AJ	ReRun
21	I.BLK	I.BLK	PD091160.D	13 Nov 2025 16:19		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PD091161.D	13 Nov 2025 16:32		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	I.BLK	I.BLK	PD091193.D	14 Nov 2025 20:01		AR\AJ	Ok
2	PEM	PEM	PD091194.D	14 Nov 2025 20:15		AR\AJ	Ok,M
3	RESCHK	RESCHK	PD091195.D	14 Nov 2025 20:43		AR\AJ	Ok
4	PSTDICC100	PSTDICC100	PD091196.D	14 Nov 2025 20:56		AR\AJ	Ok
5	PSTDICC075	PSTDICC075	PD091197.D	14 Nov 2025 21:10		AR\AJ	Ok
6	PSTDICC050	PSTDICC050	PD091198.D	14 Nov 2025 21:24		AR\AJ	Ok
7	PSTDICC025	PSTDICC025	PD091199.D	14 Nov 2025 21:37		AR\AJ	Ok
8	PSTDICC005	PSTDICC005	PD091200.D	14 Nov 2025 21:51		AR\AJ	Ok,M
9	PCHLORICC1000	PCHLORICC1000	PD091201.D	14 Nov 2025 22:04		AR\AJ	Ok
10	PCHLORICC750	PCHLORICC750	PD091202.D	14 Nov 2025 22:18		AR\AJ	Ok
11	PCHLORICC500	PCHLORICC500	PD091203.D	14 Nov 2025 22:32		AR\AJ	Ok
12	PCHLORICC250	PCHLORICC250	PD091204.D	14 Nov 2025 22:46		AR\AJ	Ok,M
13	PCHLORICC50	PCHLORICC50	PD091205.D	14 Nov 2025 22:59		AR\AJ	Ok,M
14	PTOXICC1000	PTOXICC1000	PD091206.D	14 Nov 2025 23:13		AR\AJ	Ok
15	PTOXICC750	PTOXICC750	PD091207.D	14 Nov 2025 23:27		AR\AJ	Ok
16	PTOXICC500	PTOXICC500	PD091208.D	14 Nov 2025 23:40		AR\AJ	Ok
17	PTOXICC250	PTOXICC250	PD091209.D	14 Nov 2025 23:54		AR\AJ	Ok
18	PTOXICC100	PTOXICC100	PD091210.D	15 Nov 2025 00:08		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PSTDICV050	ICVPD111525	PD091211.D	15 Nov 2025 00:21		AR\AJ	Ok
20	PCHLORICV050	ICVPD111525CHLOR	PD091212.D	15 Nov 2025 00:49		AR\AJ	Ok
21	PTOXICV050	ICVPD111525TOX	PD091213.D	15 Nov 2025 01:30		AR\AJ	Ok
22	I.BLK	I.BLK	PD091214.D	15 Nov 2025 01:44		AR\AJ	Ok
23	PEM	PEM	PD091215.D	15 Nov 2025 01:57		AR\AJ	Ok,M
24	PSTDCCC050	PSTDCCC050	PD091216.D	15 Nov 2025 02:11		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111725

Review By	Abdul	Review On	11/18/2025 11:05:20 AM
Supervise By		Supervise On	
SubDirectory	PD111725	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091217.D	17 Nov 2025 09:35		AR\AJ	Ok
2	I.BLK	I.BLK	PD091218.D	17 Nov 2025 09:51		AR\AJ	Ok
3	PEM	PEM	PD091219.D	17 Nov 2025 10:04		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PD091220.D	17 Nov 2025 10:18		AR\AJ	Ok
5	PB170527BS	PB170527BS	PD091221.D	17 Nov 2025 12:54		AR\AJ	Ok
6	PB170506BS	PB170506BS	PD091222.D	17 Nov 2025 13:33		AR\AJ	Ok
7	PB170575BL	PB170575BL	PD091223.D	17 Nov 2025 13:46		AR\AJ	Ok
8	PB170575BS	PB170575BS	PD091224.D	17 Nov 2025 14:00		AR\AJ	Ok
9	Q3638-01	MOO-25-0327	PD091225.D	17 Nov 2025 14:14		AR\AJ	Ok,NS
10	Q3638-03	MOO-25-0328	PD091226.D	17 Nov 2025 14:27		AR\AJ	Ok,NS
11	Q3640-02	LAW-25-0179-0180	PD091227.D	17 Nov 2025 14:41		AR\AJ	Ok,NS
12	Q3640-03	LAW-25-0182	PD091228.D	17 Nov 2025 14:55	TCMX low in 2nd column	AR\AJ	Ok,NS
13	Q3641-01	TR-05-111325	PD091229.D	17 Nov 2025 15:08		AR\AJ	Ok,NS
14	Q3645-01	HD-02-11142025	PD091230.D	17 Nov 2025 15:49		AR\AJ	Ok
15	Q3646-01	SU-03-11142025	PD091231.D	17 Nov 2025 16:02		AR\AJ	Ok,NS
16	Q3647-01	VNJ-212	PD091232.D	17 Nov 2025 16:16		AR\AJ	Ok,NS
17	Q3648-01	OR-02-111425	PD091233.D	17 Nov 2025 16:30		AR\AJ	Ok,NS
18	I.BLK	I.BLK	PD091234.D	17 Nov 2025 19:21		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111725

Review By	Abdul	Review On	11/18/2025 11:05:20 AM
Supervise By		Supervise On	
SubDirectory	PD111725	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PEM	PEM	PD091235.D	17 Nov 2025 19:35		AR\AJ	Ok,NS
20	PSTDCCC050	PSTDCCC050	PD091236.D	17 Nov 2025 20:02		AR\AJ	Ok
21	Q3649-01	B5-0.0-0.5-20251114	PD091237.D	17 Nov 2025 21:11		AR\AJ	Ok,NS
22	Q3649-03	DUPE-0.0-0.5-20251114	PD091238.D	17 Nov 2025 21:24		AR\AJ	Ok,NS
23	Q3649-03MS	DUPE-0.0-0.5-20251114	PD091239.D	17 Nov 2025 21:38		AR\AJ	Ok,NS
24	Q3649-03MSD	DUPE-0.0-0.5-20251114	PD091240.D	17 Nov 2025 21:52		AR\AJ	Ok,NS
25	Q3649-06	B4-0.0-0.5-20251114	PD091241.D	17 Nov 2025 22:05		AR\AJ	Ok,NS
26	Q3649-08	B3-0.0-0.5-20251114	PD091242.D	17 Nov 2025 22:19		AR\AJ	Ok,NS
27	I.BLK	I.BLK	PD091243.D	17 Nov 2025 22:33		AR\AJ	Ok
28	PSTDCCC050	PSTDCCC050	PD091244.D	17 Nov 2025 23:00		AR\AJ	Ok

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/13/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 11/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 11/13/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137869

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3609-01	AU-713-COMP-01	1	1.15	10.53	11.68	9.74	81.6	
Q3609-02	AU-713-VOC-01	2	1.18	10.57	11.75	10.37	86.9	
Q3609-04	AU-713-COMP-02	3	1.19	10.53	11.72	10.1	84.6	
Q3609-05	AU-713-VOC-02	4	1.15	10.68	11.83	9.23	75.7	
Q3609-07	AU-713-COMP-03	5	1.19	10.33	11.52	10.00	85.3	
Q3609-08	AU-713-VOC-03	6	1.16	10.90	12.06	10.44	85.1	
Q3609-10	AU-713-COMP-04	7	1.18	10.34	11.52	9.51	80.6	
Q3609-11	AU-713-VOC-04	8	1.13	10.46	11.59	10.16	86.3	
Q3609-13	AU-713-COMP-05	9	1.14	10.49	11.63	10.05	84.9	
Q3609-14	AU-713-VOC-05	10	1.13	10.40	11.53	10.00	85.3	
Q3609-15	AU-713-COMP-05	11	1.14	10.49	11.63	10.05	84.9	
Q3610-01	AU-713-TPH-01	12	1.18	10.35	11.53	10.37	88.8	
Q3610-02	AU-713-TPH-02	13	1.19	10.50	11.69	9.48	79.0	
Q3610-03	AU-713-TPH-03	14	1.11	10.33	11.44	9.75	83.6	
Q3610-04	AU-713-TPH-04	15	1.18	10.42	11.6	9.1	76.0	
Q3610-05	AU-713-TPH-05	16	1.18	10.56	11.74	10.33	86.6	
Q3610-06	AU-713-TPH-06	17	1.19	10.37	11.56	10.03	85.2	
Q3610-07	AU-713-TPH-07	18	1.15	10.40	11.55	9.08	76.3	
Q3610-08	AU-713-TPH-08	19	1.19	10.43	11.62	10.04	84.9	
Q3610-09	AU-713-TPH-09	20	1.18	10.71	11.89	9.9	81.4	
Q3610-10	AU-713-TPH-10	21	1.16	10.24	11.4	9.96	85.9	
Q3610-11	AU-713-TPH-11	22	1.14	9.99	11.13	8.96	78.3	
Q3610-12	AU-713-TPH-12	23	1.19	10.49	11.68	9.99	83.9	
Q3610-13	AU-713-TPH-13	24	1.15	10.84	11.99	10.4	85.3	
Q3610-14	AU-713-TPH-14	25	1.18	10.50	11.68	10.44	88.2	
Q3610-15	AU-713-TPH-15	26	1.14	10.48	11.62	10.34	87.8	
Q3611-01	AU-713-01	27	1.13	10.54	11.67	9.83	82.5	
Q3611-02	AU-713-02	28	1.12	11.27	12.39	9.7	76.1	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/13/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 11/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 11/13/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137869

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3611-03	AU-713-03	29	1.12	10.07	11.19	9.69	85.1	
Q3611-04	AU-713-04	30	1.16	10.64	11.8	9.27	76.2	
Q3611-05	AU-713-05	31	1.16	10.31	11.47	10.55	91.1	
Q3611-06	AU-713-06	32	1.18	10.70	11.88	10.14	83.7	
Q3611-07	AU-713-07	33	1.15	10.28	11.43	9.16	77.9	
Q3611-08	AU-713-08	34	1.19	10.71	11.9	10.05	82.7	
Q3611-09	AU-713-09	35	1.19	10.66	11.85	9.64	79.3	
Q3611-10	AU-713-10	36	1.17	10.13	11.3	9.79	85.1	
Q3611-11	AU-713-11	37	1.13	11.03	12.16	9.74	78.1	
Q3611-12	AU-713-12	38	1.11	10.13	11.24	9.17	79.6	
Q3611-13	AU-713-13	39	1.13	10.59	11.72	9.54	79.4	
Q3611-14	AU-713-14	40	1.16	10.23	11.39	10.02	86.6	
Q3611-15	AU-713-15	41	1.16	10.44	11.6	10.08	85.4	
Q3614-01	COMP-1	42	1.12	10.94	12.06	10.12	82.3	
Q3614-02	COMP-2	43	1.18	10.44	11.62	9.71	81.7	
Q3614-03	COMP-3	44	1.18	10.30	11.48	9.78	83.5	
Q3615-01	East Bathhouse-Concrete	45	1.00	1.00	2.00	2.00	100.0	Concreate sample
Q3618-01	EME-TOP-SOIL	46	1.18	10.33	11.51	8.44	70.3	
Q3618-02	EME-GENERAL-FILL	47	1.19	10.38	11.57	10.34	88.2	
Q3621-01	VNJ-231	48	1.16	10.38	11.54	11.06	95.4	
Q3621-02	VNJ-231	49	1.19	10.21	11.4	10.7	93.1	
Q3622-01	RT3071-SOLID	50	1.15	10.44	11.59	8.75	72.8	
Q3623-01	RBR200007-SOIL	51	1.13	10.84	11.97	10.5	86.4	
Q3623-02	RBR200007-SILICA	52	1.13	10.86	11.99	3.71	23.8	sludge sample
Q3624-01	OK-02-11122025	53	1.13	10.91	12.04	10.54	86.3	
Q3624-02	OK-02-11122025-E2	54	1.16	10.34	11.5	10.12	86.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

UP 137869

WorkList Name : %1-111225

WorkList ID : 193056

Department : Wet-Chemistry

Date : 11-12-2025 07:31:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3609-01	AU-713-COMP-01	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-02	AU-713-VOC-01	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-04	AU-713-COMP-02	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-05	AU-713-VOC-02	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-07	AU-713-COMP-03	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-08	AU-713-VOC-03	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-10	AU-713-COMP-04	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-11	AU-713-VOC-04	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-13	AU-713-COMP-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-14	AU-713-VOC-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3609-15	AU-713-COMP-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-01	AU-713-TPH-01	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-02	AU-713-TPH-02	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-03	AU-713-TPH-03	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-04	AU-713-TPH-04	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-05	AU-713-TPH-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-06	AU-713-TPH-06	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-07	AU-713-TPH-07	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-08	AU-713-TPH-08	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-09	AU-713-TPH-09	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-10	AU-713-TPH-10	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO

Date/Time 11-12-25 15:30

Raw Sample Received by: JP WCI

Raw Sample Relinquished by: JP WCI

Date/Time 11-12-25

Raw Sample Received by: JP WCI

Raw Sample Relinquished by: JP WCI

WORKLIST(Hardcopy Internal Chain)

137869

WorkList Name : %1-111225

WorkList ID : 193056

Department : Wet-Chemistry

Date : 11-12-2025 07:31:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3610-11	AU-713-TPH-11	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-12	AU-713-TPH-12	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-13	AU-713-TPH-13	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-14	AU-713-TPH-14	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3610-15	AU-713-TPH-15	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/11/2025	Chemtech -SO
Q3611-01	AU-713-01	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-02	AU-713-02	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-03	AU-713-03	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-04	AU-713-04	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-05	AU-713-05	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-06	AU-713-06	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-07	AU-713-07	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-08	AU-713-08	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-09	AU-713-09	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-10	AU-713-10	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-11	AU-713-11	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-12	AU-713-12	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-13	AU-713-13	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-14	AU-713-14	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3611-15	AU-713-15	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	11/12/2025	Chemtech -SO
Q3614-01	COMP-1	Solid	Percent Solids	Cool 4 deg C	POWE02	J21	11/11/2025	Chemtech -SO

Date/Time 11-12-25 13:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11-12-25 17:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

137869

WorkList Name : %1-111225

WorkList ID : 193056

Department : Wet-Chemistry

Date : 11-12-2025 07:31:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3614-02	COMP-2	Solid	Percent Solids	Cool 4 deg C	POWE02	J21	11/11/2025	Chemtech -SO
Q3614-03	COMP-3	Solid	Percent Solids	Cool 4 deg C	POWE02	J21	11/11/2025	Chemtech -SO
Q3615-01	East Bathroom- Concrete	Solid	Percent Solids	Cool 4 deg C	CORE02	J32	11/12/2025	Chemtech -SO
Q3618-01	EME-TOP-SOIL	Solid	Percent Solids	Cool 4 deg C	ENTA05	D41	11/12/2025	Chemtech -SO
Q3618-02	EME-GENERAL-FILL	Solid	Percent Solids	Cool 4 deg C	ENTA05	D41	11/12/2025	Chemtech -SO
Q3621-01	VNJ-231	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3621-02	VNJ-231	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3622-01	RT3071-SOLID	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3623-01	RBR200007-SOIL	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3623-02	RBR200007-SILICA	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/12/2025	Chemtech -SO
Q3624-01	OK-02-11122025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/12/2025	Chemtech -SO
Q3624-02	OK-02-11122025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/12/2025	Chemtech -SO

Date/Time 11-12-25 15:30

Raw Sample Received by: SO (WC)

Raw Sample Relinquished by: [Signature]

Date/Time 11-12-25

Raw Sample Received by: CP [Signature]

Raw Sample Relinquished by: [Signature]

1740

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Florisil

Matrix : Solid

Weigh By: EH

Balance check: EH

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 11/13/2025

Extraction Start Time : 08:50

Extraction End Date : 11/13/2025

Extraction End Time : 12:00

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24766
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2660
Baked Na2SO4	N/A	EP2655
Sand	N/A	E3951
Hexane	N/A	E3981
Florisil	N/A	E3976
9:1 Hexane:Acetone Mixture	N/A	EP2644
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: N-EVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/13/25	RS (Ext Lab)	Y-P.P. 9711/25
12:05	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 11/13/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170527BL	PBLK527	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10			U3-1
PB170527BS	PLCS527	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10			2
Q3609-01	AU-713-COMP-01	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	E		3
Q3609-04	AU-713-COMP-02	Pesticide-TCL	30.08	N/A	ritesh	Evelyn	10	E		4
Q3609-07	AU-713-COMP-03	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	E		5
Q3609-10	AU-713-COMP-04	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10	E		6
Q3609-13	AU-713-COMP-05	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	E		U4-1
Q3609-13MS	AU-713-COMP-05MS	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	E		2
Q3609-13MS D	AU-713-COMP-05MSD	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	E		3
Q3614-01	COMP-1	PESTICIDE Group1	30.09	N/A	ritesh	Evelyn	10	E		4
Q3614-02	COMP-2	PESTICIDE Group1	30.05	N/A	ritesh	Evelyn	10	E		5
Q3614-03	COMP-3	PESTICIDE Group1	30.03	N/A	ritesh	Evelyn	10	E		6
Q3618-01	EME-TOP-SOIL	Pesticide-TCL	30.08	N/A	ritesh	Evelyn	10	E		U6-1
Q3618-02	EME-GENERAL-FILL	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	F		2
Q3621-02	VNJ-231	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	B		3
Q3624-01	OK-02-11122025	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	E		4

RS
11/13

170527x
8:50

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3614 WorkList ID : 193097 Department : Extraction Date : 11-13-2025 08:41:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3609-01	AU-713-COMP-01	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/11/2025	8081B
Q3609-04	AU-713-COMP-02	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/11/2025	8081B
Q3609-07	AU-713-COMP-03	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/11/2025	8081B
Q3609-10	AU-713-COMP-04	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/11/2025	8081B
Q3609-13	AU-713-COMP-05	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/11/2025	8081B
Q3614-01	COMP-1	Solid	PESTICIDE Group1	Cool 4 deg C	POWE02	J21	11/11/2025	8081B
Q3614-02	COMP-2	Solid	PESTICIDE Group1	Cool 4 deg C	POWE02	J21	11/11/2025	8081B
Q3614-03	COMP-3	Solid	PESTICIDE Group1	Cool 4 deg C	POWE02	J21	11/11/2025	8081B
Q3618-01	EME-TOP-SOIL	Solid	Pesticide-TCL	Cool 4 deg C	ENTA05	D41	11/12/2025	8081B
Q3618-02	EME-GENERAL-FILL	Solid	Pesticide-TCL	Cool 4 deg C	ENTA05	D41	11/12/2025	8081B
Q3621-02	VNJ-231	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/12/2025	8081B
Q3624-01	OK-02-11122025	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D41	11/12/2025	8081B

Date/Time 11/13/25 8:45
Raw Sample Received by: RJ (Ext 66)
Raw Sample Relinquished by: [Signature]

Date/Time 11/12/25 9:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: RJ (Ext 66)

Prep Standard - Chemical Standard Summary

Order ID : Q3614

Test : PESTICIDE Group1

Prepbatch ID : PB170527,

Sequence ID/Qc Batch ID: pd111325,pd111725,

Standard ID :

EP2655,EP2660,PP24651,PP24663,PP24738,PP24739,PP24740,PP24741,PP24742,PP24744,PP24745,PP24746,PP24747,PP24748,PP24749,PP24750,PP24751,PP24752,PP24753,PP24754,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764,PP24766,PP24942,

Chemical ID :

E3875,E3941,E3944,E3951,E3955,E3956,E3971,E3976,E3981,P12604,P12610,P13038,P13041,P13196,P13246,P13774,P13780,P13786,P13788,P13862,P9053,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2655	10/24/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/24/2025
<u>FROM</u> 4000.00000gram of E3875 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2660	11/03/2025	05/03/2026	RUPESHKUMAR SHAH	None	None	Riteshkumar Patel
<u>FROM</u>								

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4027	Pesticide resolution Check Mixture 8081	PP24651	06/16/2025	12/11/2025	Abdul Mirza	None	None	Yogesh Patel
								07/22/2025

FROM 1.00000ml of P13246 + 99.00000ml of E3941 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel
								07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3629	20 PPM PEST stock Solution 1st source(RESTEK)	PP24738	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13038 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP24739	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13041 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1273	20 PPM Mirex Stock (Primary Source)	PP24740	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P9053 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP24741	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13196 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24742	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13788 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3630	100/100 PPB PEST Working std.1st Source(RESTEK)	PP24744	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 98.50000ml of E3956 + 0.50000ml of PP24738 + 0.50000ml of PP24740 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
80	100/100 PPB Pesticide Working Solution 2nd Source	PP24745	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 98.50000ml of E3956 + 0.50000ml of PP24739 + 0.50000ml of PP24741 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
386	1000/100 PPB Chlordane STD (Restek)	PP24746	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12604 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3746	1000/100 ppb Chlordane STD-RESTEK 2ND SOURCE	PP24747	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12610 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
383	1000/100 PPB Toxaphene STD (Restek)	PP24748	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13774 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3669	1000/100 PPB TOXAPHENE STD 2nd source (RESTEK)	PP24749	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13862 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3631	75 PPB ICAL PEST STD(RESTEK)	PP24750	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.25000ml of E3956 + 0.75000ml of PP24744 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3632	50 PPB ICAL PEST STD(RESTEK)	PP24751	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24744 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3633	25 PPB ICAL PEST STD(RESTEK)	PP24752	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.75000ml of E3956 + 0.25000ml of PP24744 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3634	5 PPB ICAL PEST STD(RESTEK)	PP24753	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.90000ml of E3956 + 0.10000ml of PP24751 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3988	50 PPB PEST ICV STD(RESTEK)	PP24754	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24745 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
528	CHLOR 750 PPB STD	PP24755	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.25000ml of E3956 + 0.75000ml of PP24746 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
529	CHLOR 500 PPB STD	PP24756	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24746 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
530	CHLOR 250 PPB STD	PP24757	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.75000ml of E3956 + 0.25000ml of PP24746 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3408	CHLOR 50 PPB STD	PP24758	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.90000ml of E3956 + 0.10000ml of PP24756 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
532	CHLOR 500 PPB ICV STD	PP24759	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24747 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
533	TOX 750 PPB STD	PP24760	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.25000ml of E3956 + 0.75000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
534	TOX 500 PPB STD	PP24761	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
535	TOX 250 PPB STD	PP24762	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.75000ml of E3956 + 0.25000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2217	TOX 100 PPB STD	PP24763	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.90000ml of E3956 + 0.10000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3670	TOX 500 PPB ICV std (RESTEK)	PP24764	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of PP24749 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
79	500 PPB Pesticide Spike Solution	PP24766	07/28/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								08/19/2025

FROM 95.00000ml of E3955 + 2.50000ml of PP24739 + 2.50000ml of PP24741 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
758	PEM Mix w/Surr	PP24942	09/24/2025	03/24/2026	Abdul Mirza	None	None	Yogesh Patel
								10/10/2025

FROM 1.00000ml of P13780 + 99.00000ml of E3971 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	07/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	12/11/2025	06/11/2025 / Rajesh	06/04/2025 / Rajesh	E3941

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	25A2756718	12/31/2028	07/09/2025 / RUPESH	04/28/2020 / RUPESH	E3951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	04/18/2027	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3956

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	09/15/2025 / Evelyn	09/04/2025 / Riteshkumar	E3971

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agela Technologies Inc.	FS0006 / Cleanert Florisil cartridge	M07456	04/03/2026	10/03/2025 / Evelyn	10/03/2025 / Evelyn	E3976

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/30/2026	10/27/2025 / RUPESH	10/22/2025 / RUPESH	E3981

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12604

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12610

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	01/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13038

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0199099	07/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13041

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	042022	01/21/2026	07/21/2025 / Abdul	01/17/2024 / Abdul	P13196

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	19161 / 8081 pesticide resolution check mixture	013124	12/17/2025	06/17/2025 / Abdul	02/09/2024 / Abdul	P13246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203038	01/21/2026	07/21/2025 / Abdul	05/03/2024 / Ankita	P13774

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32074 / Pesticide Performance Evaluation Mix w/Surrogate	A0213999	03/24/2026	09/24/2025 / Abdul	11/19/2024 / Ankita	P13780

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	01/21/2026	07/21/2025 / Abdul	11/19/2024 / Ankita	P13788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0210240	01/21/2026	07/21/2025 / Abdul	12/09/2024 / Abdul	P13862

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	112018	01/21/2026	07/21/2025 / Abdul	11/01/2019 / Stephen	P9053



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian

Recd. by RI on 6/11/25

E3941

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Material	BDH9274-2.5KG
Material Description	BDH SAND STDD OTTAWA W+I 2.5KG
Grade	NOT APPLICABLE
Batch	25A2756718
Reassay Date	12/31/2028
CAS Number	14808-60-7
Molecular Formula	SiO ₂
Molecular Mass	60.09
Date of Manufacture	12/05/2024
Storage	Room Temperature

Characteristics	Specifications	Measured Values
Appearance	Beige granules.	Beige granules.
Moisture	<= 0.1 %	0.1 %
Particle Size 30-40 mesh	>= 80 %	99 %
CUSTOMER PART # BDH9274-2.5KG		

Received on 1/12/25.

E3951

Internal ID #: 793

Signature

We certify that this batch conforms to the specifications listed above.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits

Product meets analytical specifications of the grades listed.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E 3955

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors**[™]



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E3956



Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3971

Jamie Croak
Director Quality Operations, Bioscience Production

Cleanert Florisil

1g/6ml 30/pkg

LOT#:M07456

MFG#:G02147

CAT#FS0006



固相萃取产品

Made in China

Agela Technologies

E 3976



n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™



Material No.: 9262-03

Batch No.: 25C0362006

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	4
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Received on 10/22/25

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3981

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291

Lot No.: A0199099

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : June 30, 2027

Storage: 10°C or colder

Ship: Ambient

P130397 5
↓
P13043 1
✓
12-26-2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.0 µg/mL	+/- 8.9732
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	200.1 µg/mL	+/- 8.9762
3	beta-BHC	319-85-7	BCCC6425	99%	200.3 µg/mL	+/- 8.9844
4	delta-BHC	319-86-8	14450800	98%	200.0 µg/mL	+/- 8.9740
5	Heptachlor	76-44-8	813251	99%	200.1 µg/mL	+/- 8.9754
6	Aldrin	309-00-2	14389400	98%	200.0 µg/mL	+/- 8.9718
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.1 µg/mL	+/- 8.9754
8	trans-Chlordane	5103-74-2	32943	98%	199.9 µg/mL	+/- 8.9696
9	cis-Chlordane	5103-71-9	31766	98%	200.1 µg/mL	+/- 8.9762
10	Endosulfan I	959-98-8	BCCF4060	99%	200.1 µg/mL	+/- 8.9754
11	4,4'-DDE	72-55-9	GHYQG	99%	200.1 µg/mL	+/- 8.9777
12	Dieldrin	60-57-1	11129900	98%	200.0 µg/mL	+/- 8.9718
13	Endrin	72-20-8	14123200	98%	199.9 µg/mL	+/- 8.9696
14	4,4'-DDD	72-54-8	HAN02	99%	200.1 µg/mL	+/- 8.9777
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	200.0 µg/mL	+/- 8.9718

17	Endrin aldehyde	7421-93-4	30720	98%	200.1	µg/mL	+/- 8.9784
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.0	µg/mL	+/- 8.9732
19	Methoxychlor	72-43-5	13668200	99%	200.1	µg/mL	+/- 8.9777
20	Endrin ketone	53494-70-5	1-ABS-16-7	98%	200.0	µg/mL	+/- 8.9740

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13039
 ↓
 P13043
 5
 1
 JAW
 12/26/23

Quality Confirmation Test

Column:
 30m x .25mm x .2um
 Rtx-CLP II (cat.# 11323)

Carrier Gas:
 helium-constant pressure 20 psi.

Temp. Program:
 150°C to 300°C
 @ 4°C/min. (hold 5 min.)


Inj. Temp:
 200°C

Det. Temp:
 300°C

Det. Type:
 ECD

Split Vent:
 Split ratio 50:1

Inj. Vol
 1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
 Josh McCloskey - Operations Technician I

Date Mixed: 19-Jun-2023

Balance Serial # 1128360905

Jennifer Pollino
 Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

P 12603
↓
P 12605
[3]
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician


Date Mixed: 11-May-2023

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 12603 } (3)
↓
P 12605

7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *chromatographic plus*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32021 Lot No.: A0197993
Description: Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size: 2 mL Pkg Amt: > 1 mL
Expiration Date: August 31, 2029 Storage: 10°C or colder
Ship: Ambient

Handwritten: P 12606
P 12610
5 Five
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. { hold 10 min.}

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 11-May-2023

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291

Lot No.: A0200423

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : July 31, 2027

Storage: 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number:	19161		
Lot Number:	013124		
Description:	CLP Pesticides & PCBs Resolution Check Standard		
Expiration Date:	013129	Solvent(s):	Lot#
Recommended Storage:	Refrigerate (4 °C)	Hexane	273615 (50%)
Nominal Concentration (µg/mL):	Varied	Toluene	28508 (50%)
NIST Test ID#:	6UTB		
Volume(s) shown below were combined and diluted to (mL):	100.0	5E-05	Balance Uncertainty
		0.021	Pipet Uncertainty

Formulated By:	Lawrence Barry
	013124
	DATE
Reviewed By:	Pedro L. Rentes
	013124
	DATE

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty (mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (±) µg/mL	CAS#	SDS Information	
1. trans-Chlordane	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	5103-74-2	0.5mg/m3 (skin)	or-rat 500mg/kg
2. Endosulfan I	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	959-98-8	0.1mg/m3 (skin)	or-rat 18mg/kg
3. 4,4'-DDE	19361	013124	0.010	1.00	0.004	201.6	2.0	0.03	72-55-9	N/A	or-rat 880mg/kg
4. Dieldrin	19361	013124	0.010	1.00	0.004	202.8	2.0	0.03	60-57-1	0.25mg/m3 (skin)	or-rat 38300µg/kg
5. Endosulfan sulfate	19361	013124	0.010	1.00	0.004	204.2	2.0	0.03	1031-07-8	N/A	or-rat 18mg/kg
6. Endrin ketone	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	53494-70-5	N/A	N/A
7. 4,4-Methoxychlor	19361	013124	0.010	1.00	0.004	1000.7	10.0	0.09	72-43-5	10mg/m3	or-rat 6000mg/kg
8. 2,4,5,6-Tetrachloro-m-xylene	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	877-09-8	N/A	N/A
9. Decachlorobiphenyl (209)	19361	013124	0.010	1.00	0.004	202.0	2.0	0.03	2051-24-3	N/A	N/A

P13243
P13244
(5)

P13244
P13245
P13246
P13247
P13248
P13249
P13250
P13251
P13252
P13253
P13254
P13255
P13256
P13257
P13258
P13259
P13260
P13261
P13262
P13263
P13264
P13265
P13266
P13267
P13268
P13269
P13270
P13271
P13272
P13273
P13274
P13275
P13276
P13277
P13278
P13279
P13280
P13281
P13282
P13283
P13284
P13285
P13286
P13287
P13288
P13289
P13290
P13291
P13292
P13293
P13294
P13295
P13296
P13297
P13298
P13299
P13300
P13301
P13302
P13303
P13304
P13305
P13306
P13307
P13308
P13309
P13310
P13311
P13312
P13313
P13314
P13315
P13316
P13317
P13318
P13319
P13320
P13321
P13322
P13323
P13324
P13325
P13326
P13327
P13328
P13329
P13330
P13331
P13332
P13333
P13334
P13335
P13336
P13337
P13338
P13339
P13340
P13341
P13342
P13343
P13344
P13345
P13346
P13347
P13348
P13349
P13350
P13351
P13352
P13353
P13354
P13355
P13356
P13357
P13358
P13359
P13360
P13361
P13362
P13363
P13364
P13365
P13366
P13367
P13368
P13369
P13370
P13371
P13372
P13373
P13374
P13375
P13376
P13377
P13378
P13379
P13380
P13381
P13382
P13383
P13384
P13385
P13386
P13387
P13388
P13389
P13390
P13391
P13392
P13393
P13394
P13395
P13396
P13397
P13398
P13399
P13400
P13401
P13402
P13403
P13404
P13405
P13406
P13407
P13408
P13409
P13410
P13411
P13412
P13413
P13414
P13415
P13416
P13417
P13418
P13419
P13420
P13421
P13422
P13423
P13424
P13425
P13426
P13427
P13428
P13429
P13430
P13431
P13432
P13433
P13434
P13435
P13436
P13437
P13438
P13439
P13440
P13441
P13442
P13443
P13444
P13445
P13446
P13447
P13448
P13449
P13450
P13451
P13452
P13453
P13454
P13455
P13456
P13457
P13458
P13459
P13460
P13461
P13462
P13463
P13464
P13465
P13466
P13467
P13468
P13469
P13470
P13471
P13472
P13473
P13474
P13475
P13476
P13477
P13478
P13479
P13480
P13481
P13482
P13483
P13484
P13485
P13486
P13487
P13488
P13489
P13490
P13491
P13492
P13493
P13494
P13495
P13496
P13497
P13498
P13499
P13500
P13501
P13502
P13503
P13504
P13505
P13506
P13507
P13508
P13509
P13510
P13511
P13512
P13513
P13514
P13515
P13516
P13517
P13518
P13519
P13520
P13521
P13522
P13523
P13524
P13525
P13526
P13527
P13528
P13529
P13530
P13531
P13532
P13533
P13534
P13535
P13536
P13537
P13538
P13539
P13540
P13541
P13542
P13543
P13544
P13545
P13546
P13547
P13548
P13549
P13550
P13551
P13552
P13553
P13554
P13555
P13556
P13557
P13558
P13559
P13560
P13561
P13562
P13563
P13564
P13565
P13566
P13567
P13568
P13569
P13570
P13571
P13572
P13573
P13574
P13575
P13576
P13577
P13578
P13579
P13580
P13581
P13582
P13583
P13584
P13585
P13586
P13587
P13588
P13589
P13590
P13591
P13592
P13593
P13594
P13595
P13596
P13597
P13598
P13599
P13600
P13601
P13602
P13603
P13604
P13605
P13606
P13607
P13608
P13609
P13610
P13611
P13612
P13613
P13614
P13615
P13616
P13617
P13618
P13619
P13620
P13621
P13622
P13623
P13624
P13625
P13626
P13627
P13628
P13629
P13630
P13631
P13632
P13633
P13634
P13635
P13636
P13637
P13638
P13639
P13640
P13641
P13642
P13643
P13644
P13645
P13646
P13647
P13648
P13649
P13650
P13651
P13652
P13653
P13654
P13655
P13656
P13657
P13658
P13659
P13660
P13661
P13662
P13663
P13664
P13665
P13666
P13667
P13668
P13669
P13670
P13671
P13672
P13673
P13674
P13675
P13676
P13677
P13678
P13679
P13680
P13681
P13682
P13683
P13684
P13685
P13686
P13687
P13688
P13689
P13690
P13691
P13692
P13693
P13694
P13695
P13696
P13697
P13698
P13699
P13700
P13701
P13702
P13703
P13704
P13705
P13706
P13707
P13708
P13709
P13710
P13711
P13712
P13713
P13714
P13715
P13716
P13717
P13718
P13719
P13720
P13721
P13722
P13723
P13724
P13725
P13726
P13727
P13728
P13729
P13730
P13731
P13732
P13733
P13734
P13735
P13736
P13737
P13738
P13739
P13740
P13741
P13742
P13743
P13744
P13745
P13746
P13747
P13748
P13749
P13750
P13751
P13752
P13753
P13754
P13755
P13756
P13757
P13758
P13759
P13760
P13761
P13762
P13763
P13764
P13765
P13766
P13767
P13768
P13769
P13770
P13771
P13772
P13773
P13774
P13775
P13776
P13777
P13778
P13779
P13780
P13781
P13782
P13783
P13784
P13785
P13786
P13787
P13788
P13789
P13790
P13791
P13792
P13793
P13794
P13795
P13796
P13797
P13798
P13799
P13800
P13801
P13802
P13803
P13804
P13805
P13806
P13807
P13808
P13809
P13810
P13811
P13812
P13813
P13814
P13815
P13816
P13817
P13818
P13819
P13820
P13821
P13822
P13823
P13824
P13825
P13826
P13827
P13828
P13829
P13830
P13831
P13832
P13833
P13834
P13835
P13836
P13837
P13838
P13839
P13840
P13841
P13842
P13843
P13844
P13845
P13846
P13847
P13848
P13849
P13850
P13851
P13852
P13853
P13854
P13855
P13856
P13857
P13858
P13859
P13860
P13861
P13862
P13863
P13864
P13865
P13866
P13867
P13868
P13869
P13870
P13871
P13872
P13873
P13874
P13875
P13876
P13877
P13878
P13879
P13880
P13881
P13882
P13883
P13884
P13885
P13886
P13887
P13888
P13889
P13890
P13891
P13892
P13893
P13894
P13895
P13896
P13897
P13898
P13899
P13900
P13901
P13902
P13903
P13904
P13905
P13906
P13907
P13908
P13909
P13910
P13911
P13912
P13913
P13914
P13915
P13916
P13917
P13918
P13919
P13920
P13921
P13922
P13923
P13924
P13925
P13926
P13927
P13928
P13929
P13930
P13931
P13932
P13933
P13934
P13935
P13936
P13937
P13938
P13939
P13940
P13941
P13942
P13943
P13944
P13945
P13946
P13947
P13948
P13949
P13950
P13951
P13952
P13953
P13954
P13955
P13956
P13957
P13958
P13959
P13960
P13961
P13962
P13963
P13964
P13965
P13966
P13967
P13968
P13969
P13970
P13971
P13972
P13973
P13974
P13975
P13976
P13977
P13978
P13979
P13980
P13981
P13982
P13983
P13984
P13985
P13986
P13987
P13988
P13989
P13990
P13991
P13992
P13993
P13994
P13995
P13996
P13997
P13998
P13999
P14000
P14001
P14002
P14003
P14004
P14005
P14006
P14007
P14008
P14009
P14010
P14011
P14012
P14013
P14014
P14015
P14016
P14017
P14018
P14019
P14020
P14021
P14022
P14023
P14024
P14025
P14026
P14027
P14028
P14029
P14030
P14031
P14032
P14033
P14034
P14035
P14036
P14037
P14038
P14039
P14040
P14041
P14042
P14043
P14044
P14045
P14046
P14047
P14048
P14049
P14050
P14051
P14052
P14053
P14054
P14055
P14056
P14057
P14058
P14059
P14060
P14061
P14062
P14063
P14064
P14065
P14066
P14067
P14068
P14069
P14070
P14071
P14072
P14073
P14074
P14075
P14076
P14077
P14078
P14079
P14080
P14081
P14082
P14083
P14084
P14085
P14086
P14087
P14088
P14089
P14090
P14091
P14092
P14093
P14094
P14095
P14096
P14097
P14098
P14099
P14100
P14101
P14102
P14103
P14104
P14105
P14106
P14107
P14108
P14109
P14110
P14111
P14112
P14113
P14114
P14115
P14116
P14117
P14118
P14119
P14120
P14121
P14122
P14123
P14124
P14125
P14126
P14127
P14128
P14129
P14130
P14131
P14132
P14133
P14134
P14135
P14136
P14137
P14138
P14139
P14140
P14141
P14142
P14143
P14144
P14145
P14146
P14147
P14148
P14149
P14150
P14151
P14152
P14153
P14154
P14155
P14156
P14157
P14158
P14159
P14160
P14161
P14162
P14163
P14164
P14165
P14166
P14167
P14168
P14169
P14170
P14171
P14172
P14173
P14174
P14175
P14176
P14177
P14178
P14179
P14180
P14181
P14182
P14183
P14184
P14185
P14186
P14187
P14188
P14189
P14190
P14191
P14192
P14193
P14194
P14195
P14196
P14197
P14198
P14199
P14200
P14201
P14202
P14203
P14204
P14205
P14206
P14207
P14208
P14209
P14210
P14211
P14212
P14213
P14214
P14215
P14216
P14217
P14218
P14219
P14220
P14221
P14222
P14223
P14224
P14225
P14226
P14227
P14228
P14229
P14230
P14231
P14232
P14233
P14234
P14235
P14236
P14237
P14238
P14239
P14240
P14241
P14242
P14243
P14244
P14245
P14246
P14247
P14248
P14249
P14250
P14251
P14252
P14253
P14254
P14255
P14256
P14257
P14258
P14259
P14260
P14261
P14262
P14263
P14264
P14265
P14266
P14267
P14268
P14269
P14270
P14271
P14272
P14273
P14274
P14275
P14276
P14277
P14278
P14279
P14280
P14281
P14282
P14283
P14284
P14285
P14286
P14287
P14288
P14289
P14290
P14291
P14292
P14293
P14294
P14295
P14296
P14297
P14298
P14299
P14300
P14301
P14302
P14303
P14304
P14305
P14306
P14307
P14308
P14309
P14310
P14311
P14312
P14313
P14314
P14315
P14316
P14317
P14318
P14319
P14320
P14321
P14322
P14323
P14324
P14325
P14326
P14327
P14328
P14329
P14330
P14331
P14332
P14333
P14334
P14335
P14336
P14337
P14338
P14339
P14340
P14341
P14342
P14343
P14344
P14345
P14346
P14347
P14348
P14349
P14350
P14351
P14352
P14353
P14354
P14355
P14356
P14357
P14358
P14359
P14360
P14361
P14362
P14363
P14364
P14365
P14366
P14367
P14368
P14369
P14370
P14371
P14372
P14373
P14374
P14375
P14376
P14377
P14378
P14379
P14380
P14381
P14382
P14383
P14384
P14385
P14386
P14387
P14388
P14389
P14390
P14391
P14392
P14393
P14394
P14395
P14396
P14397
P14398
P14399
P14400
P14401
P14402
P14403
P14404
P14405
P14406
P14407
P14408
P14409
P14410
P14411
P14412
P14413
P14414
P14415
P14416
P14417
P14418
P14419
P14420
P14421
P14422
P14423
P14424
P14425
P14426
P14427
P14428
P14429
P14430
P14431
P14432
P14433
P14434
P14435
P14436
P14437
P14438
P14439




110 Benner Circle
Belleville, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *chromatographic plus*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	—%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Handwritten notes and signatures:

- Signature: *[Signature]*
- Date: *11/19/24*
- Text: *P13779*
- Text: *P13773*

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

J.P.

Dakota Parson - Operations Technician I

Date Mixed: 10-Oct-2023

Balance Serial # 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397

(P)
P13773
P19779
A7
11/19/24
10-5-20





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32074 **Lot No.:** A0213999

Description: Pesticide Performance Eval Mix w/Surrogate

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: July 31, 2028 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

918780 AJ
11/19/24

918784

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	2.0 µg/mL	+/- 0.1506
2	alpha-BHC	319-84-6	15279000	99%	1.0 µg/mL	+/- 0.0749
3	gamma-BHC (Lindane)	58-89-9	14931300	97%	1.0 µg/mL	+/- 0.0756
4	beta-BHC	319-85-7	BCCC6425	99%	1.0 µg/mL	+/- 0.0749
5	Endrin	72-20-8	15076100	98%	5.0 µg/mL	+/- 0.3769
6	4,4'-DDT	50-29-3	231103JLM	97%	10.0 µg/mL	+/- 0.7512
7	Methoxychlor	72-43-5	14979500	98%	24.8 µg/mL	+/- 1.8543
8	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	2.0 µg/mL	+/- 0.1513

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2µm
Rtx-CLP II (cat.# 111323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

150°C to 300°C
@ 4°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 50:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan

Malina Homan - Operations Technician I

Date Mixed: 17-Jul-2024

Balance Serial # 1128353505

Jennifer Pollino

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Jul-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent:

Acetone

CAS # 67-64-1

Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0210240
Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2028 **Storage:** 10°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.3 µg/mL	+/- 56.0105

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13861
P13862

12/9/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024

Balance Serial # B442140311



Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 26-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13861 } ②
P13862 }


12/9/2024



CERTIFIED WEIGHT REPORT


Part Number: 72072
Lot Number: 112018
Description: n-Tetracosane-d50
Expiration Date: 112028
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 1000
NIST Test ID#: 2684186
Weight(s) shown below were combined and diluted to (mL): 200.0

Solvent(s): Methylene chloride
Lot#: 102669
Received by
SG on 11/11/19
p9044-p9053
5E-05 Balance Uncertainty
0.058 Flask Uncertainty


Formulated By:	Prashant Chauhan	112018	DATE
Reviewed By:	Pedro Rentas	112018	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)
1. n-Tetracosane-d50	2072	PR-17753/09216TC1	1000	98	0.2	0.20411	0.20415	1000.2	4.2	18416-32-3 N/A
Method GC8MSD-3.M: Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 250°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Candice Warren.										

TC 720720



The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
• Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Run 40, "P72072 L112018 [1000µg/mL in MeCl2]"

Run Length: 35.00 min, 20999 points at 10 points/second.

Created: Thu, Nov 22, 2018 at 7:23:18 AM.

Sampled: Sequence "112018-GC4M1", Method "GC4-M1".

Analyzed using Method "GC4-M1".

Comments

GC4-M1 Analysis by Melissa Stonier


Column ID SPB5 L#60062-01A : 30 meter x 0.53mm x 1.5µm Film Thickness

Flow rates: Total Flow = 300 mL/min, Helium (carrier) = 6.5 mL, Helium (make-up) = 25 mL, Hydrogen (detector) = 30 mL, Air (detector) = 360 mL

Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes.

Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDAQ Channel 1.

Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 µL, Range = 3





SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: **Kleinfelder**
ADDRESS: **180 Sheree Blvd. Suite 3800**
CITY **Exton** STATE: **PA** ZIP: **19341**
ATTENTION: **Mark Warchol**
PHONE: **484-883-3892** FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **Logan School**
PROJECT NO.: LOCATION: **Philadelphia, PA**
PROJECT MANAGER: **Mark Warchol**
e-mail: **mwarchol@kleinfelder.com**
PHONE: **484-883-3892** FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS: **Same**
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) **5** DAYS*
HARDCOPY (DATA PACKAGE): **5** DAYS*
EDD: **5** DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☒ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

PADEP Historic
Clean Cell Parameters

1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS		
			COMP	GRAB	DATE	TIME		E										← Specify Preservatives A-HCl B-HNO3 C-H2SO4	D-NaOH E-ICE F-OTHER
1.	COMP-1	Soil	✓		11/1/25	10:05	6	✓											
2.	COMP-2	↓	↓		↓	10:45	↓	↓											
3.	COMP-3	↓	↓		↓	11:30	↓	↓											
4.																			
5.																			
6.																			
7.																			
8.																			
9.																			
10.																			

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. [Signature]	DATE/TIME: 11/1/25 15:00	RECEIVED BY: 1. [Signature]	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.6°C
RELINQUISHED BY SAMPLER: 2. [Signature]	DATE/TIME: 11/2/25 11:45	RECEIVED BY: 2. [Signature]	Comments: 2.6°C + 0.5 = 2.6°C
RELINQUISHED BY SAMPLER: 3. [Signature]	DATE/TIME:	RECEIVED BY: 3. [Signature]	Page 1 of 1 CLIENT: ☐ Hand Delivered ☒ Other FedEx Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3614	POWE02	Order Date : 11/12/2025 11:51:00 AM	Project Mgr :
Client Name : Kleinfelder		Project Name : Logan School	Report Type : Results+QC
Client Contact : Mark Warchol		Receive DateTime : 11/12/2025 11:45:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : Kleinfelder		Purchase Order :	Hard Copy Date :
Invoice Contact : Mark Warchol			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3614-01	COMP-1	Solid	11/11/2025	10:05					
					VOCMS Group1		8260D	5 Bus. Days	
Q3614-02	COMP-2	Solid	11/11/2025	10:45					
					VOCMS Group1		8260D	5 Bus. Days	
Q3614-03	COMP-3	Solid	11/11/2025	11:30					
					VOCMS Group1		8260D	5 Bus. Days	

Relinquished By :

Date / Time :

OP
11/12/25 1250

Received By :

Date / Time :

[Signature]
11/12/25 1250

Storage Area : VOA Refridgerator Room