

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3614</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PD091221.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB170527BS (Column 1)	Aldrin	16.65	16.2	ug/kg	97				82	124		
	Dieldrin	16.65	16.3	ug/kg	98				85	121		
	4,4'-DDE	16.65	16.3	ug/kg	98				81	123		
	4,4'-DDD	16.65	16.0	ug/kg	96				80	131		
	4,4'-DDT	16.65	16.5	ug/kg	99				70	129		
PB170527BS (Column 2)	Aldrin	16.65	17.2	ug/kg	103				82	124		
	Dieldrin	16.65	17.1	ug/kg	103				85	121		
	4,4'-DDE	16.65	17.1	ug/kg	103				81	123		
	4,4'-DDD	16.65	16.6	ug/kg	100				80	131		
	4,4'-DDT	16.65	17.6	ug/kg	106				70	129		