

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3614

Analytical Method: 8270E

Client: Kleinfelder

DataFile: BG064758.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170528BS	Naphthalene	1700	1400	ug/Kg	82				62	100	
	Fluorene	1700	1400	ug/Kg	82				75	99	
	Phenanthrene	1700	1400	ug/Kg	82				75	100	
	Anthracene	1700	1400	ug/Kg	82				75	103	
	Pyrene	1700	1400	ug/Kg	82				59	103	
	Benzo(a)anthracene	1700	1400	ug/Kg	82				75	104	
	Chrysene	1700	1400	ug/Kg	82				75	103	
	Benzo(b)fluoranthene	1700	1600	ug/Kg	94				62	109	
	Benzo(a)pyrene	1700	1600	ug/Kg	94				79	109	
	Indeno(1,2,3-cd)pyrene	1700	1600	ug/Kg	94				63	101	
	Benzo(g,h,i)perylene	1700	1600	ug/Kg	94				70	108	