

Surrogate Summary

SDG No.: **Q3614**

Client: **Kleinfelder**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3614-01	COMP-1	1,2-Dichloroethane-d4	50	59.5	119		63	155
		Dibromofluoromethane	50	51.1	102		70	134
		Toluene-d8	50	48.5	97		74	123
		4-Bromofluorobenzene	50	45.2	90		52	138
Q3614-02	COMP-2	1,2-Dichloroethane-d4	50	61.9	124		63	155
		Dibromofluoromethane	50	52.0	104		70	134
		Toluene-d8	50	49.0	98		74	123
		4-Bromofluorobenzene	50	47.0	94		52	138
Q3614-03	COMP-3	1,2-Dichloroethane-d4	50	59.4	119		63	155
		Dibromofluoromethane	50	51.0	102		70	134
		Toluene-d8	50	48.8	98		74	123
		4-Bromofluorobenzene	50	45.1	90		52	138
VY1112SBL01	VY1112SBL01	1,2-Dichloroethane-d4	50	57.4	115		63	155
		Dibromofluoromethane	50	51.1	102		70	134
		Toluene-d8	50	48.1	96		74	123
		4-Bromofluorobenzene	50	42.4	85		52	138
VY1112SBS01	VY1112SBS01	1,2-Dichloroethane-d4	50	49.4	99		63	155
		Dibromofluoromethane	50	50.6	101		70	134
		Toluene-d8	50	51.3	103		74	123
		4-Bromofluorobenzene	50	51.8	104		52	138
VY1112SBSD01	VY1112SBSD01	1,2-Dichloroethane-d4	50	50.9	102		63	155
		Dibromofluoromethane	50	50.2	100		70	134
		Toluene-d8	50	50.5	101		74	123
		4-Bromofluorobenzene	50	50.1	100		52	138