

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3614</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>VY023754.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1112SBS01	cis-1,2-Dichloroethene	20	19.6	ug/Kg	98			82	123	
	1,1,1-Trichloroethane	20	20.1	ug/Kg	101			80	126	
	Benzene	20	20.1	ug/Kg	101			84	121	
	Trichloroethene	20	20.1	ug/Kg	101			83	122	
	Toluene	20	20.6	ug/Kg	103			83	122	
	Ethyl Benzene	20	20.2	ug/Kg	101			82	124	
	m/p-Xylenes	40	41.0	ug/Kg	103			83	124	
	o-Xylene	20	19.8	ug/Kg	99			83	123	
	Isopropylbenzene	20	19.8	ug/Kg	99			82	124	