

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3614</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>VY023755.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1112SBSD01	cis-1,2-Dichloroethene	20	20.7	ug/Kg	104	6		82	123	15
	1,1,1-Trichloroethane	20	21.1	ug/Kg	106	5		80	126	15
	Benzene	20	20.8	ug/Kg	104	3		84	121	15
	Trichloroethene	20	20.6	ug/Kg	103	2		83	122	15
	Toluene	20	20.7	ug/Kg	104	1		83	122	15
	Ethyl Benzene	20	21.1	ug/Kg	106	5		82	124	15
	m/p-Xylenes	40	42.1	ug/Kg	105	2		83	124	15
	o-Xylene	20	20.9	ug/Kg	104	5		83	123	15
	Isopropylbenzene	20	21.0	ug/Kg	105	6		82	124	15