



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

CASE NARRATIVE

Core Environmental Consultants and Services, Inc.

Project Name: Jones Beach Concrete

Project # N/A

Order ID # Q3615

Test Name: VOC-TCLVOA-10,SVOC-TCL BNA -20,PCB,TPH

GC,Mercury,Metals ICP-TAL

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 11/12/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOC-TCLVOA-10,SVOC-TCL BNA -20,PCB,TPH GC,Mercury,Metals ICP-TAL. This data package contains results for VOC-TCLVOA-10(8260D),SVOC-TCL BNA -20(8270E),PCB(8082A),TPH GC(8015D),Mercury(7471B),Metals ICP-TAL(6010D).

C. Analytical Techniques:

VOC-TCLVOA-10 : The analysis performed on instrument MSVOA_Y were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis of VOC-TCLVOA-10 was based on method 8260D.

SVOC-TCL BNA -20 : The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The samples were analyzed on instrument BNA_G using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3541.

TPH GC : The analysis were performed on instrument FID_F. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 13302.The analysis of TPH GC was based on method 8015D and extraction was done based on method 3541.

PCB : The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11.The analyses were performed on instrument GCECD_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

11. The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

Mercury, Metals ICP-TAL : The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils). The analysis and digestion of Mercury was based on method 7471B

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds except following
PCB : The MS {Q3609-07MS} with File ID: PO115077.D recoveries met the requirements for all compounds except for [AR1016(1)440% - AR1016(2)218%] and [AR1260(2)164%] due to matrix interference.

Mercury, Metals ICP-TAL : The Matrix Spike (AU-713-COMP-01MS) analysis met criteria for all compounds except for Antimony, Potassium, Selenium, Sodium, Vanadium and Zinc due to Chemical Interference during Digestion Process. The Matrix Spike (COMP-3MS) analysis met criteria for all compounds except for Mercury due to Sample matrix interference.

The MSD recoveries met the requirements for all compounds except following
PCB : The MSD {Q3609-07MSD} with File ID: PO115078.D recoveries met the requirements for all compounds except for [AR1016(1)456% - AR1016(2)213%] and [AR1260(2)170%] due to matrix interference.

Mercury, Metals ICP-TAL : The Matrix Spike Duplicate (AU-713-COMP-01MSD) analysis met criteria for all compounds except for Antimony, Barium, Potassium, Selenium, Sodium, Vanadium and Zinc due to Chemical Interference during Digestion Process.

The RPD were met for all analysis except following
VOC-TCLVOA-10 : The RPD for {VY1113SBSD01} with File ID: VY023777.D met criteria except for 1,2,3-Trichlorobenzene[20%] due to difference in BS and BSD concentrations.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements except following



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

SVOC-TCL BNA -20 : The %RSD is greater than 20% in the Initial Calibration (Method 8270- BG111225.M) for 2,4-Dinitrophenol, Pentachlorophenol, These Compounds are passing on Linear regression.

The %RSD is greater than 20% in the Initial Calibration (Method 8270-BP102925.M) for 2-Nitrophenol, 2-Nitroaniline, 2,6-Dinitrotoluene, 3-Nitroaniline, 2,4-Dinitrotoluene, Butylbenzylphthalate, Bis(2-ethylhexyl)phtha... These Compounds are passing on Linear regression and 2,4-Dinitrophenol, 4,6-Dinitro-2-methylphenol.. are passing on Quadratic regression.

The Continuous Calibration met the requirements except following
VOC-TCLVOA-10 : The Continuous Calibration File ID VY023774.D met the requirements except for 4-Bromofluorobenzene which is not our target compound, therefore no corrective action was taken.

SVOC-TCL BNA -20 : The Continuous Calibration File ID BP026126.D met the requirements except for 2,3,4,6-Tetrachlorophenol, 2,4-Dinitrophenol, 2-Nitrophenol, 4,6-Dinitro-2-methylphenol, Di-n-octyl phthalate and 2,4,6-Tribromophenol, failing high but the associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Tuning criteria met requirements.

The Duplicate analysis met criteria for all compoundsexcept following
Mercury, Metals ICP-TAL : The Duplicate (AU-713-COMP-01MSD) analysis met criteria for all compounds except for Iron, Manganese, Sodium and Zinc due to Chemical Interference during Digestion Process.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

SVOC-TCL BNA -20: The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Mercury, Metals ICP-TAL : In analytical Sequence LB137889, The % Recovery outside limit for Potassium and Sodium for CCV06 and CCV07 but, no any sample associated under this CCVs.

In analytical Sequence LB137889, The Result outside limit for Potassium and Sodium for CCB06 and CCB07 but, no any sample associated under this CCBs.

VOC-TCLVOA-10 : Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

Trip Blank was not provided with this set of samples.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____