

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137881

Review By	mohan	Review On	11/14/2025 3:10:06 PM
Supervise By	jaswal	Supervise On	11/14/2025 4:14:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP87933,MP87934,MP87935,MP87936,MP87937,MP87938
ICV Standard	MP87939
CCV Standard	MP87941
ICSA Standard	
CRI Standard	MP87943
LCS Standard	
Chk Standard	MP87940,MP87942,MP87944,MP87946

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/13/25 11:22		Ayul	OK
2	S0.2	S0.2	CAL2	11/13/25 11:24		Ayul	OK
3	S2.5	S2.5	CAL3	11/13/25 11:27		Ayul	OK
4	S5	S5	CAL4	11/13/25 11:29		Ayul	OK
5	S7.5	S7.5	CAL5	11/13/25 11:31		Ayul	OK
6	S10	S10	CAL6	11/13/25 11:39		Ayul	OK
7	ICV113	ICV113	ICV	11/13/25 11:43		Ayul	OK
8	ICB113	ICB113	ICB	11/13/25 11:46		Ayul	OK
9	CCV72	CCV72	CCV	11/13/25 11:48		Ayul	OK
10	CCB72	CCB72	CCB	11/13/25 11:50		Ayul	OK
11	CRA	CRA	CRDL	11/13/25 11:53		Ayul	OK
12	HighStd	HighStd	HIGH STD	11/13/25 11:55		Ayul	OK
13	ChkStd	ChkStd	SAM	11/13/25 11:57		Ayul	OK
14	PB170524BL	PB170524BL	MB	11/13/25 11:59		Ayul	OK
15	PB170524BS	PB170524BS	LCS	11/13/25 12:02		Ayul	OK
16	Q3614-01	COMP-1	SAM	11/13/25 12:04		Ayul	OK
17	Q3614-02	COMP-2	SAM	11/13/25 12:06		Ayul	OK
18	Q3614-03	COMP-3	SAM	11/13/25 12:14		Ayul	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137881

Review By	mohan	Review On	11/14/2025 3:10:06 PM
Supervise By	jaswal	Supervise On	11/14/2025 4:14:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP87933,MP87934,MP87935,MP87936,MP87937,MP87938
ICV Standard	MP87939
CCV Standard	MP87941
ICSA Standard	
CRI Standard	MP87943
LCS Standard	
Chk Standard	MP87940,MP87942,MP87944,MP87946

19	Q3614-03DUP	COMP-3DUP	DUP	11/13/25 12:16		Ayul	OK
20	Q3614-03MS	COMP-3MS	MS	11/13/25 12:18		Ayul	OK
21	CCV73	CCV73	CCV	11/13/25 12:20		Ayul	OK
22	CCB73	CCB73	CCB	11/13/25 12:23		Ayul	OK
23	Q3614-03MSD	COMP-3MSD	MSD	11/13/25 12:25		Ayul	OK
24	Q3615-01	East Bathhouse- Conc	SAM	11/13/25 12:27		Ayul	OK
25	Q3618-01	EME-TOP-SOIL	SAM	11/13/25 12:30		Ayul	OK
26	Q3618-02	EME-GENERAL-FILL	SAM	11/13/25 12:32		Ayul	OK
27	Q3621-02	VNJ-231	SAM	11/13/25 12:34		Ayul	OK
28	Q3624-01	OK-02-11122025	SAM	11/13/25 12:36		Ayul	OK
29	Q3614-03L	COMP-3L	SD	11/13/25 12:39		Ayul	OK
30	Q3614-03A	COMP-3A	PS	11/13/25 12:41		Ayul	OK
31	CCV74	CCV74	CCV	11/13/25 12:43		Ayul	OK
32	CCB74	CCB74	CCB	11/13/25 12:45		Ayul	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137889

Review By	jaswal	Review On	11/14/2025 1:56:13 PM
Supervise By	mohan	Supervise On	11/14/2025 4:13:03 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/13/25 12:46		Jaswal	OK
2	S1	S1	CAL2	11/13/25 12:51		Jaswal	OK
3	S2	S2	CAL3	11/13/25 12:55		Jaswal	OK
4	S3	S3	CAL4	11/13/25 12:59		Jaswal	OK
5	S4	S4	CAL5	11/13/25 13:03		Jaswal	OK
6	S5	S5	CAL6	11/13/25 13:07		Jaswal	OK
7	ICV01	ICV01	ICV	11/13/25 13:12		Jaswal	OK
8	LLICV01	LLICV01	LLICV	11/13/25 13:16		Jaswal	OK
9	ICB01	ICB01	ICB	11/13/25 13:21		Jaswal	OK
10	CRI01	CRI01	CRDL	11/13/25 13:25		Jaswal	OK
11	ICSA01	ICSA01	ICSA	11/13/25 13:29		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	11/13/25 13:33		Jaswal	OK
13	ICSADL	ICSADL	ICSA	11/13/25 13:37		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	11/13/25 13:42		Jaswal	OK
15	CCV01	CCV01	CCV	11/13/25 13:46		Jaswal	OK
16	CCB01	CCB01	CCB	11/13/25 13:50		Jaswal	OK
17	PB170522BL	PB170522BL	MB	11/13/25 13:54		Jaswal	OK
18	PB170522BS	PB170522BS	LCS	11/13/25 13:58		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137889

Review By	jaswal	Review On	11/14/2025 1:56:13 PM
Supervise By	mohan	Supervise On	11/14/2025 4:13:03 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	Q3615-01	East Bathhouse- Conc	SAM	11/13/25 14:03		Jaswal	OK
20	Q3609-01	AU-713-COMP-01	SAM	11/13/25 14:07		Jaswal	OK
21	Q3609-01DUP	AU-713-COMP-01DUP	DUP	11/13/25 14:11		Jaswal	OK
22	Q3609-01L	AU-713-COMP-01L	SD	11/13/25 14:15		Jaswal	OK
23	Q3609-01MS	AU-713-COMP-01MS	MS	11/13/25 14:19		Jaswal	OK
24	Q3609-01MSD	AU-713-COMP-01MSD	MSD	11/13/25 14:23		Jaswal	OK
25	Q3609-01A	AU-713-COMP-01A	PS	11/13/25 14:27	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
26	Q3609-04	AU-713-COMP-02	SAM	11/13/25 14:31		Jaswal	OK
27	CCV02	CCV02	CCV	11/13/25 14:35		Jaswal	OK
28	CCB02	CCB02	CCB	11/13/25 14:39		Jaswal	OK
29	LR-1	LR-1	HIGH STD	11/13/25 14:44		Jaswal	OK
30	LR-2	LR-2	HIGH STD	11/13/25 15:08		Jaswal	OK
31	LR-3	LR-3	HIGH STD	11/13/25 15:15		Jaswal	OK
32	Q3609-07	AU-713-COMP-03	SAM	11/13/25 15:19		Jaswal	OK
33	Q3609-10	AU-713-COMP-04	SAM	11/13/25 15:23		Jaswal	OK
34	Q3609-13	AU-713-COMP-05	SAM	11/13/25 15:27		Jaswal	OK
35	Q3614-01	COMP-1	SAM	11/13/25 15:31		Jaswal	OK
36	Q3614-02	COMP-2	SAM	11/13/25 15:35		Jaswal	OK
37	Q3614-03	COMP-3	SAM	11/13/25 15:39		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137889

Review By	jaswal	Review On	11/14/2025 1:56:13 PM
Supervise By	mohan	Supervise On	11/14/2025 4:13:03 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

38	Q3618-01	EME-TOP-SOIL	SAM	11/13/25 15:43		Jaswal	OK
39	CCV03	CCV03	CCV	11/13/25 15:48		Jaswal	OK
40	CCB03	CCB03	CCB	11/13/25 15:51		Jaswal	OK
41	Q3618-02	EME-GENERAL-FILL	SAM	11/13/25 15:56		Jaswal	OK
42	Q3619-03	GP-B1(8-12)	SAM	11/13/25 16:00		Jaswal	OK
43	Q3619-04	GP-B1(12-16)	SAM	11/13/25 16:04		Jaswal	OK
44	Q3619-09	GP-B2(12-16)	SAM	11/13/25 16:08		Jaswal	OK
45	Q3619-14	GP-B3(13.2)	SAM	11/13/25 16:12		Jaswal	OK
46	Q3619-17	GP-B4(8-12)	SAM	11/13/25 16:16		Jaswal	OK
47	Q3619-18	GP-B4(13.2)	SAM	11/13/25 16:20		Jaswal	OK
48	Q3621-02	VNJ-231	SAM	11/13/25 16:24		Jaswal	OK
49	Q3624-01	OK-02-11122025	SAM	11/13/25 16:28		Jaswal	OK
50	PB170535BL	PB170535BL	MB	11/13/25 16:32		Jaswal	OK
51	CCV04	CCV04	CCV	11/13/25 16:36		Jaswal	OK
52	CCB04	CCB04	CCB	11/13/25 16:40		Jaswal	OK
53	PB170535BS	PB170535BS	LCS	11/13/25 16:45		Jaswal	OK
54	PB170517TB	PB170517TB	MB	11/13/25 16:49		Jaswal	OK
55	Q3609-03	AU-713-COMP-01	SAM	11/13/25 16:53		Jaswal	OK
56	Q3609-06	AU-713-COMP-02	SAM	11/13/25 16:57		Jaswal	OK
57	Q3609-09	AU-713-COMP-03	SAM	11/13/25 17:02		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137889

Review By	jaswal	Review On	11/14/2025 1:56:13 PM
Supervise By	mohan	Supervise On	11/14/2025 4:13:03 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

58	Q3609-12	AU-713-COMP-04	SAM	11/13/25 17:06		Jaswal	OK
59	Q3609-15	AU-713-COMP-05	SAM	11/13/25 17:10		Jaswal	OK
60	Q3622-01	RT3071-SOLID	SAM	11/13/25 17:14		Jaswal	OK
61	Q3623-01	RBR200007-SOIL	SAM	11/13/25 17:18		Jaswal	OK
62	Q3623-01DUP	RBR200007-SOILDUP	DUP	11/13/25 17:23		Jaswal	OK
63	CCV05	CCV05	CCV	11/13/25 17:27		Jaswal	OK
64	CCB05	CCB05	CCB	11/13/25 17:31		Jaswal	OK
65	Q3623-01L	RBR200007-SOILL	SD	11/13/25 17:35		Jaswal	OK
66	Q3623-01MS	RBR200007-SOILMS	MS	11/13/25 17:39		Jaswal	OK
67	Q3623-01MSD	RBR200007-SOILMSI	MSD	11/13/25 17:43		Jaswal	OK
68	Q3623-01A	RBR200007-SOILA	PS	11/13/25 17:47	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
69	PB170531BL	PB170531BL	MB	11/13/25 17:51		Jaswal	OK
70	PB170531BS	PB170531BS	LCS	11/13/25 17:56		Jaswal	OK
71	Q3616-05	Composite	SAM	11/13/25 18:00		Jaswal	OK
72	Q3616-05DUP	Composite DUP	DUP	11/13/25 18:04		Jaswal	OK
73	Q3616-05L	Composite L	SD	11/13/25 18:08		Jaswal	OK
74	Q3616-05MS	Composite MS	MS	11/13/25 18:13		Jaswal	OK
75	CCV06	CCV06	CCV	11/13/25 18:17	CCV fail for K,Na	Jaswal	OK
76	CCB06	CCB06	CCB	11/13/25 18:25	CCV fail for K,Na	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137889

Review By	jaswal	Review On	11/14/2025 1:56:13 PM
Supervise By	mohan	Supervise On	11/14/2025 4:13:03 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

77	Q3616-05A	Composite A	PS	11/13/25 18:29	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
78	Q3616-05MSD	Composite MSD	MSD	11/13/25 18:33		Jaswal	OK
79	CCV07	CCV07	CCV	11/13/25 18:38	CCV fail for K,Na	Jaswal	OK
80	CCB07	CCB07	CCB	11/13/25 18:42	CCV fail for K,Na	Jaswal	OK