

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3615

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 21:19

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	6.33	6.33	6.23	6.43	0.00
Aroclor-1254-2 (2)	6.54	6.55	6.45	6.65	0.01
Aroclor-1254-3 (3)	6.90	6.91	6.81	7.01	0.01
Aroclor-1254-4 (4)	7.19	7.19	7.09	7.29	0.01
Aroclor-1254-5 (5)	7.60	7.60	7.50	7.70	0.00
Tetrachloro-m-xylene	4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl	9.93	9.93	9.83	10.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3615

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 21:19

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.52	5.53	5.43	5.63	0.01
Aroclor-1254-2 (2)	5.67	5.67	5.57	5.77	0.00
Aroclor-1254-3 (3)	6.07	6.08	5.98	6.18	0.01
Aroclor-1254-4 (4)	6.30	6.30	6.20	6.40	0.00
Aroclor-1254-5 (5)	6.72	6.72	6.62	6.82	0.00
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.64	8.64	8.54	8.74	0.00