

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q3615
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115082.D **Time Analyzed:** 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.485	5.387	5.587	497.520	500.000	-0.5
Aroclor-1016-2	5.507	5.408	5.608	485.700	500.000	-2.9
Aroclor-1016-3	5.569	5.470	5.670	506.150	500.000	1.2
Aroclor-1016-4	5.665	5.566	5.766	492.600	500.000	-1.5
Aroclor-1016-5	5.956	5.858	6.058	478.720	500.000	-4.3
Aroclor-1260-1	7.070	6.971	7.171	469.500	500.000	-6.1
Aroclor-1260-2	7.324	7.225	7.425	504.990	500.000	1.0
Aroclor-1260-3	7.681	7.582	7.782	499.300	500.000	-0.1
Aroclor-1260-4	7.903	7.803	8.003	486.490	500.000	-2.7
Aroclor-1260-5	8.210	8.110	8.310	482.670	500.000	-3.5
Decachlorobiphenyl	9.930	9.830	10.030	48.900	50.000	-2.2
Tetrachloro-m-xylene	4.346	4.248	4.448	50.980	50.000	2.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q3615
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115082.D **Time Analyzed:** 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	485.160	500.000	-3.0
Aroclor-1016-2	4.745	4.647	4.847	486.630	500.000	-2.7
Aroclor-1016-3	4.920	4.821	5.021	496.470	500.000	-0.7
Aroclor-1016-4	4.961	4.864	5.064	493.430	500.000	-1.3
Aroclor-1016-5	5.173	5.076	5.276	495.990	500.000	-0.8
Aroclor-1260-1	6.203	6.106	6.306	460.310	500.000	-7.9
Aroclor-1260-2	6.391	6.294	6.494	452.250	500.000	-9.6
Aroclor-1260-3	6.544	6.446	6.646	457.870	500.000	-8.4
Aroclor-1260-4	7.015	6.917	7.117	466.890	500.000	-6.6
Aroclor-1260-5	7.256	7.158	7.358	466.250	500.000	-6.8
Decachlorobiphenyl	8.639	8.541	8.741	50.550	50.000	1.1
Tetrachloro-m-xylene	3.648	3.548	3.748	49.510	50.000	-1.0