

Surrogate Summary

SDG No.: Q3615

Client: Core Environmental Consultants and Ser

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114870.D	PIBLK-PO114870.D	Tetrachloro-m-xyl	1	20	15.5	77		60	140
		Decachlorobiphen	1	20	16.8	84		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	17.3	86		60	140
I.BLK-PO115071.D	PIBLK-PO115071.D	Tetrachloro-m-xyl	1	20	17.3	87		60	140
		Decachlorobiphen	1	20	17.4	87		60	140
		Tetrachloro-m-xyl	2	20	17.3	86		60	140
		Decachlorobiphen	2	20	18.5	93		60	140
PB170526BL	PB170526BL	Tetrachloro-m-xyl	1	20	21.9	109		21	165
		Decachlorobiphen	1	20	22.6	113		10	170
		Tetrachloro-m-xyl	2	20	20.9	105		21	165
		Decachlorobiphen	2	20	23.6	118		10	170
PB170526BS	PB170526BS	Tetrachloro-m-xyl	1	20	23.6	118		21	165
		Decachlorobiphen	1	20	23.7	118		10	170
		Tetrachloro-m-xyl	2	20	21.8	109		21	165
		Decachlorobiphen	2	20	25.0	125		10	170
Q3609-07MS	AU-713-COMP-03MS	Tetrachloro-m-xyl	1	20	21.8	109		21	165
		Decachlorobiphen	1	20	18.5	92		10	170
		Tetrachloro-m-xyl	2	20	21.1	106		21	165
		Decachlorobiphen	2	20	19.5	97		10	170
Q3609-07MSD	AU-713-COMP-03MSD	Tetrachloro-m-xyl	1	20	21.9	109		21	165
		Decachlorobiphen	1	20	18.4	92		10	170
		Tetrachloro-m-xyl	2	20	21.1	105		21	165
		Decachlorobiphen	2	20	19.4	97		10	170
I.BLK-PO115086.D	PIBLK-PO115086.D	Tetrachloro-m-xyl	1	20	16.8	84		60	140
		Decachlorobiphen	1	20	17.0	85		60	140
		Tetrachloro-m-xyl	2	20	14.6	73		60	140
		Decachlorobiphen	2	20	17.8	89		60	140
I.BLK-PP075887.D	PIBLK-PP075887.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	20.4	102		60	140
		Tetrachloro-m-xyl	2	20	18.8	94		60	140
		Decachlorobiphen	2	20	19.3	96		60	140
I.BLK-PP076487.D	PIBLK-PP076487.D	Tetrachloro-m-xyl	1	20	22.6	113		60	140
		Decachlorobiphen	1	20	22.8	114		60	140
		Tetrachloro-m-xyl	2	20	20.7	103		60	140
		Decachlorobiphen	2	20	18.7	94		60	140
Q3615-01	East Bathhouse- Concrete	Tetrachloro-m-xyl	1	20	29.3	146		21	165
		Decachlorobiphen	1	20	28.4	142		10	170
		Tetrachloro-m-xyl	2	20	26.6	133		21	165
		Decachlorobiphen	2	20	23.6	118		10	170
I.BLK-PP076502.D	PIBLK-PP076502.D	Tetrachloro-m-xyl	1	20	22.1	111		60	140

Surrogate Summary

SDG No.: Q3615

Client: Core Environmental Consultants and Ser

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP076502.D	PIBLK-PP076502.D	Decachlorobiphen	1	20	21.6	108		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	18.2	91		60	140