

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: CORE02
ProjectID: Jones Beach Concrete
Lab Code: ACE SDG No.: Q3615
DataFile: FF016667.D Analyst Name: YP\AJ Analyst Date: 11-12-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	86283341	101510	106517	4.701