

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG No.: Q3615

Instrument ID: BNA_G

Calibration Date(s): 11/12/2025 11/12/2025

Calibration Time(s): 11:17 16:04

LAB FILE ID:		RRF2.5 = BG064745.D			RRF005 = BG064746.D		RRF010 = BG064747.D	
		RRF020 = BG064748.D			RRF040 = BG064749.D		RRF050 = BG064750.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.084	1.128	1.165	1.170	1.181	1.145	2.9
Benzaldehyde			1.027	0.980	1.029	0.949	0.997	3.0
Phenol-d6		1.388	1.466	1.539	1.580	1.616	1.523	5.0
Phenol		1.501	1.584	1.666	1.711	1.774	1.657	5.4
bis(2-Chloroethyl)ether		1.093	1.141	1.224	1.234	1.224	1.179	4.4
2-Chlorophenol		1.208	1.221	1.268	1.303	1.318	1.258	3.4
2-Methylphenol		1.072	1.149	1.196	1.218	1.234	1.173	4.6
2,2-oxybis(1-Chloropropane)		2.713	2.796	2.959	2.834	2.861	2.792	3.6
Acetophenone		0.465	0.536	0.581	0.561	0.555	0.545	7.0
3+4-Methylphenols			1.480	1.628	1.649	1.670	1.607	4.1
n-Nitroso-di-n-propylamine	0.924	0.875	1.015	1.116	1.097	1.138	1.030	8.9
Nitrobenzene-d5		0.371	0.378	0.410	0.414	0.416	0.405	5.5
Hexachloroethane		0.584	0.549	0.572	0.563	0.550	0.558	2.8
Nitrobenzene		0.353	0.398	0.431	0.415	0.417	0.409	6.6
Isophorone		0.723	0.760	0.796	0.767	0.785	0.770	3.1
2-Nitrophenol		0.133	0.175	0.189	0.193	0.194	0.184	13.1
2,4-Dimethylphenol		0.309	0.301	0.336	0.331	0.334	0.327	4.6
bis(2-Chloroethoxy)methane		0.383	0.419	0.444	0.425	0.432	0.423	4.6
2,4-Dichlorophenol		0.283	0.331	0.367	0.364	0.371	0.351	9.4
Naphthalene		1.076	1.163	1.170	1.141	1.123	1.134	2.8
4-Chloroaniline		0.357	0.452	0.477	0.471	0.472	0.454	9.6
Hexachlorobutadiene		0.365	0.353	0.370	0.366	0.358	0.363	2.1
Caprolactam			0.090	0.104	0.103	0.114	0.105	8.0
4-Chloro-3-methylphenol		0.348	0.388	0.399	0.390	0.402	0.390	5.1
2-Methylnaphthalene		0.820	0.842	0.902	0.844	0.850	0.848	3.0
Hexachlorocyclopentadiene			0.478	0.547	0.560	0.557	0.558	8.3
2,4,6-Trichlorophenol		0.386	0.432	0.472	0.465	0.474	0.457	7.9
2-Fluorobiphenyl		1.403	1.445	1.509	1.420	1.400	1.433	2.6
2,4,5-Trichlorophenol		0.467	0.460	0.546	0.521	0.520	0.512	6.8
1,1-Biphenyl		1.416	1.417	1.498	1.428	1.405	1.436	2.2
2-Chloronaphthalene		1.124	1.112	1.167	1.120	1.116	1.132	1.7
2-Nitroaniline		0.291	0.345	0.373	0.371	0.383	0.364	9.9
Dimethylphthalate		1.535	1.569	1.654	1.549	1.580	1.580	2.4
Acenaphthylene		1.839	1.928	1.996	1.930	1.932	1.933	2.5

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG No.: Q3615

Instrument ID: BNA_G

Calibration Date(s): 11/12/2025 11/12/2025

Calibration Time(s): 11:17 16:04

LAB FILE ID:		RRF2.5 = BG064745.D			RRF005 = BG064746.D		RRF010 = BG064747.D	
		RRF020 = BG064748.D			RRF040 = BG064749.D		RRF050 = BG064750.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.265	0.281	0.313	0.305	0.317	0.304	7.4
3-Nitroaniline		0.247	0.277	0.301	0.306	0.320	0.300	9.6
Acenaphthene		1.072	1.104	1.169	1.166	1.177	1.159	4.6
2,4-Dinitrophenol			0.060	0.114	0.169	0.189	0.158	38.0
4-Nitrophenol			0.181	0.224	0.241	0.257	0.238	13.2
Dibenzofuran		1.897	1.991	1.994	1.867	1.892	1.916	2.7
2,4-Dinitrotoluene		0.376	0.426	0.477	0.462	0.472	0.456	9.1
Diethylphthalate		1.443	1.521	1.585	1.492	1.532	1.513	2.8
4-Chlorophenyl-phenylether		0.892	0.927	0.940	0.880	0.882	0.902	2.5
Fluorene		1.481	1.569	1.595	1.471	1.462	1.503	3.6
4-Nitroaniline		0.245	0.288	0.318	0.320	0.335	0.311	10.7
4,6-Dinitro-2-methylphenol			0.073	0.099	0.117	0.121	0.111	19.4
n-Nitrosodiphenylamine		0.483	0.531	0.544	0.538	0.519	0.524	3.8
2,4,6-Tribromophenol		0.346	0.384	0.410	0.386	0.398	0.390	5.5
4-Bromophenyl-phenylether		0.241	0.263	0.269	0.269	0.259	0.263	3.9
Hexachlorobenzene		0.312	0.312	0.326	0.320	0.318	0.318	1.7
Atrazine		0.245	0.251	0.257	0.253	0.254	0.253	1.5
Pentachlorophenol			0.088	0.138	0.160	0.177	0.157	24.9
Phenanthrene		1.075	1.118	1.123	1.099	1.081	1.094	1.8
Anthracene		1.098	1.122	1.152	1.113	1.099	1.116	1.6
Carbazole		0.926	0.987	0.982	0.971	0.944	0.959	2.3
Di-n-butylphthalate		0.998	1.107	1.105	1.077	1.076	1.070	3.4
Fluoranthene		1.508	1.558	1.531	1.494	1.461	1.496	2.6
Pyrene		1.076	1.089	1.145	1.111	1.164	1.121	2.8
Terphenyl-d14		1.018	1.028	1.060	1.012	1.040	1.017	3.0
Butylbenzylphthalate		0.377	0.391	0.402	0.391	0.404	0.394	2.4
3,3-Dichlorobenzidine			0.501	0.518	0.517	0.507	0.501	3.3
Benzo (a) anthracene		1.338	1.371	1.395	1.379	1.357	1.360	1.7
Chrysene		1.266	1.292	1.301	1.309	1.267	1.282	1.4
Bis (2-ethylhexyl) phthalate		0.582	0.581	0.598	0.583	0.571	0.576	2.5
Di-n-octyl phthalate			0.993	1.026	1.038	0.983	0.996	3.0
Benzo (b) fluoranthene		1.190	1.250	1.299	1.291	1.279	1.273	3.2
Benzo (k) fluoranthene		1.245	1.296	1.321	1.249	1.279	1.279	2.3
Benzo (a) pyrene		1.131	1.176	1.197	1.191	1.182	1.181	2.0
Indeno (1,2,3-cd) pyrene		1.262	1.324	1.388	1.399	1.404	1.374	4.4

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG No.: Q3615

Instrument ID: BNA_G

Calibration Date(s): 11/12/2025 11/12/2025

Calibration Time(s): 11:17 16:04

LAB FILE ID:		RRF2.5 = BG064745.D		RRF005 = BG064746.D		RRF010 = BG064747.D		
		RRF020 = BG064748.D		RRF040 = BG064749.D		RRF050 = BG064750.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.026	1.092	1.128	1.151	1.155	1.126	4.6
Benzo (g,h,i) perylene		1.000	1.044	1.091	1.105	1.108	1.083	4.2
1,2,4,5-Tetrachlorobenzene		0.756	0.746	0.811	0.782	0.759	0.778	3.2
1,4-Dioxane		0.441	0.455	0.445	0.468	0.452	0.443	3.9
2,3,4,6-Tetrachlorophenol		0.418	0.480	0.521	0.501	0.536	0.506	8.9