

Surrogate Summary

SDG No.: Q3615

Client: Core Environmental Consultants and Services, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3615-01	East Bathhouse- Concrete	1,2-Dichloroethane-d4	50	57.3	115		63	155
		Dibromofluoromethane	50	39.4	79		70	134
		Toluene-d8	50	48.6	97		74	123
		4-Bromofluorobenzene	50	42.8	86		52	138
VY1113SBL01	VY1113SBL01	1,2-Dichloroethane-d4	50	56.7	113		63	155
		Dibromofluoromethane	50	50.5	101		70	134
		Toluene-d8	50	48.7	97		74	123
		4-Bromofluorobenzene	50	50.5	101		52	138
VY1113SBS01	VY1113SBS01	1,2-Dichloroethane-d4	50	50.7	101		63	155
		Dibromofluoromethane	50	50.1	100		70	134
		Toluene-d8	50	51.1	102		74	123
		4-Bromofluorobenzene	50	62.2	124		52	138
VY1113SBSD01	VY1113SBSD01	1,2-Dichloroethane-d4	50	51.8	104		63	155
		Dibromofluoromethane	50	51.6	103		70	134
		Toluene-d8	50	52.9	106		74	123
		4-Bromofluorobenzene	50	53.2	106		52	138