

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CORE02
 Lab Code: ACE SDG NO.: Q3615
 Lab File ID: VY023774.D Date Analyzed: 11/13/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:48
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	535734	7.71	769382	8.62	787003	11.41
UPPER LIMIT	1071470	8.207	1538760	9.116	1574010	11.914
LOWER LIMIT	267867	7.207	384691	8.116	393502	10.914
EPA SAMPLE NO.						
East Bathhouse- Concrete	759197	7.71	1341213	8.62	1126773	11.41
VY1113SBL01	770600	7.71	1347630	8.62	1134546	11.41
VY1113SBS01	545029	7.71	820567	8.62	760023	11.41
VY1113SBSD01	571249	7.71	854908	8.62	744787	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.