

Initial and Continuing Calibration Verification

Client: Dal-Tile

SDG No.: Q3616

Project: Waste Water - Dickson Plant

RunNo.: LB137875

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	11/12/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	11/12/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	11/12/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile

SDG No.: Q3616

Project: Waste Water - Dickson Plant

RunNo.: LB137905

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.094	0.099	95	85-115	11/14/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	11/14/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	11/14/2025
Sample ID: Cyanide	CCV3	mg/L	0.24	0.25	96	90-110	11/14/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile

SDG No.: Q3616

Project: Waste Water - Dickson Plant

RunNo.: LB137922

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	11/17/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.95	1	95	90-110	11/17/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	11/17/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.97	1	97	90-110	11/17/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile

SDG No.: Q3616

Project: Waste Water - Dickson Plant

RunNo.: LB138013

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.494	0.50	99	90-110	11/21/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.526	0.50	105	90-110	11/21/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.517	0.50	103	90-110	11/21/2025