

Report of Analysis

Client: Dal-Tile
Project: Waste Water - Dickson Plant
Client Sample ID: Composite
Lab Sample ID: Q3616-05
Level (low/med): low

Date Collected: 11/11/25
Date Received: 11/12/25
SDG No.: Q3616
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	47.8		1	2.56	10.0	ug/L	11/13/25 10:50	11/13/25 18:00	6010D	M3010A
7440-43-9	Cadmium	0.27	J	1	0.25	3.00	ug/L	11/13/25 10:50	11/13/25 18:00	6010D	M3010A
7440-47-3	Chromium	91.4		1	1.06	5.00	ug/L	11/13/25 10:50	11/13/25 18:00	6010D	M3010A
7440-50-8	Copper	8.98	J	1	2.30	10.0	ug/L	11/13/25 10:50	11/13/25 18:00	6010D	M3010A
7439-92-1	Lead	9.88		1	1.15	6.00	ug/L	11/13/25 10:50	11/13/25 18:00	6010D	M3010A
7439-97-6	Mercury	4.26	N	1	0.076	0.20	ug/L	11/14/25 12:40	11/17/25 13:25	7470A	M7470A
7440-02-0	Nickel	12.1	J	1	1.53	20.0	ug/L	11/13/25 10:50	11/13/25 18:00	6010D	M3010A
7782-49-2	Selenium	563		1	4.82	10.0	ug/L	11/13/25 10:50	11/13/25 18:00	6010D	M3010A
7440-66-6	Zinc	414		1	8.33	20.0	ug/L	11/13/25 10:50	11/13/25 18:00	6010D	M3010A

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits