

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Dal-Tile

SDG No.: Q3616

Contract: DALT01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV117	Mercury	4.19	4.0	105	90 - 110	CV	11/17/2025	12:54	LB137914

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Dal-Tile

SDG No.: Q3616

Contract: DALT01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Mercury	5.03	5.0	101	90 - 110	CV	11/17/2025	13:06	LB137914
CCV02	Mercury	5.02	5.0	100	90 - 110	CV	11/17/2025	13:39	LB137914
CCV03	Mercury	5.11	5.0	102	90 - 110	CV	11/17/2025	13:52	LB137914

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Dal-Tile

SDG No.: Q3616

Contract: DALT01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	4010	4000	100	90 - 110	P	11/13/2025	13:12	LB137889
	Cadmium	1990	2000	99	90 - 110	P	11/13/2025	13:12	LB137889
	Chromium	793	800	99	90 - 110	P	11/13/2025	13:12	LB137889
	Copper	999	1000	100	90 - 110	P	11/13/2025	13:12	LB137889
	Lead	3920	4000	98	90 - 110	P	11/13/2025	13:12	LB137889
	Nickel	1950	2000	97	90 - 110	P	11/13/2025	13:12	LB137889
	Selenium	4000	4000	100	90 - 110	P	11/13/2025	13:12	LB137889
	Zinc	1940	2000	97	90 - 110	P	11/13/2025	13:12	LB137889

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Dal-Tile

SDG No.: Q3616

Contract: DALT01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	19.4	20.0	97	80 - 120	P	11/13/2025	13:16	LB137889
	Cadmium	6.14	6.0	102	80 - 120	P	11/13/2025	13:16	LB137889
	Chromium	9.82	10.0	98	80 - 120	P	11/13/2025	13:16	LB137889
	Copper	22.5	20.0	112	80 - 120	P	11/13/2025	13:16	LB137889
	Lead	12.0	12.0	100	80 - 120	P	11/13/2025	13:16	LB137889
	Nickel	40.8	40.0	102	80 - 120	P	11/13/2025	13:16	LB137889
	Selenium	23.9	20.0	120	80 - 120	P	11/13/2025	13:16	LB137889
	Zinc	42.2	40.0	106	80 - 120	P	11/13/2025	13:16	LB137889

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Dal-Tile

SDG No.: Q3616

Contract: DALT01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4930	5000	99	90 - 110	P	11/13/2025	13:46	LB137889
	Cadmium	2470	2500	99	90 - 110	P	11/13/2025	13:46	LB137889
	Chromium	1000	1000	100	90 - 110	P	11/13/2025	13:46	LB137889
	Copper	1240	1250	99	90 - 110	P	11/13/2025	13:46	LB137889
	Lead	4930	5000	99	90 - 110	P	11/13/2025	13:46	LB137889
	Nickel	2470	2500	99	90 - 110	P	11/13/2025	13:46	LB137889
	Selenium	5000	5000	100	90 - 110	P	11/13/2025	13:46	LB137889
	Zinc	2490	2500	100	90 - 110	P	11/13/2025	13:46	LB137889
CCV02	Arsenic	4900	5000	98	90 - 110	P	11/13/2025	14:35	LB137889
	Cadmium	2410	2500	96	90 - 110	P	11/13/2025	14:35	LB137889
	Chromium	973	1000	97	90 - 110	P	11/13/2025	14:35	LB137889
	Copper	1220	1250	98	90 - 110	P	11/13/2025	14:35	LB137889
	Lead	4830	5000	96	90 - 110	P	11/13/2025	14:35	LB137889
	Nickel	2420	2500	97	90 - 110	P	11/13/2025	14:35	LB137889
	Selenium	4960	5000	99	90 - 110	P	11/13/2025	14:35	LB137889
	Zinc	2470	2500	99	90 - 110	P	11/13/2025	14:35	LB137889
CCV03	Arsenic	4890	5000	98	90 - 110	P	11/13/2025	15:48	LB137889
	Cadmium	2420	2500	97	90 - 110	P	11/13/2025	15:48	LB137889
	Chromium	982	1000	98	90 - 110	P	11/13/2025	15:48	LB137889
	Copper	1220	1250	98	90 - 110	P	11/13/2025	15:48	LB137889
	Lead	4860	5000	97	90 - 110	P	11/13/2025	15:48	LB137889
	Nickel	2420	2500	97	90 - 110	P	11/13/2025	15:48	LB137889
	Selenium	4910	5000	98	90 - 110	P	11/13/2025	15:48	LB137889
	Zinc	2490	2500	100	90 - 110	P	11/13/2025	15:48	LB137889
CCV04	Arsenic	5010	5000	100	90 - 110	P	11/13/2025	16:36	LB137889
	Cadmium	2500	2500	100	90 - 110	P	11/13/2025	16:36	LB137889
	Chromium	1020	1000	102	90 - 110	P	11/13/2025	16:36	LB137889
	Copper	1260	1250	101	90 - 110	P	11/13/2025	16:36	LB137889
	Lead	5010	5000	100	90 - 110	P	11/13/2025	16:36	LB137889
	Nickel	2490	2500	100	90 - 110	P	11/13/2025	16:36	LB137889
	Selenium	5040	5000	101	90 - 110	P	11/13/2025	16:36	LB137889
	Zinc	2580	2500	103	90 - 110	P	11/13/2025	16:36	LB137889
CCV05	Arsenic	4900	5000	98	90 - 110	P	11/13/2025	17:27	LB137889
	Cadmium	2420	2500	97	90 - 110	P	11/13/2025	17:27	LB137889

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Dal-Tile

SDG No.: Q3616

Contract: DALT01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Chromium	983	1000	98	90 - 110	P	11/13/2025	17:27	LB137889
	Copper	1230	1250	98	90 - 110	P	11/13/2025	17:27	LB137889
	Lead	4860	5000	97	90 - 110	P	11/13/2025	17:27	LB137889
	Nickel	2420	2500	97	90 - 110	P	11/13/2025	17:27	LB137889
	Selenium	4910	5000	98	90 - 110	P	11/13/2025	17:27	LB137889
	Zinc	2510	2500	100	90 - 110	P	11/13/2025	17:27	LB137889
CCV06	Arsenic	4940	5000	99	90 - 110	P	11/13/2025	18:17	LB137889
	Cadmium	2440	2500	98	90 - 110	P	11/13/2025	18:17	LB137889
	Chromium	1010	1000	101	90 - 110	P	11/13/2025	18:17	LB137889
	Copper	1240	1250	99	90 - 110	P	11/13/2025	18:17	LB137889
	Lead	4900	5000	98	90 - 110	P	11/13/2025	18:17	LB137889
	Nickel	2440	2500	98	90 - 110	P	11/13/2025	18:17	LB137889
CCV07	Selenium	4940	5000	99	90 - 110	P	11/13/2025	18:17	LB137889
	Zinc	2580	2500	103	90 - 110	P	11/13/2025	18:17	LB137889
	Arsenic	5010	5000	100	90 - 110	P	11/13/2025	18:38	LB137889
	Cadmium	2470	2500	99	90 - 110	P	11/13/2025	18:38	LB137889
	Chromium	1010	1000	101	90 - 110	P	11/13/2025	18:38	LB137889
	Copper	1260	1250	100	90 - 110	P	11/13/2025	18:38	LB137889
	Lead	4960	5000	99	90 - 110	P	11/13/2025	18:38	LB137889
	Nickel	2470	2500	99	90 - 110	P	11/13/2025	18:38	LB137889
	Selenium	4990	5000	100	90 - 110	P	11/13/2025	18:38	LB137889
	Zinc	2610	2500	104	90 - 110	P	11/13/2025	18:38	LB137889