



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Dal-Tile

SDG No.: Q3616

Contract: DALT01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	22.3	20.0	111	65 - 135	P	11/13/2025	13:25	LB137889
	Cadmium	6.14	6.0	102	65 - 135	P	11/13/2025	13:25	LB137889
	Chromium	10.3	10.0	103	65 - 135	P	11/13/2025	13:25	LB137889
	Copper	22.9	20.0	115	65 - 135	P	11/13/2025	13:25	LB137889
	Lead	14.4	12.0	120	65 - 135	P	11/13/2025	13:25	LB137889
	Nickel	41.1	40.0	103	65 - 135	P	11/13/2025	13:25	LB137889
	Selenium	25.8	20.0	129	65 - 135	P	11/13/2025	13:25	LB137889
	Zinc	42.5	40.0	106	65 - 135	P	11/13/2025	13:25	LB137889
CRA	Mercury	0.22	0.2	108	70 - 130	CV	11/17/2025	13:11	LB137914