

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Dal-Tile

SDG No.: Q3616

Contract: DALT01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	7.42			-20	20	11/13/2025	13:29	LB137889
	Cadmium	-0.18	1.0	18	-5	7	11/13/2025	13:29	LB137889
	Chromium	49.7	52.0	96	42	62	11/13/2025	13:29	LB137889
	Copper	2.75	2.0	138	-18	22	11/13/2025	13:29	LB137889
	Lead	-8.14			-12	12	11/13/2025	13:29	LB137889
	Nickel	6.06	2.0	303	-38	42	11/13/2025	13:29	LB137889
	Selenium	-5.23			-20	20	11/13/2025	13:29	LB137889
	Zinc	1.59			-40	40	11/13/2025	13:29	LB137889
ICSAB01	Arsenic	110	104	106	88.4	120	11/13/2025	13:33	LB137889
	Cadmium	982	972	101	826	1120	11/13/2025	13:33	LB137889
	Chromium	555	542	102	460	624	11/13/2025	13:33	LB137889
	Copper	487	511	95	434	588	11/13/2025	13:33	LB137889
	Lead	42.2	49.0	86	37	61	11/13/2025	13:33	LB137889
	Nickel	1000	954	105	810	1100	11/13/2025	13:33	LB137889
	Selenium	54.7	46.0	119	26	66	11/13/2025	13:33	LB137889
	Zinc	1010	952	106	809	1095	11/13/2025	13:33	LB137889