

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3617

Project: Linden Generating Station - 402

Run No.: LB137878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137878BS							
TPH	mg/L	20.0	19.0		95	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3617

Project: Linden Generating Station - 402

Run No.: LB137878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137878BSD							
TPH	mg/L	20.0	17.8		89	1	78-114	11/13/2025