



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

RunNo.: LB137879

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90–110	11/13/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90–110	11/13/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90–110	11/13/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

RunNo.: LB137886

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.488	0.5	98	90-110	11/13/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/13/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.491	0.5	98	90-110	11/13/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.485	0.5	97	90-110	11/13/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

RunNo.: LB137901

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.095	0.099	96	90–110	11/14/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90–110	11/14/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90–110	11/14/2025
Sample ID: CCV3 Cyanide	mg/L	0.24	0.25	96	90–110	11/14/2025
Sample ID: CCV4 Cyanide	mg/L	0.25	0.25	100	90–110	11/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

RunNo.: LB137936

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	1080	1000	108	90–110	10/23/2025
Sample ID: CCV1 TOC	mg/L	1020	1000	102	90–110	11/18/2025
Sample ID: CCV2 TOC	mg/L	1020	1000	102	90–110	11/18/2025