

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

Run No.: LB137936

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137936BSS							
TOC	mg/Kg	1000	1020		102	1	90-110	11/18/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

Run No.: LB137886

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170537BS							
Hexavalent Chromium	mg/Kg	20	20.6		103	1	84-110	11/13/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q3618

Project: Whittaker Coatings Site – E9125B

Run No.: LB137901

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170546BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	11/14/2025