

Continuing Calibration Report for SequenceID : FF111725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016716.D

Aromatic C10-C12	5495483.000	40.000	4.502	6.378	137387.075	131963.544	-4.110
Aromatic C12-C16	8357375.000	60.000	6.379	9.083	139289.583	135166.617	-3.050
Aromatic C16-C21	10464575.000	80.000	9.084	13.405	130807.188	133743.127	2.195
Aromatic C21-C36	22365155.000	180.000	13.406	18.870	124250.861	117666.204	-5.596
Aromatic EPH	46682588.000	360.000	4.502	18.870	129673.856	125744.182	-3.125