

Continuing Calibration Report for SequenceID : FF111725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016731.D

Aromatic C10-C12	5413826.000	40.000	4.502	6.378	135345.650	131963.544	-2.563
Aromatic C12-C16	8196999.000	60.000	6.379	9.083	136616.650	135166.617	-1.073
Aromatic C16-C21	10258552.000	80.000	9.084	13.405	128231.900	133743.127	4.121
Aromatic C21-C36	22208488.000	180.000	13.406	18.870	123380.489	117666.204	-4.856
Aromatic EPH	46077865.000	360.000	4.502	18.870	127994.069	125744.182	-1.789