

Continuing Calibration Report for SequenceID : FG111725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016925.D

Aliphatic C9-C12	8312412.000	60.000	3.260	6.901	138540.200	122003.227	-13.555
Aliphatic C12-C16	5689870.000	40.000	6.902	10.358	142246.750	122851.240	-15.788
Aliphatic C16-C21	8956347.000	60.000	10.359	13.749	149272.450	132396.253	-12.747
Aliphatic C21-C28	11592307.000	80.000	13.750	17.436	144903.838	139034.659	-4.221
Aliphatic C28-C40	13646977.000	120.000	17.437	22.460	113724.808	113794.545	0.061
Aliphatic EPH	48197913.000	360.000	3.260	22.460	133883.092	124878.157	-7.211