

Continuing Calibration Report for SequenceID : FG111725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016938.D

Aliphatic C9-C12	7912613.000	60.000	3.260	6.901	131876.883	122003.227	-8.093
Aliphatic C12-C16	5407778.000	40.000	6.902	10.358	135194.450	122851.240	-10.047
Aliphatic C16-C21	8530754.000	60.000	10.359	13.749	142179.233	132396.253	-7.389
Aliphatic C21-C28	10984596.000	80.000	13.750	17.436	137307.450	139034.659	1.242
Aliphatic C28-C40	13085578.000	120.000	17.437	22.460	109046.483	113794.545	4.172
Aliphatic EPH	45921319.000	360.000	3.260	22.460	127559.219	124878.157	-2.147