

Continuing Calibration Report for SequenceID : FG112025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016956.D

Aliphatic C9-C12	8189375.000	60.000	3.258	6.899	136489.583	122003.227	-11.874
Aliphatic C12-C16	5612132.000	40.000	6.900	10.356	140303.300	122851.240	-14.206
Aliphatic C16-C21	8831800.000	60.000	10.357	13.743	147196.667	132396.253	-11.179
Aliphatic C21-C28	11474845.000	80.000	13.744	17.429	143435.563	139034.659	-3.165
Aliphatic C28-C40	13319255.000	120.000	17.430	22.445	110993.792	113794.545	2.461
Aliphatic EPH	47427407.000	360.000	3.258	22.445	131742.797	124878.157	-5.497