

Continuing Calibration Report for SequenceID : FG112025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016963.D

Aliphatic C9-C12	7779432.000	60.000	3.258	6.899	129657.200	122003.227	-6.274
Aliphatic C12-C16	5321494.000	40.000	6.900	10.356	133037.350	122851.240	-8.291
Aliphatic C16-C21	8375437.000	60.000	10.357	13.743	139590.617	132396.253	-5.434
Aliphatic C21-C28	10879586.000	80.000	13.744	17.429	135994.825	139034.659	2.186
Aliphatic C28-C40	13293839.000	120.000	17.430	22.445	110781.992	113794.545	2.647
Aliphatic EPH	45649788.000	360.000	3.258	22.445	126804.967	124878.157	-1.543