

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Instrument ID: ECD_S

Calibration Date(s): 11/03/2025 11/03/2025

Calibration Times: 09:40 11:17

GC Column: RTX-CLP

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS032226.D</u>	CF 500 =	<u>PS032227.D</u>		
CF 750 =		<u>PS032228.D</u>	CF 1000 =	<u>PS032229.D</u>	CF 1500 =	<u>PS032230.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	24885400000	24929800000	24576300000	22857100000	21942500000	23838200000	6
2,4,5-TP(Silvex)	29002100000	27213000000	26386400000	24343100000	23113100000	26011600000	9
2,4-D	5213180000	4699390000	4506050000	4152780000	3987670000	4511810000	11
2,4-DB	2485210000	2665610000	2725680000	2615770000	2683300000	2635120000	4
2,4-DCAA	5119490000	4707160000	4538440000	4190170000	4101110000	4531270000	9
DICAMBA	20442900000	19695500000	19208400000	17857400000	17126700000	18866200000	7
DICHLORPROP	5368470000	4545610000	4374370000	4053670000	3901680000	4448760000	13
Dinoseb	24209400000	21719600000	20866000000	19171400000	18071500000	20807600000	11

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Instrument ID: ECD_S

Calibration Date(s): 11/03/2025 11/03/2025

Calibration Times: 09:40 11:17

GC Column: RTX-CLP2

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS032226.D</u>	CF 500 =	<u>PS032227.D</u>		
CF 750 =		<u>PS032228.D</u>	CF 1000 =	<u>PS032229.D</u>	CF 1500 =	<u>PS032230.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	24492800000	24868100000	24604000000	22804200000	21821100000	23718000000	6
2,4,5-TP(Silvex)	30276100000	28543900000	27314100000	24890000000	23245400000	26853900000	10
2,4-D	5684310000	5535780000	5360230000	4911890000	4736080000	5245660000	8
2,4-DB	3827300000	3872610000	3837710000	3693680000	3625410000	3771340000	3
2,4-DCAA	3895870000	3715630000	3686230000	3472610000	3457140000	3645500000	5
DICAMBA	18687400000	18066800000	17633600000	16322600000	15651800000	17272400000	7
DICHLORPROP	5449030000	4964760000	4780180000	4399810000	4251930000	4769140000	10
Dinoseb	19878000000	18406900000	17876900000	16527900000	15747300000	17687400000	9