

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 11/03/2025

11/03/2025

Continuing Calib Time: 10:13

Initial Calibration Time(s): 09:40

11:17

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.46	7.47	7.37	7.57	0.01
2,4-DCAA	7.28	7.28	7.18	7.38	0.00
DICHLORPROP	8.17	8.17	8.07	8.27	0.00
2,4-D	8.41	8.41	8.31	8.51	0.00
2,4,5-TP(Silvex)	9.28	9.28	9.18	9.38	0.00
2,4,5-T	9.58	9.58	9.48	9.68	0.00
2,4-DB	10.15	10.16	10.06	10.26	0.01
Dinoseb	11.34	11.35	11.25	11.45	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3618

Continuing Calib Date: 11/13/2025

Initial Calibration Date(s): 11/03/2025 11/03/2025

Continuing Calib Time: 10:13

Initial Calibration Time(s): 09:40 11:17

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.81	7.81	7.71	7.91	0.00
2,4-DCAA	7.62	7.62	7.52	7.72	0.00
DICHLORPROP	8.52	8.52	8.42	8.62	0.00
2,4-D	8.87	8.86	8.76	8.96	0.00
2,4,5-TP(Silvex)	9.75	9.74	9.64	9.84	0.00
2,4,5-T	10.17	10.17	10.07	10.27	0.00
2,4-DB	10.74	10.74	10.64	10.84	0.00
Dinoseb	11.11	11.11	11.01	11.21	0.01