

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032399.D **Time Analyzed:** 15:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.575	9.478	9.678	796.320	712.500	11.8
2,4,5-TP(Silvex)	9.275	9.180	9.380	780.830	712.500	9.6
2,4-D	8.411	8.314	8.514	766.670	705.000	8.7
2,4-DB	10.153	10.055	10.255	773.600	712.500	8.6
2,4-DCAA	7.282	7.183	7.383	819.180	750.000	9.2
DICAMBA	7.463	7.366	7.566	781.910	705.000	10.9
DICHLORPROP	8.171	8.074	8.274	751.230	705.000	6.6
Dinoseb	11.339	11.245	11.445	753.550	705.000	6.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/13/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032399.D **Time Analyzed:** 15:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.171	10.072	10.272	786.230	712.500	10.3
2,4,5-TP(Silvex)	9.743	9.644	9.844	783.760	712.500	10.0
2,4-D	8.864	8.764	8.964	777.750	705.000	10.3
2,4-DB	10.737	10.637	10.837	777.340	712.500	9.1
2,4-DCAA	7.615	7.517	7.717	827.470	750.000	10.3
DICAMBA	7.806	7.708	7.908	787.340	705.000	11.7
DICHLORPROP	8.521	8.422	8.622	776.340	705.000	10.1
Dinoseb	11.104	11.005	11.205	756.440	705.000	7.3