

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032413.D **Time Analyzed:** 21:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.572	9.478	9.678	805.390	712.500	13.0
2,4,5-TP(Silvex)	9.274	9.180	9.380	781.920	712.500	9.7
2,4-D	8.407	8.314	8.514	773.900	705.000	9.8
2,4-DB	10.149	10.055	10.255	825.530	712.500	15.9
2,4-DCAA	7.280	7.183	7.383	810.900	750.000	8.1
DICAMBA	7.461	7.366	7.566	779.510	705.000	10.6
DICHLORPROP	8.168	8.074	8.274	751.080	705.000	6.5
Dinoseb	11.338	11.245	11.445	747.920	705.000	6.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3618
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032413.D **Time Analyzed:** 21:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.170	10.072	10.272	795.220	712.500	11.6
2,4,5-TP(Silvex)	9.743	9.644	9.844	774.190	712.500	8.7
2,4-D	8.862	8.764	8.964	774.300	705.000	9.8
2,4-DB	10.735	10.637	10.837	771.420	712.500	8.3
2,4-DCAA	7.615	7.517	7.717	839.420	750.000	11.9
DICAMBA	7.807	7.708	7.908	787.460	705.000	11.7
DICHLORPROP	8.520	8.422	8.622	776.600	705.000	10.2
Dinoseb	11.104	11.005	11.205	751.770	705.000	6.6